

# WEST BENGAL MEDICAL SERVICES CORPORATION LIMITED

(Wholly owned by the Government of West Bengal)

Regd. Off: Swasthya Bhawan, GN-29, Sector-V, Salt Lake, Kolkata-91

Phone No. 033-40335423

Email: [deo.wbmsc@gmail.com](mailto:deo.wbmsc@gmail.com), [mp.wbmsc@gmail.com](mailto:mp.wbmsc@gmail.com)

Memo No: WBMSC/824

Date: 04/04/2014

## INVITATION FOR QUOTATION FOR QUALITY ANALYSIS OF COMBINED URINE STICK

West Bengal Medical Service Corporation Limited is inviting sealed quotation for Quality Analysis (QA) of the Combined Urine Stick from reputed NABL Accredited Laboratories.

Detail of job is noted in the table below:

Schedule No.	Name of the work	QA Test
01.	Quality Analysis of Combined Urine Stick for detection of Protein and Glucose in Urine	03 (batches)

West Bengal Medical Services Corporation Ltd. intends to engage reputed & qualified NABL accredited Laboratories having the qualifications noted hereunder.

1. NABL Accreditation.
2. IT Return of last 3 years (2010-11, 2011-12, 2012-13)
3. Valid PAN Number.
4. Valid Trade License.

The Laboratory should submit documents in support of the above criteria along with their quotation.

### Quoted Rate:

- The quoted rate shall be valid for the entire quantity as mentioned above. Corrections if any shall be made by crossing out, initialing, dating and re writing.
- All duties, taxes and other levies payable by the selected bidder for the work shall be included in the quoted price.
- The rates quoted by the bidder shall include the cost of courier/postage charges to deliver test report at WBMSCL office.
- The price to be quoted in Indian rupees only

Each bidder shall submit only one quotation.

### Validity of quotation:

- Quotation shall remain valid for a period not less than 3 months after the deadline date specified for submission

### Evaluation of quotation:

- Purchaser will evaluate and compare the quotations determined to be substantially responsive i.e. which
  - (a) Are properly signed
  - (b) Conform to the terms and conditions

**Schedule:**

<b>1.</b>	Last date of submission of quotation	<b>16.04.2014 up to 1:00 P.M</b>
<b>2.</b>	Date of opening of quotation	<b>16.04.2014 at 2:00 P.M</b>

**Submission & Acceptance of quotation:**

- Prospective bidders are requested to submit the quotations in a sealed cover via courier/post addressing Managing Director, WBMSC Ltd. at Swasthya Bhawan, GN-29, Sector-V, Salt Lake, Kolkata – 700091 within the stipulated date & time.
- The purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest evaluated quotation price.
- The bidder whose bid is accepted will be notified of the award of contract by the purchaser prior to expiration of the quotation validity period. The terms of the accepted offer shall be incorporated in the purchase order.

**Timeline for submission of report:**

The Laboratory should send their test report within 8 (eight) days of the receipt of sample.

**Payment terms**

Payment should be process within 7 (Seven) days after giving test report and Bills.

WBMSCL reserves the right to accept or reject any quotation and to cancel the bidding process and reject all quotations at any time without assigning any reason whatsoever.

Sd/-  
**Managing Director**

## Specification of Combined Urine Stick

Combined Stick for detection of Protein and Glucose in urine for identification of pregnant mother in Pre Eclamptic stage. The stick should be

### Features

- Dip Stick method
- Single use
- Rapid test result within 2 minutes approx from freshly collected urine
- Results to be given both in mmol /dl & mg/dl
- Test result comparative marker indicator should be printed on the unit container/bottle.
- Guidelines of use (Leaflet should be inserted in each 100 stick unit)
- Manufacturer Should have GMP, product should preferably be ISO / CE / US FDA Approved
- Room Temperature Storage between (2 to 40°C). Should be suitable in Indian Climate
- Testing temperature 15 to 30° C

**The Colour patch of the dipstick should change for the following for Glucose (with tolerance of  $\pm 15\%$ )**

- Negative:
- Trace: 5 mmol /dl
- Glucose 1 +: 15 mmol /dl
- Glucose 2 +: 30 mmol /dl
- Glucose 3 +: 60 mmol /dl
- Glucose 4 +: 110 mmol /dl

**The Colour patch of the dipstick should change for the following for Protein (with tolerance of  $\pm 15\%$ )**

- Negative: < 10 mg/dl
- Trace: 10 – 20 mg/dl
- Protein 1 +: 30 mg/dl
- Protein 2 +: 100 mg/dl
- Protein 3 +: 300 mg/dl
- Protein 4 +: 1000 mg/dl

**Quoting Rate**

<b>Sl No.</b>	<b>Name of the work</b>	<b>QA Test</b>	<b>Unit rate for a batch in (Rs.)</b>	<b>Total Amount (Rs.)</b>
<b>(1)</b>	<b>(2)</b>	<b>(3)</b>	<b>(4)</b>	<b>(5)</b>
01.	Quality Analysis of Combined Urine Stick for detection of Protein and Glucose in Urine	03 (batches)		

**Signature of bidder**