

Notice Inviting e-Tender
West Bengal Medical Services Corporation Limited
Swasthya Bhawan Complex
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 23577070/40340307/40340308

E mail: deo.wbmsc@gmail.com, md.wbmsc@gmail.com

SUPPLY OF BLOOD-BANK EQUIPMENTS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 48 /2014

Dated- 02.08.2014

The following amendment have been made in the tender document,

Amendment - IX

UPDATED & REVISED TECHNICAL SPECIFICATION

SCHEDULE - VI

Di Electric Tube Sealer

(a) Type	:	Vertical
(b) Power supply	:	100-240 V AC (universal input)
(c) Input frequency	:	50/60 Hz
(d) Compliance	:	EN 61010-1
(e) Operation	:	Continuous operation
(f) Power consumption	:	Operating 170 W / Standby 20W
(g) RF output power	:	20 W
(h) RF output frequency	:	40.68 MHz
(i)Weight	:	3 - 7 kg
(j) Tube detection	:	Automatic
(k) Max. dia of tube that can be Sealed	:	6 mm

(l) Sealing time	:	1 -3 sec. (Adjustable)
(m) Colour coded light indicator	:	<ul style="list-style-type: none"> i) Cover open ii) Power iii) Ready iv) Seal
(n)Dimensions (WxDxH) mm	:	Suitable for table top installation
(o)Fuse	:	250V, 3.15A type M, slow blow
(p) Standards & approvals	:	CE / FDA / BIS or any other indian or internationally approved safety standard certification
(q) Manufacturing Standard	:	ISO 9001:2008 or any other Indian or internationally approved standard certification, EN ISO 13485:2012 (for electrical & biosafety hazard)