

Notice Inviting e-Tender
West Bengal Medical Services Corporation Limited
Swasthya Bhawan Complex
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 23577070/40340307/40340308

E mail: deo.wbmsc@gmail.com, md.wbmsc@gmail.com

**SUPPLY OF RADIANT WARMER WITH BASSINET IN THE HOSPITALS AND MEDICAL COLLEGES
OF THE GOVERNMENT OF WEST BENGAL.**

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-63/2014	Dated-29.11.2014
---------------------------------------	------------------

2nd call of tender no. WBMSC/ NIT-50/2014, dated 08.08.2014

The following amendment have been made in the tender document,

Amendment – III

**Replace the following (marked in Red) in Amendment I (Revised
Technical Specification) published on 09.12.2014**

2. The radiant warmer should have a heating element with power consumption of 800 Watts or less. It should have calrod (tubular) / quartz / non-ceramic infrared heating element with parabolic stainless steel reflector for uniform radiation of heat waves to all corners of the bassinet. Heater should be protected by a metal grid to prevent shock hazards.

AS UNDER

2. The radiant warmer should have a heating element with power consumption of 800 Watts or less. It should have calrod (tubular) / quartz / non-ceramic infrared heating element with parabolic stainless steel / aluminium reflector for uniform radiation of heat waves to all corners of the bassinet. Heater should be protected by a metal grid to prevent shock hazards.