



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/20

E mail: procurement@wbmsc.gov.in

Supply and Commissioning of different types of Medical Equipment at Medical College and Hospital of the Govt. of West Bengal

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-215/2018

Dated-12.10.2018

Amendment-IV

REVISED TECHNICAL SPECIFICATION

600 mA X-RAY :

Parameter / Feature	Threshold Limit / Range
XRAY TUBE	
mAs in 0.9 Sec & at 90 KV	300 or more
Must Have Specification	
Parameter / Feature	Yes / No

Accessories

LEAD HANGER STAND (TO HOLD 6 HANGERS)-02 NOS	
---	--