

Notice Inviting e-Tender
West Bengal Medical Services Corporation Limited
Swasthya Bhawan Complex
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 23577070/40340307/40340308

E mail: deo.wbmisc@gmail.com, md.wbmisc@gmail.com

SUPPLY OF BLOOD-BANK EQUIPMENTS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-60/2014

Dated- 31.10.2014

The following amendment have been made in the tender document,

Amendment - I

REVISED TECHNICAL SPECIFICATION

Schedule V

DI ELECTRIC TUBE SEALER (PORTABLE)

- 1 Blood Bag tube sealer is a compact equipment to seal the Blood Bag pilot tubing.
- 2 The system should be able to seal the Blood Bag tubes quickly and effectively.
- 3 Should be simple to handle.
- 4 System should gently seal the tubing with no haemolysis using radiofrequency.
- 5 Should be capable of making wide seal of 2 mm thickness.
- 6 The sealing time should not be more than 2 seconds.
- 7 Minimum weight should not exceed 2.5 Kg with all accessories & rechargeable battery & charger. A carrying case should be included. The battery should have at least warranty of 1 year. Price of the battery should be quoted in Form 10 (b).
- 8 Should have indication lamps for "Sealing Process" on handle or main unit.

- 9 No warm – up time should be required.
- 10 Should ensure easy separation of tube segments after the sealing.
- 11 System should run on battery with 8 - 10 hrs. back up.
- 12 Back up battery should seal more than 800 seals on PVC - tubes in continuous mode.
- 13 The units shall be capable of operating continuously in ambient temperature of 10 – 40 °C and relative humidity of 15 – 90%.
- 14 Power input: 220 – 240V/50 Hz AC Single phase fitted with appropriate Indian plugs and sockets.
- 15 Electrodes should be well protected by a cover.
- 16 Certification:
 - Product certification: CE / US FDA / BIS certified.
 - Quality certification: ISO certified.