

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Bhawan Complex
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 23577070/40340307/40340308

E mail: deo.wbmsc@gmail.com, md.wbmsc@gmail.com

SUPPLY OF BLOOD-BANK EQUIPMENTS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 48 /2014

Dated- 02.08.2014

The following amendment have been made in the tender document,

Amendment - I

REVISED TECHNICAL SPECIFICATION

SCHEDULE - IV

Donor Station (Donor Bed with Couch System)

- | | |
|----------------------------------|---|
| (a) Power source | : 230 ± 10% VAC, 50 Hz, single phase |
| (b) Lifting capacity (approx.) | : 150 kg (Donor Weight). |
| (c) Interface | : For connecting Blood Collection Monitor |
| (d) Classification | : Class I, (similar to Type B applied part) protection against electric shock |
| (e) Power Consumption | : Operating 100W (max.) /Standby 5W max. |
| (f) Movement / Position | : Any intermediate position between seating position / donation position and vasovagal attack position.
Head end side & foot end side should be adjustable |
| (g) Movement actuation | : DC motor and torque multiplier |
| (h) Initiation of motor movement | : By operating the rocker switch / remote control |
| (i) Arm rests | : Both swivel able and height adjustable |
| (j) Upholstery | : Soft moulded washable (easily cleanable) upholstery of 2.5" thickness |

(k) Manoeuverability	: Provided with lockable castor wheels
(l) Accessories	: BCM tray, BP apparatus tray, tray for Accessories and telescopic IV stand.
(m) Weight	: 75 – 90 kg
(n) Length of fixed seat	: 100 cm +/- 5 cm
(o) Length of swivel seat	: 90 cm +/- 5 cm
(p) Width	: 67 cm +/-5 cm
(q) Length of arm rest	: 60 cm +/- 5 cm
(r) Width of arm rest	: 15 cm +/- 2 cm
(s) Height of leg rest	: 80 cm +/- 5 cm
(t) Height of Seat	: 51 cm +/- 5 cm
(u) Vasovagal attack position	: 66 cm +/- 5 cm
(v) Max. height of back rest	: 110 cm +/- 5 cm
(w) Fuse	: 250 V, 1A, 5 * 20mm type M, SB
(x) Standards & approvals	: CE / FDA / BIS or any other indian or internationally approved safety standard certification
(y) Manufacturing Standard	: ISO 9001:2008 or any other indian or internationally approved standard certification, EN ISO 13485:2012 (for electrical & bio safety hazard)
(z) Warning	: Replace fuses with specified type and rating