

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Bhawan Complex
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/40340308

E mail: procurement@wbmsc.gov.in

SUPPLY OF BLOOD-BANK EQUIPMENTS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT- 6/2015

Dated- 29.01.2015

The following amendment have been made in the tender document,

Amendment - I

REVISED TECHNICAL SPECIFICATION

INCUBATOR WITH THERMOSTATIC CONTROL (45 LTR.)

- a) Control: Equipped with Microprocessor controlled with a large digital display.
- b) Work Chamber Volume: **minimum** 45 Litres.
- c) Working temp.: From room temp. +5 °C to 75 °C.
- d) Temp. deviation with respect to time: ± 0.2 °C.
- e) Display: Digital display located **in front panel of the equipment**.
- f) Easy to clean: An interchangeable silicon door gasket & removal of shelf supports make thorough cleaning and decontamination of the working chamber easy & safe.
- g) Inner casing: Should be made of corrosion resistant stainless steel.
- h) Shelves: Perforated shelves made of corrosion resistant stainless steel. Standard supply **2 / 3** nos. , inter shelves gap – 6" **approx**.
- i) Heat conduction: Gravity convection **or air circulation**.
- j) Heater on **rear / side / bottom of the equipment**.
- k) Power requirement: 230 V, single phase, 50 Hz.
- l) After sales service: Prompt & efficient after sales service should be available from Kolkata.
- m) Product certification: CE / US FDA / BIS certified.
- n) Quality Certification: ISO certificate.