

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340431/308

E mail: procurement@wbmsc.gov.in

PROCUREMENT OF MEDICAL EQUIPMENT FOR FIVE NEW MEDICAL COLLEGE AND HOSPITALS

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-207/2018

Dated –11.10.2018

Amendment-IV

Technical Specifications

Ammendment-IV

The specification of the equipment has been ammended in colour in the specification is detail below;

Name of the Discipline: ANATOMY		
Sl. No.	Name of the Equipments (Item)	Specification
3	Mortuary cooler with arrangement to keep 8 bodies	<p>Temperature Range : 2°C to 8°C Temperature Control : Digital Micro-processor Temperature Display : LED / LCD Display Construction : The entire unit should be 304 Stainless Steel Insulation : High Grade Poly Uretherane minimum 80 mm thickness (Green Panel/CFC Free) Doors : Made of Steel Sheets w/ Magnetic Gasket & External Lock Trolleys : Made of Steel and sliding on telescopic rails Compressor (2 nos) : ISI Marked & CFC Free Defrosting Type : Automatic Evaporator : Internal evaporator system Forced draught Internal drainage : Yes Integral temp gauge : Yes Refrigeration System : Air cooled Hermetically sealed Air Circulation : Forced Air Circulation Alarm : High / Low alarm system Internal Lighting : Waterproof fluorescent ceiling lamps Locking System : Standard Key Locks 2. Size of each chamber (inside): 2100 – 2400 mm (length) × 500 – 600 mm (breadth) × 300– 450 mm (height) Entire cabinet Size: 4000-4200 mm(Length) x 2300-2400 mm(Depth) x 1650-1785(Hight) 2 storied (4 column & 2 rows; i.e. maximum 2 bodies vertically) The product should be CE/USFDA/BIS. Power Supply : 220 Volts / 50 Hz Essentials : User & Technical Manual, Optional Accessories : Caters, Stabilizer, Loading Trolley, Instrument Trolley, Data Logger, Temperature Chart Recorder, Authorization Certificate</p>