

**Notice Inviting e-Tender**  
**West Bengal Medical Services Corporation Limited**  
**Swasthya Bhawan Complex**  
**GN-29, Salt Lake, Sector-V**  
**Kolkata-700091**

Phone No (033) 23577070/40340307/40340308

E mail: [deo.wbmisc@gmail.com](mailto:deo.wbmisc@gmail.com), [md.wbmisc@gmail.com](mailto:md.wbmisc@gmail.com)

**SUPPLY OF BLOOD-BANK EQUIPMENTS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.**

(Submission of Bid through *online*)

**Bid Reference No.: WBMISCL/NIT- 48 /2014**

**Dated- 02.08.2014**

**The following amendment have been made in the tender document,**

**Amendment - I**

**REVISED TECHNICAL SPECIFICATION**

**SCHEDULE - X**

**Centrifuge (Table Top)**

- a) Capacity : 16 holes x15ml
- b) Built in time and speed regulator (digital controlled) with suitable speed indication (digital display) and lid lock system
- c) RPM : 500 to 6000 rpm
- d) Power Supply : 220/240 volts, single phase, 50 Hz
- e) A line voltage corrector of suitable rating should form part of the configuration as per IS:9815/89
- f) Technical Literature: The firm shall submit printed illustrated technical literature/ leaflet indicating the model number. If quoted model is a modified version of their any standard product that also be indicated in the offer
- g) Standards & approvals : CE / FDA / BIS or any other indian or internationally approved safety standard certification

h) Manufacturing Standard : ISO 9001:2008 or any other indian or internationally approved standard certification, EN ISO 13485:2012 (for electrical & biosafety hazard)