

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi Complex

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/40340308

E mail: procurement@wbmsc.gov.in

SUPPLY OF C-ARM MACHINE & C-ARM COMPATIBLE OT TABLE IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT- 42/2015

Dated – 03.12.2015

The following amendment have been made in the tender document,

Amendment – IV

For Schedule I

C-ARM MACHINE

1. TECHNICAL SPECIFICATION

C-Arm Machine

1. I.I.T.V. SYSTEM:

- c) The center resolution should be minimum 48lp/cm at 9 field.