



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

SUPPLY AND COMMISSIONING OF MEDICAL EQUIPMENTS (SET-III) FOR TRAUMA CARE FACILITY FOR THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVT. OF WEST BENGAL

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT- 66/2017

Dated – 23.09.2017

2nd call of Bid reference no. WBMSCL/NIT-21/2017, dated: 26.04.2017 of Schedule –I, II, III, IV & V

SCHEDULE – II

The following amendments have been made in the tender document,

Amendment – IV

High Vaccum Suction Machine

1. Shall have **OEM (Original Equipment Manufacturer)** Motor of minimum ¼ H.P. capacity. The machine should be mounted on four wheels and handle for transportation
2. **Technical Specification:**
 - 2.1. The Suction pump should be oil immersed fitted on Motor shaft
 - 2.2. Suction pump should have line grinding internally.
 - 2.3. To facilitate maintenance the cover of machine should be easily to open from the top & sides

- 2.4. The suction machine should be capable of producing minimum vacuum of 500 approx mm Hg. Which should be adjustable and monitored by vacuum gauge of suitable range.
 - 2.5. It should have two bottles of 1 or 2 liters (As per requirement) with synthetic rubber lids. The bottle shall be fitted with the arrangement to prevent overflow of fluid.
 - 2.6. ON/OFF Switch and Power indicator should be available
 - 2.7. Body material: Base, top & panel made of rust proof and corrosion resistant moulded ABS/Stainless Steel. Jar/Bottle material: Autoclavable polycarbonate.
 - 2.8. Inbuilt maintenance free battery. Battery backup 45 minutes or more on full charge. Should be with built in charger.
 - 2.9. Should operate on 230V, 50 Hz supply.
- 3. Accessories, Spares and Consumables:**
- 3.1. System as specifies
 - 3.2. Power cord 2 metre long with 15A three pin plug top
 - 3.3. Lid set - 01
 - 3.4. Seal for bottle - 01 nos
 - 3.5. Suction tubing set - 02 set
- 4. Standards, Safety And Training:**
- 4.1. Manufacturer should have ISO certification
 - 4.2. System should have CE or US FDA approval.
 - 4.3. On Site Comprehensive training for staff and support services till customer satisfaction with the system.
- 5. Documentation:**
- 5.1. User/Technical/Maintenance manuals to be supplied in the English.
 - 5.2. Certificate of calibration and inspection.