



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

PROCUREMENT OF MEDICAL EQUIPMENT FOR FIVE NEW MEDICAL COLLEGE AND HOSPITALS

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-99/2020

Dated –13.07.2020

(2nd call of Bid Reference No.: WBMSCL/NIT-209/2018, Dated- 11.10.2018)

The following amendment has been made in the tender document,

Amendment – II

(In Page: 34)

1. TECHNICAL SPECIFICATIONS

Sl. No.	Equipment	Specifications
4	Autoclave Electric	<ol style="list-style-type: none">1. High Pressure, electrically heated vertical steam sterilizer is used for sterilization of surgical instruments, dressing material, lined, rubber, plastic material by means of saturated steam under pressure of 15 to 20 psi (adjustable). A part from health services it is also used in scientific research institutions and universities.2. Sturdy double walled construction with boiler made of stainless steel of SS-304 sheet is easy to operate. Outer shell is made of Stainless Steel. Stainless steel Bucket with handle

Sl. No.	Equipment	Specifications
		<p>with perforated wall. The boiler and outer shell is provided with air insulation: Lid is made of stainless steel plate and is tightened around by wing nuts. Moulded, jointless gaskets are made of Neoprene rubber: Sterilizer is hydraulically tested upto 40 psi.</p> <p>3. Fitted with water level arrangements to indicate water position inside the boiler, pressure guage, Air/steam release cock. 2 (two) Loaded safety Valve and one pressure control switch which can be set at any selected point from 10 psi. to 20 psi. +- 3 psi. and drain, ISI marked immersion heating element (approx KW) heats the water and steam to desired temperature and pressure.</p> <p>4. Size: Dia X Depth = 400 mm X 600mm</p> <p>5. Should be CE/US FDA/BIS</p>



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

PROCUREMENT OF MEDICAL EQUIPMENT FOR FIVE NEW MEDICAL COLLEGE AND HOSPITALS

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-99/2020

Dated –13.07.2020

(2nd call of Bid Reference No.: WBMSCL/NIT-209/2018, Dated- 11.10.2018)

The following amendment has been made in the tender document,

Amendment – III

Procurement of Item No.2 Starling's long extension Kymographs with time markers is dropped.