



## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340308/319  
E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

PROCUREMENT OF MEDICAL EQUIPMENT FOR MEDICAL COLLEGES AND HOSPITALS  
(GROUP 3)

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-52/2019

Dated –18.06.2019

(2<sup>nd</sup> call of Bid Reference No.: WBMSCL /NIT-217/2018, Dated –12.10.2018)

The following amendment has been made in the tender document,

### Amendment – 1

#### Schedule X

DLCO	
Parameter /Feature	THRESHOLD LIMIT/RANGE
Flow in L/sec	(+/-)16
volume per second (L/sec)	12
Flow Accuracy	2% or 0.020 L/sec
Volume Accuracy	+ 2% or 0.050L
CO/CO2 Analyzer	Type: infrared analyzer
Measuring Range in % of CO	(0-0.15) or more
Resolution in %	0.005 or less
Accuracy in %	0.05 or less

Tracer gas sensor(He/N2)	
Range in %	(0-50) or more
Resolution in %	0.05 or less
Accuracy in %	0.2 or less
Colour monitor (Touch screen/ non touch screen) size (diagonal) in inch	12 or more
Accessories propriotor based/ universal	
MUST HAVE SPECIFICATIONS	
In built Spirometer	
Automatic calibration	
In Spirometry: FVC, SVC, MVV, Pre & Post challenge	
Single breath DLCO	
Single & multi breath FRC lung volume studies interpretation as per ATS/ ERS guidelines	
Quality Standard and Safety Certification	
European CE or USFDA approved	
Scope of Supply	
Computer Trolley – 1 No.	
Cylinder Trolley – 1 No.	
Colour deskjet Printer – 1 No.	
Nose Clips – 50 Nos	
Gas Cylinder (DLCO Gas) – 2 No.	
Mouth Piece – 400 Nos	