



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/20

E mail: procurement@wbmsc.gov.in

Supply and Commissioning of different types of Medical Equipment at Medical College and Hospital of the Govt. of West Bengal

(Submission of Bid through *online*)

Bid Reference No.: WBMSC/NIT-215/2018

Dated-12.10.2018

Amendment-II

REVISED TECHNICAL SPECIFICATION

OPHTHALMIC USG B-SCAN :

PARAMETER / FEATURE	THRESHOLD VALUE / LIMIT
PROBE	
PROBE FREQUENCY IN MHZ	MINIMUM 12.5
SECTOR ANGLE	20°-60° OR MORE
DISPLAY	
DEPTH RANGE IN MM	30-60 OR MORE
AXIAL RESOLUTION (MM)	0.15 OR MORE
LATERAL RESOLUTION MM	0.2 OR LESS
MEASUREMENT	
GAIN IN DB	128 OR LESS
FILE EXPORT FEATURES	
VIDEO IN FRAMES	600
CINE WEIGHT	UP TO 256 FRAMES
MUST HAVE SPECIFICATION	

PARAMETER / FEATURE	YES/NO
PROBE	
MOTORIZED SECTOR B-PROBE	
DISPLAY	
LIVE ZOOM	
GAIN CURVES	
MEASUREMENT	
DISTANCE CALLIPER, AREA & ARROW TOOLS	
FILE EXPORT FEATURES	
EXPORT REPORT IN PDF	
EXPORT SNAPSHOT IN JPEG	
EXPORT MOVIE IN AVI	
DYNAMIC RECORDING	