

## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Bhawan Complex  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340307/40340308

E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

SUPPLY OF BLOOD-BANK EQUIPMENTS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT- 25 /2015

Dated – 13.07.2015

The following amendment have been made in the tender document,

### Amendment - I

## REVISED TECHNICAL SPECIFICATION

### SCHEDULE - VIII

## BINOCULAR COMPOUND MICROSCOPE

- a) **Eyepieces:** WF 10X (field number – 18 mm or better)
- b) **IPD:** 48 – 75mm
- c) **Objectives:** Parfocal, antifungus, infinity optical system, plan achromatic 4X, 10X, 40X, 100X (oil immersion)
- d) **Binocular Head:** Eyepiece tube Siedentopf type 30 ° inclined, 360 ° rotatable
- e) **Nosepiece:** Reversed Inward Quadruple
- f) **Condenser:** N.A. 1.25 Abbe condenser, continuously variable iris / diaphragm with filter holder (blue / green / yellow)
- g) **Stage:** Built in low position mechanical stage around 100 - 140mm x 140mm with fine vernier graduations designed with convenient coaxial adjustment for slide manipulation
- h) **Antifungus:** To protect the system from fungus formation in high-humidity environments, an anti-fungus coating is to be used to prolong the life of both microscope and objective

- i) **Illumination:** Built in 3W – 6W LED illumination with intensity control LED Bulb Life  
10,000 -30,000 hours
- j) **Body:** All gear, housing, body are metallic
- k) **Accessories included:** Universal power supply 100 - 240V, CE power cord & dust cover, immersion oil
- l) **Standards & approvals:** CE / US FDA / BIS or any other Indian or internationally approved safety standard certification
- m) **Manufacturing Standard:** ISO 9001:2008 certificate or any other Indian or internationally approved safety standard certification, EN ISO 13485:2012 (for electrical & biosafety hazard)