



**COVID-19
URGENT**

Notice Inviting e-Tender

**West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091**

Phone No (033) 40340308/412
E mail: procurement@wbmsc.gov.in

Procurement of medical equipment for COVID 19 care facilities of the State of West Bengal

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-207/2020

Dated –14.10.2020

The following amendment has been made in the tender document,

Schedule I

Amendment – II

Revised Specifications of Computerized Radiographic (CR) System with Printer

1. The system should comprise of:
 - a. Image recording system (Image plate and cassette)
 - b. Image reader
 - c. Image processing work station
 - d. Accessories and consumables
2. Should be of Single cassette (size – 14" x 17") speed – 50 cassette/hr or more at resolution of 10 pixel/mm
3. Acquisition of image at monitor after insertion of the cassette: 40 sec or less
4. Cassettes: Should provide the following cassettes;-
 - i) 10" X 12" cassette – 1 no.
 - ii) 8" X 10" cassette –1 no.
 - iii) 14" x 17" cassette – 4 no.

5. Standard resolution of 10 pixel/mm
6. Console with monitor of 17" or more – 1 no.
7. UPS from reputed manufacturer having at least 30 minute back up with appropriate KVA for the entire system.
8. System should have capability to pan, zoom, annotate, crop, noise suppression/control multiple image layout, customized formatting, CD/DVD and USB writing etc.
9. System should be DICOM compatible and scalable for connectivity to PACS.
10. Should provide with Dry Imager camera with **DPI 300 or more**. The camera should print all standard film sizes. The camera should have 2 online trays.
11. The equipment should be new and unused. The manufacturing date should not be more than 180 days when it would reach the consignee address.
12. The printout copy should bear the patient identification particular.
13. **Standard & Safety**
Should be of US FDA or CE ("Conformité Européene")
14. **Warranty & CMC**
The machine, UPS and all other accessories should be under warranty for 2 years.

It is mandatory that the supplier would configure LASER / Thermal printer of any reputed make available in Indian market on demand of WBMSCCL / purchaser for printing of images.