



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319

E mail: procurement@wbmsc.gov.in / mp@wbmsc.gov.in

SUPPLY AND COMMISSIONING OF MEDICAL EQUIPMENTS FOR BLOOD COMPONENT SEPARATION UNIT FOR THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVT. OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 76 /2018

Dated-25.04.2018

The following amendments have been made in the tender document,

Amendment-XI (Applicable for Schedule V)

TECHNICAL SPECIFICATION

PH Meter

1. It should be digital electronic pH Meter with combination pH electrodes with thin glass type suitable for use in eppendorf type tube (500 μ l to 1 ml).
2. Should be portable.
3. It should have pH range: 0.0 to 14.0 pH with 3^{1/2} digital display.
4. It should have 0.0 to 19999 mv range.
5. Its have accuracy – repeatability and resolution: 0.01 pH
6. 3 point calibrators should be provided
7. Temperature compensation & pH standardization: 00 to 100°C (manual)
8. It should be operational on 220-240 volts, 50 Hz, single phase supply

9. It should work properly on ambient Temperature: 00 to 40°C. 40°C at 95% RH.
10. It should have suitable in-built replaceable battery.
11. The details of battery should be provided.
12. Equipment should be supplied with suitable stand & carrying case to be provided.
13. Warranty: 2 years; CMC: 5 years
14. **Product Certification:** CE; Should be supported with documentations or US FDA (The product must be enlisted under US FDA and the quoted model must have US FDA permission to market in US & export to other countries.
15. **Quality certifications:** ISO certified