



CORRIGENDUM – II

to *Bidding Documents* for

Procurement of 30(thirty) Advance Life Support Ambulances for Department of Health & Family Welfare, Government of West Bengal

Bid Reference No.: WBMSCL/NIT-306/2022

Dated – 13.07.2022

Sl. No.	Page No	Section	Clause	Present clause / Provision	Changed clause / Provision
1	5	II	A-2	EMD-2 Lakhs online + 8 Lakhs BG	EMD- 2 Lakhs online + 8 Lakhs BG OR 10 Lakhs online
2	5	II	A-3	Timeline for Supply: 60 days	Time line for supply :90 days after award of contact
3	7	II	5(d)	Minium experience of supply 10(ten) units of ALS Ambulance in any single consignment during last three years to any Govt/Semi Govt organisation/PSUs	Minium experience of supply 08 (eight) units of ALS Ambulance in any single consignment during last three years to any Govt/Semi Govt organisation/PSUs
4	16	II	C (A)	During the tender	Addition Qualified Bidder will have to arrange for physical demonstration of the offered ALS ambulance (complete in all respect) at Swastha Bhawan, Kolkata – 700091 on the notified date (within 15 days of the notice) before the Technical Bid Evaluation Committee. The decision of the Technical Bid Evaluation Committee regarding acceptance of the offered model is final.
5	19	II	H (b)	After Sales Service:.....	Addition The supplier should provide a clear undertaking that they will arrange after sales service of the base vehicle by nearest authorised Service Centre throughout the West Bengal and liable to provide a credit facility for minimum 30 days to pay the bills (<i>if payable</i>) from the respective Health Units under the Department of Health & Family Welfare. Non – Adherence of this clause will be treated as legal obligation. (To be furnished on Non – Judicial Stamp Paper of ₹ 10/- at the time of delivery of the Ambulances)
6	25	III	1. A.		Generic Vehicle Parameters 1. Category of the vehicle: M2 & M3 1-A-2 to 1-A-10 : As per AIS-125 (Type – D ALS Ambulance)

7	26	III	1B(a)	Patient Cabin : Patient examination light (12V,55W)	Patient Cabin: Patient examination light should be minimum of 1650 Lux, measured at the stretcher surface in its lowest position. The minimum distance of the measurement shall be 750 mm below the light and in an area with a minimum diameter of 200mm.
8	26	III	1B (c)	Stair Chair & Spine Board	Stair Chair and Spine Board should be complaint with product standards such as EN-1865 and should be manufactured by ISO-13485 certified manufactures. All certificates should be from accredited/notified bodies.
9	26	III	1B (d)	Ambulance Cot –Auto Loading stretcher (Length-1850 mm x width-530mm) with two quick release patient restraints	Ambulance Cot –Auto Loading stretcher (Length-1850 max width-530mm) with two quick release patient restraints as per AIS-125, Part-1 is mandatory.
10	26	III	1B(f)	One Aluminium Portable Cylinder	One Aluminium portable Oxygen Cylinder with capacity of 2.2 litres. With Medical Grade Certified (European CE/US FDA) pressure regulator with integrated flowmeter and pressure port for delivery mask.
11	26	III	1B(g)	Patient Monitor & Defibrillator with AED Pad Extension Cable – NASAN, SANJEEVANI 1006 (Compact) / Comen-S5	Patient Monitor & Defibrillator with AED Pad Extension Cable – SPO₂, NIBP, ECG, Printer. NASAN, SANJEEVANI 1006 (Compact). In addition require to provide BP machine and stethoscope, Medicine tray, Thermometer, Oxygen cylinder with moks regulator Laryngoscope , Torch glucometer
12	26	III	1B(j)	Syringe pump – Skanray FloSkan / Comen-M260 (2nos)	Syringe pump – Skanray FloSkan (1no)
13	26 & 27	III	1B(k)	GAS SOURCE: Complete gas pipeline with tubing embedded in panel having superior STAINLESS STEEL outlet points for oxygen (1 no. 02 outlet for oral mask and 1 no. 02 outlet for ventilator) with Rail mounting systems (RMS) for	GAS SOURCE: Complete gas pipeline with tubing embedded in panel having superior STAINLESS STEEL outlet points for oxygen (1 no. 02 outlet for oral mask and 1 no. 02 outlet for ventilator) with Rail mounting systems (RMS) for loading and unloading the 2 nos. of “D” type bulk cylinders(with Oxygen),

				loading and unloading the 2 nos. of “D” type bulk cylinders (cylinder are not included), Low 02 Alarm, 02 Cylinder pressure Manometer and 0.2 micron dust filter	Low 02 Alarm, 02 Cylinder pressure Manometer and 0.2 micron dust filter
14	27	III	1B(i)	Patient Cabin: S. S. Wash basin, electrically foot switch operated with 25 liter water tank capacity.	Patient Cabin: S. S. Wash basin, electrically foot switch operated with Fresh Water Storage Tank & waste water tank (Total Capacity 20-25 liter)
15	27	III	2	<u>Medical Accessories related to patient:</u> A foldable seat for Doctor /Paramedic should be installed.....	<u>Medical Accessories related to patient:</u> Seat for Doctor /Paramedic should be installed.....
16	29	III	3Changing from one cylinder to the other should not affect the distribution pressure in any way and this changeover should occur automatically/ManuallyChanging from one cylinder to the other should not affect the distribution pressure in any way and this changeover should occur Manually
17	34	III	4EThese should be made from non-wood & non ferrous fire retardant material (ABS not necessary) in sync with the ambulance’s internal look and feel.....	These should be made from fire retardant material (ABS not necessary) in sync with the ambulance’s internal look and feel.....
18	40	III	5A	<u>Emblems, Marking & Colour Scheme</u> Complete body exterior should be uniform white in colour. All external marking should be retro-reflective in nature and materials used for the same should meet or exceed the requirements of ASTM D 4956, Standard Specification for Retro-reflective sheeting for Traffic Control. Guidelines in regards to Emblems and Markings for Ambulances issued by the Government from time to time shall be applicable. However, the quality parameters of the markings indicated above shall remain constant.	<u>Emblems, Marking & Colour Scheme</u> Complete body exterior should be uniform white in colour. All external marking should be retro- reflective in nature and materials used for the same should meet or exceed the requirements of ASTM D 4956, Standard Specification for Retro-reflective sheeting for Traffic Control. All exterior Graphic design & lettering should be as per requirement of the TIA (As per specific requirement of customer).
19	51	III	--	Form 4B	Deleted