



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

Supply and Commissioning of Rotary Microtome for Department of SDLD at IPGEM&R and SSKM Hospital

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-163/2023

Dated-10.04.2023

Amendment-I

REVISED TECHNICAL SPECIFICATION

Manual Rotary Microtome

- Manual Rotary Microtome with simultaneous Clockwise and Anti Clockwise coarse feed wheel movement as per the choice of the end users.
- Knife holder base with lateral displacement features, including locking lever.
- Standard specimen clamp with adapter should be provided with standard delivery
- Sectioning thickness selection should be selectable from 0.5 to 60 μ m with Sectioning thickness increments range from 0 to 2 μ m in 0.5 μ m increments, from 2 to 10 μ m in 1 μ m increments, from 10 to 20 μ m in 2 μ m increments, from 20 to 60 μ m in 5 μ m increments
- Vertical Stroke Length should be at least 58 mm along with Horizontal Specimen Feed of approx 25 mm
- Specimen orientation facility must be available 8 degree (through X,Y axis)
- Trimming thickness selection must be available with 2 different steps (preferably at 10 μ m and 50 μ m)
- Specimen Retraction must be available with provision of switching on/off
- System should come with high profile disposable blade system
- System should come with 1 pack of high profile disposable blade (at least 50 pcs./ pack).
- The supplier should have at least over 10 numbers of different installation bases in Kolkata.(preferably in govt. Hospital)

- The supplier must have Kolkata based after sales service support with manufacturers factory trained engineers and established track record for providing efficient after sales service.
- Certification: European CE(4 digit notified body)/USFDA/BIS