



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

SUPPLY AND COMMISSIONING OF MEDICAL EQUIPMENT AT MEDICAL COLLEGE AND HOSPITAL OF THE GOVERNMENT OF WEST BENGAL (Group 1).

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 216/2018

Dated-12.10.2018

The following amendments have been made in the tender document,

Amendment –II (Revision of Technical Specification)

The revised technical specifications for the item is given below,

Automated Cell Counter:

PARAMETER / VALUE	THRESHOLD RANGE / LIMIT
WHITE BLOOD CELL DIFFERENTIALS	6
NUMBER OF STANDARD PARAMETERS	28 OR MORE
MACHINE THROUGHPUT (NUMBER OF TEST) PER HOUR	80 OR MORE
MUST HAVE SPECIFICATIONS	
REAGENTS ARE COMPANY/BRAND SPECIFIC	
CYANIDE FREE METHOD FOR HEMOGLOBIN ESTIMATION	

CONTINUOUS SAMPLE LOADING FACILITY WITH SAMPLE LOADER	
COMPATIBILITY TO LIS	
BAR CODE READING FACILITY	
ONLINE UPS (30 MINUTES BACKUP) OF APPROPRIATE CAPACITY FOR THE ENTIRE SYSTEM (INDICATIVE MAKE: APC/LUMINOUS/EMMERSON/MICROTEK/AVO/EXIDE)	
CALIBRATOR AND 1 SET OF 3 LEVELS QUALITY CONTROL; REAGENTS WORTH OF 5000 TESTS; THE REAGENTS SHOULD BE SUPPLIED IN STAGGERED MANNER AS AND WHEN REQUIRED BY THE END USER	
ADDITIONAL 3 YEARS WARRANTY (i.e., TOTAL OF 5 YEARS WARRANTY)	
Quality Standard and Safety Certification	
EQUIPMENT: EUROPEAN CE IVD OR US FDA	
REAGENT: EUROPEAN CE IVD OR US FDA	

- 1) Indicative sample load per year is 45,000. Bidders are to quote for cost of 1 (one) reportable test result.
- 2) The reagent rate will be fixed for 5 years.
- 3) Financial bid will be evaluated based on the aggregated sum of the cost of equipment and cost for 2, 25,000 reportable test results (i.e. 45,000 per year x 5 years)
- 4) **The Cost per reportable test result s should include:**
 - a) All reagents
 - b) Priming
 - c) Shutdown
 - d) Cleaning
 - e) Calibration
 - f) Comprehensive Maintenance of the equipment
- 5) In case the numbers of samples are more than 60,000 per year, the bidder should supply reagent at 10% discounted rate. And when the case load would cross 1, 00,000 tests per year the Company should provide an addition new machine of same make and model free of cost as backup.
The bidders should quote cost of commensurate reagent packs in Form 9(a).

6) The selected bidder should provide free Comprehensive Maintenance Services in case the test load at facilities is maintained (60,000 per year for one unit and 1, 00,000 tests per year for 2 units) as detailed above at point 5.