



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340308/319

E mail: procurement@wbmsc.gov.in

SUPPLY AND COMMISSIONING OF MEDICAL EQUIPMENTS (SET-I) FOR TRAUMA CARE FACILITY FOR THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVT. OF WEST BENGAL

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT- 105/2017

Dated – 30.11.2017

SCHEDULE – I

The following amendments have been made in the tender document. The changes are highlighted in yellow in the document below,

Amendment – I

Revised technical specifications for Otoacoustic emissions (OAE) system: Facility for doing TEOAE, DPOAE

TEOAE Specifications:

1. Facility of click stimulus and Tone burst stimulus
2. Configurable stimulus intensity
3. Band analysis from 1 KHz to 5 KHz or more
4. Reproducibility in half octave bands
5. Should have suppression facility
6. Full cross correlation, frequency analysis with reproducibility and signal to noise data on single test or between test pairs
7. Customized TEOAE protocol

DPOAE specifications:

1. Frequency range Minimum of 500-10,000 Hz.
2. Number of test points per octave: Upto 32 points per octave
3. Intensity: f1 and f2 levels from 0 to 70 dB SPL.
4. Customizable measurement protocols.
5. Variable Ratio: f2/f1.
6. DP Definition Points: f2-f1; 2f2-f1; 2f1-f2; 3f1-2f2; 3f2-2f1; 4f1- 3f2; 4f2-3f1.
7. Contra lateral suppression facility
8. SNR assessment
9. Should supply printer for printing of reports.