



Notice Inviting e-Tender
West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319

E mail: procurement@wbmsc.gov.in / mp@wbmsc.gov.in

**Procurement of Medical Equipment for Medical College and Hospitals
(GROUP-5)**

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 75/2019

Dated- 08.08.2019

The following amendments have been made in the tender document,

AMENDMENT - II

TECHNICAL SPECIFICATIONS

Schedule - I

SLIT LAMP

Parameter / Feature	Threshold Limit / Range
Magnification selection in step drum rotation	5
Magnification ratio	6X,10X,16X,25X & 40X or more
Eyepieces	12.5 X
Slit width in mm	0-14
Slit length in mm	1-14
Slit angle in degree	0-180

MUST HAVE SPECIFICATIONS

Parameter / Feature	Yes / No
Galileo convergence	
Blue, red free, amber UV cut	
LED light source	
Applanation tonometer (hanging type with standard features)	
Integrated high resolution camera with Computer System: Specification of computer-Desktop (Dell/ HP/ Lenovo) Harddisk 1TB, RAM 8 GB, Processor i5 (minimum), color photo printer (inkjet), Monitor (LED) 22 inches	
Coaxial observation tube (assistance scope)-Monocular	
Illumination system for Meibomian gland study	
Motorised Table (Standard)	
Electrical Safety and Quality Certificaton	
European CE & US FDA approved	