



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

SUPPLY AND COMMISSIONING OF EYE EQUIPMENT FOR MULTI / SUPER SPECIALITY
HOSPITALS OF THE GOVT. OF WEST BENGAL
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 125/2017

Dated-27.12.2017

Amendment I:

Read Biometry Machine with Stand
as
Biometry Machine

Amendment II:

Section I: Instructions to Tenderers

A. Important information at a glance

7. Who can Bid (E)

Eye Equipment For Multi / Super Speciality Hospitals
1. Manufacturing Company or its subsidiary in India.
2. Manufacturer Authorized Distributor for the tender

Amendment III:

Section IV. Schedule of Requirements

2. TECHNICAL SPECIFICATION

Biometry Machine

A Scan Probe

1. Probe Frequency: 10 MHz
2. Fixation: Internal LED

Measurement Range

1. Axial Length: 15.0 mm to 40.0 mm or wider
2. ACD: 2.0 mm to 6.0 mm
3. Lens Thickness: 2.0 mm to 7.5 mm
4. Measurement Accuracy: ± 0.1 mm

Measurement Techniques

1. Method: Immersion & Contact
2. Mode: Auto & Manual
3. Lens Types: Aphakic, Normal, Cataract, Dense Cataract, Pseudophakic (Silicone, PMMA, Others), Silicone Oil.
4. Display: High Resolution 5 or more colour display with Touch Screen

Formula

1. SRK – T, SRK – II, Holladay, Hoffer- Q, Haigis & Post Refractive Formulas

Control

1. Storage: 100 Patient records, 10 Preset User, 25 IOL Configurations
2. Camp Mode: Fast Measurement with IOL Calculation in Single Screen
3. Printer: Fast Thermal Printer with Auto / Manual Cutter

Power

1. Electrical Requirements: 230V AC/50 Hz

Other Features

1. Auto & Manual Gain
2. Automatic Calculation of Standard deviation & average Axial length (Series of 10 measurements)
3. Simultaneous display of 2 or more different IOL Calculations
4. RTC Shows the date & Time throughout the life of the instrument
5. Various user programmable printouts Format
6. Measurement Possible for Various velocity of each Segment

Safety Features

1. Should have CE (“Conformite Europeene”) from European Union notified body having 4 digit identification number **or** US FDA

Optional Item

1. Soundless Motorized Up & Down Stand with table top which can accommodate two equipments.

Note: Bidder should quote the cost of optional item in Form 9 (a)

Amendment IV:

Section V. Bidding Forms

Form 9(a): PRICES FOR OPTIONAL, ACCESSORY & CONSUMABLE ITEMS

Sl. No	Items	Basic Price of 1(one) number in INR excluding GST
1		
2		
3		
4		

NOTE: The Tenderers should furnish the price of all the Consumables of the offered model in a separate sheet. If any Tenderer do not submit the price of any of the consumables, it will be presumed that those Consumables shall be supplied FREE OF COST by the Tenderer during Warranty & CMC period.