



## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340308/319  
E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

SUPPLY OF EQUIPMENTS FOR MOTHER &CHILD HUB OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSC/NIT- 67 /2018

Dated - 09.03.2018

2nd call of tender no. WBMSC/ NIT- 109 /2017, Dated-01.12.2017

### AMENDMENT - II

## Section IV. Schedule of Requirements

### 2. TECHNICAL SPECIFICATION

#### Schedule-II

##### Cardiotocography (CTG) equipment

1. Cardiotocography equipment with auto-correlation.
2. Wave form display and digits simultaneously.
3. Vibroacoustic stimulation and twin monitoring facilities desirable.
4. Computer compatibility software facilities.
5. Maturity sensed and automatic fetal movement detection.
6. Alert alarm for tachycardia and Bradycardia.
7. Battery (Lithium) back up for at least 3 - 4 hrs.
8. Patients data storage for at least 10 hrs
9. Should have facility for 3 displays [2 display for FHR (TWIN) and 1 for Uterine pressure]

10. Should be supplied with event marker facility for each foetal movement perceived & noted by the mother.
11. Should be supplied with a Trolley
12. Should be supplied with a Printer
13. Screen size minimum 5.5"
14. CE or US FDA certification
15. Should supply 25 rolls each of CTG thermal paper
16. Conducting jelly / paste - 02 nos.
17. Continuous supply of paper roll.
18. The rate of paper roll and other consumables should be quoted in Form 9 (a)