



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

Supply and Commissioning of Medical equipment for the Department of Urology at R.G.Kar MCH
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-206/2022

Dated-19.05.2022

Amendment -II

REVISED TECHNICAL SPECIFICATION

Schedule – V

18-22 FR Nephroscope with all accessories

Specification-

1. The product should be CE (European) and FDA approved.
2. Wide angle straight forward telescope with angled eyepiece, autoclavable, fiber optic light transmission with lock connection for inflow incorporated.
3. Should have operating telescope with offset eyepiece having angle of view between 6 deg. to 30 deg.
4. Should have working length of 18 cm and above.
5. Should have rotatable operating sheath between 20 Fr to 25 Fr compatible with telescope for continuous irrigation and suction.
6. Should have hollow obturator for operating sheath.
7. Scope should have incorporated working channel of 9 Fr and above.
8. The eyepiece of the instrument should be compatible with standard Endovision camera head.
9. Physical demonstration of the scope is mandatory after the completion of tender evaluation.
10. All equipment **European CE (4 Digit Notified body) & USFDA** certified.