

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

SUPPLY AND COMMISSIONING OF FOETAL DOPPLER FOR DIFFERENT HEALTH FACILITIES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT - 09 / 2017

Dated –07.03.2017

AMENDMENT - I:

Section I: Instructions to Tenderers

E. Submission and Opening of Bids

35. The following are to be submitted:

(ii) Statutory Documents

(b) **BID – B** (Should be in multiple page single PDF file)

| | |
|---|--|
| 4 | CE (“Conformite Europeene”) from European Union notified body having 4 digit identification number or US FDA certification . <i>Only product enlistment under US FDA will not be accepted.</i> |
|---|--|

AMENDMENT - II:

Section IV. Schedule of Requirements

1. List of Consignee

| Sl. | Description of Goods | Consignee List | Total Quantity (in number) |
|-----|----------------------|--------------------------------------|-------------------------------|
| 1 | Foetal Doppler | District Family Welfare Bureau Store | 512 |

Please note:

The Consignee Receipt Certificate (CRC) will be issued to the Vendor within 72 hours of the delivery at the Consignee address

2. TECHNICAL SPECIFICATIONS

FOETAL DOPPLER

1. It should be LCD Display (45mm X 22mm or more)
2. It should be ultrasonic emitting frequency: 2 – 2.5 MHz and power: < 5m Watt /cm²
3. Overall sensitivity at the distances 120 - 200 mm from the face of the probe: 90db
4. Automatic FHR test range 50 – 210 BPM or wider (Beat / Minute)
5. Resolution: 1 BPM
6. Audio output : 0.5 – 1 Watt
7. Should be supplied with electric charger & rechargeable battery (Li – ion) with 1 set of spare battery.
8. Good quality water proof case for carrying purpose
9. Good quality water proof probe & gel (5 bottles of 200ml or total quantity of 1000ml) should be supplied with the equipment.
10. Warranty includes the entire machine including probe.
11. Standards, Safety & Training:
 - a) Should be CE (“Conformite Europeene”) from European Union notified body having 4 digit identification number or US FDA certification. Only product enlistment under US FDA will not be accepted.
 - b) Warranty: 2 years; Extended Warranty: 3 year
 - c) On site comprehensive training for field staff and support services till customer satisfaction with the system

12. Documentation:

- a) User / Technical / Maintenance manuals to be supplied in English

NOTE:

1. Bidders should note that all the spares and consumables mentioned pre-page and above in the technical specifications of the Foetal Doppler is a part of scope of supply of the tender.
2. Bidder should quote the rate for the following items in Form 9 (b) for future requirements, if any and is not a part of scope of present supply of the tender:
 - a) Battery
 - b) Probe

The rates quoted for the items in Form 9 (b) should be reasonable and justifiable with the rate quoted for the equipment as per specifications mentioned pre-page and above. The rates for the item quoted in Form 9 (b) will not be part of financial bid evaluation of the tender.

AMENDMENT - III:

Section V. Bidding Forms

Form 2: CHECK-LIST

[Please fill in and include with your Bid]

| Non statutory documents to be submitted under <u>My Document</u> | | | | |
|--|--|-----------|--------------------|--------|
| Sl. No. | Activity | Yes/No/NA | Page No in the Bid | Remark |
| 23 | CE ("Conformite Europeene") from European Union notified body having 4 digit identification number or US FDA certification . <i>Only product enlistment under US FDA will not be accepted.</i> | | | |