



## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Bhawan Complex  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340308/40340319

E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

SUPPLY OF FULL HD DIGITAL VIDEO COLPOSCOPE & CRYO SURGICAL SYSTEM USING NITROUS OXIDE IN THE HOSPITALS OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-42/2019

Dated -04.06.2019

The following amendment has been made in the tender document,

### SCHEDULE – II

#### Amendment – I

(In Page: 35)

#### Revised Specification Cryo Surgical System using Nitrous Oxide (N<sub>2</sub>O)

1. The Cryo Surgical system should consist of the following:
  - a) Made with high quality Hypodermic needles-ensures smooth flow of cryogenic gas.
  - b) Equipped with “Silencer’ enabled quieter operation
  - c) Body made of ABS plastic
2. Should provide following interchangeable probes of 4 different sizes:
  - a) Endo-Exo cervical probe
  - b) Exo cervical probe
  - c) Deep Endo-Exo cervical probe

3. "ON- OFF control" enables cut off the gas line (flow) without closing the cylinder during interchange the probes
4. CO<sub>2</sub> gas compatibility- enable removal of any clog and cleaning the inner hypodermic needle.
5. The system should be supplied with followings:
  - a) Carrying Case
  - b) N<sub>2</sub>O cylinder with 5 ltr capacity for easy portability
  - c) Universal PIN-INDEX for easy connectivity
  - d) Pressure gauge should be present attached to the cylinder.
6. 2 years warranty