

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/40340308

E mail: procurement@wbmsc.gov.in

SUPPLY OF MEDICAL EQUIPMENTS FOR MULTI / SUPER SPECIALITY HOSPITALS OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT- 23/2016

Dated – 12.05.2016

The following amendments have been made in the tender document,

Amendment - II:

SCHEDULE – III

Trial lens sets with frames

1. Should be supplied with set of 226 trial lens with durable metal frame

Technical Specification

2. Lens set

3. Set of 226 trial lenses with polished metal rims in an elegant carrying case.
4. Minimum 64 Spherical Concave Lenses
5. Minimum 64 Spherical Convex Lenses
6. Minimum 40 Cylindrical Concave Lenses: 0.12 to 6.00 Diopters (0.25D increments)
7. Minimum 40 Cylindrical Convex Lenses: 0.12 to 6.00 Diopters (0.25D increments)
8. Prismatical Lenses: 0.5(1), 1(1), 2(1), 3, 4, 5, 6, 8, 10, 12
9. Accessory Lenses: Red, Green, Slit, Dark, Pin hole (1) Block (2), Maddox rod (1), Plain glass (1)

Metal Frame

10. Should be durable and light weight (Adult and Paediatric)
11. Number of lenses which can be inserted in to left or right lens frame simultaneously : 4 pieces
12. Degree of lens rotation around optical axis in the lens frame:360 degree

- 13 Horizontal and vertical bridge adjustments.
- 14 Free-rocking saddle bridge assures comfortable nose fit.
Range of nose rest adjustment Length:0 ~ 14 mm,angle:0 ~ 30 degree
- 15 Separate PD adjustment for each eye to compensate for asymmetrical factors in facial structure.
 - a) PD of both eye: 48 ~ 80 mm
 - b) Left or right PD: 24 ~ 40 mm
- 16 Scales with large, easy-to-read numerals.