



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

PROCUREMENT OF MEDICAL EQUIPMENT FOR MEDICAL COLLEGES AND HOSPITALS
(GROUP 3)
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-52/2019

Dated –18.06.2019

(2nd call of Bid Reference No.: WBMSCL /NIT-217/2018, Dated –12.10.2018)

The following amendment has been made in the tender document,

Amendment – 1

Schedule XI

CYTO CENTRIFUGE MACHINE	
PARAMETER / FEATURE	THRESHOLD LIMIT/ RANGE
Maximum programmable acceleration speed in RPM	2000 or more
Machine throughput (no. of tests) per hour	12 or more
Number of samples can be run at a time	12 or more
MUST HAVE SPECIFICATIONS	
PARAMETER / FEATURE	YES /NO
LCD/LED Display showing RPM & time	
Programmable centrifuge operation	
Rotor should be autoclavable with lid lock system	
Display during the run - the actual speed and the time remaining	
QUALITY STANDARD AND SAFETY CERTIFICATION	
European CE or US FDA	