



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

SUPPLY AND COMMISSIONING OF MEDICAL EQUIPMENTS (SET-III) FOR TRAUMA CARE FACILITY FOR THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVT. OF WEST BENGAL

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT- 66/2017

Dated – 23.09.2017

2nd call of Bid reference no. WBMSCL/NIT-21/2017, dated: 26.04.2017 of Schedule –I, II, III, IV & V

SCHEDULE – IV

The following amendments have been made in the tender document,

Amendment – IV

Portable Ventilator

1. Should be small and portable
2. Should operate on built in piston or turbine assembly should not use external compressor or blower.
3. Ventilator should be useful for adult, pediatric & infant patients with in-built colour touch screen of 8" or more. Display facility of loops and curves should be available.
4. **Modes of Ventilation**
 - 4.1. Volume Controlled Ventilation
 - 4.1.1. Assist / Controlled / SIMV

4.2. Pressure Controlled

4.2.1. Assist / Controlled / SIMV

4.3. CPAP/PSV

4.4. Noninvasive

5. Should have back-up ventilation in all modes of ventilation with the facility to link the backup ventilation to Apnea alarm or Low Minute Ventilation Alarm or to both.
6. Should have PEEP compensated Inspiratory trigger in pressure (sensitivity).
7. The unit should be able to provide the following parameters: Tidal volume: 30 ml to 2000ml (and beyond both limits) / Respiratory rate: Up to 99 bpm or more / PEEP / CPAP: 0 to 30 Cm H₂O / Pressure support: Up to 40 cm H₂O or more, / Pressure Control: Up to 60 cm H₂O or more/I:E ratio: 1:99 to 3:1/FiO₂: 21 % to 100% with FiO₂ monitoring
8. It should have the following alarms with audio, visual and with message display: P max, P min, High & Low Baseline pressure, High & Low Minute Volume, Apnea, Back up ventilation alert, Device alert, Power switch over alert, Low battery, low & high FiO₂.
9. Should have monitored data of: Inspired tidal volume, Inspired Minute Volume, Total Total respiratory rate Peak / Maximum airway pressure. Mean airway pressure, PEEP and FiO₂
10. Patient disconnection, power failure alarms
11. Manual Inflation upto 2-3 seconds
12. Should have 100 L / min or more peak flow
13. Battery should be rechargeable, hot swappable and life at least 8 hrs of operation or more at, normal adult ventilation settings.
14. Should have CE and US FDA approval.
15. Should also operate in Air ambulances & helicopter and should have Airworthy certificate.
16. Should have facility to change the battery (hot swappable battery)
17. Should have user selectable monitored data.
18. US FDA and European CE ("Conformité Européene") approved Inbuilt / integrated nebulizer.