

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/20
E mail: procurement@wbmsc.gov.in

SUPPLY & INSTALLATION OF 3.0 TESLA MRI MACHINE IN THE MEDICAL COLLEGE AND HOSPITAL OF GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT- 184 /2018

Dated –13.09.2018

1. West Bengal Medical Services Corporation Limited (WBMSCL) has been requested by the Government of West Bengal to procure on their behalf **3.0 Tesla MRI Machine** to be supplied to various healthcare establishments of the Govt. of West Bengal.
2. WBMSCL hereby invites bids from eligible and qualified Tenderers for the supply of **3.0 Tesla MRI MACHINE** as per Schedule of Requirement.
3. Intending Tenderer may download the tender documents from the website: **wbtenders.gov.in**. Earnest money to be drawn in favour of 'West Bengal Medical Services Corporation Limited' through online.
4. Non statutory documents, Bid – A, Bid – B & Bid – C are to be submitted concurrently.

Sd/-
Managing Director,

Table for Important Dates

Sl.	Items	Publishing date(s)
1.	Date of uploading of N.I.T. Documents (online)/ Date of Issue	19.09. 2018
2.	Documents download start date (Online)	19.09. 2018
3.	Date of Pre bid Meeting with the intending Tenderers in the Conference Hall of West Bengal Medical Services Corporation Limited	26.09.2018 at 11:00 AM
4.	Bid submission start date (On line)	28.09.2018
5.	Documents download end date (Online)	NA
6.	<p>Bid submission closing (On line) Bid submission includes: i) Non statutory documents to be submitted under <u>My Space</u> (Each sub-category item should be in multiple page single PDF file) ii) BID – A (Should be in multiple page single PDF file) iii) BID – B (Should be in multiple page single PDF file) iv) BID – C (BOQ and the Statement of Breakup of Duties and Taxes & Prices of consumables, spares etc.)</p> <p>Detailed list of documents annexed at Section V Check-List Form</p> <p>Non-statutory document (document uploaded in <u>My Space</u>), Bid – A & Bid – B constitute the technical bid and Bid – C is the financial bid.</p> <p><i>Any wrong or misleading information provided by the Tenderer during submission of bids shall lead to summary cancellation of bid and may lead to blacklisting in WBMSCL for at least 5 years.</i></p> <p>Each scanned documents should have an index page indicating the name of the documents enclosed with Page no.</p>	10.10.2018; up to 5:00 p.m.
7.	Last date of submission of all hard copies of the documents uploaded in e-tender during bid submission except BOQ at the registered office of WBMSCL in envelope. The hard copies of the uploaded documents (except BOQ) arranged in the same sequence as given in the Check List and copy of acknowledgement generated by e-tender portal against the documents uploaded during bid submission. All the documents should be appropriately flagged.	12.10.2018; up to 5:00 p.m.
8.	Bid opening date for Technical Proposals (Online)(Bid A & B)	12.10.2018; after 5:00 p.m.
9.	Bidders to remain present at WBMSCL office, Kolkata for identification of the documents for the technical bid evaluation	To be notified later
10.	Functional demonstration of the equipment	To be notified later
11.	Submission of non-statutory Wanting document (if any)	To be notified later
12.	Opening of Financial Bid	To be notified later

Section I: Instructions to Tenderers

A. Important information at a glance

Whenever there is a conflict, the provisions herein (i.e. in Important Information at a glance) shall prevail over those in the General Conditions of Contract (GCC) & Special Conditions of Contract (SCC)

(The item suffixed by "E" in bracket indicates Eligibility Criteria for a bidder)

1. Tender Details

Item	Qty	Warranty	Comprehensive Maintenance Contract (CMC)
3.0 TESLA MRI MACHINE	2	2 Year	8 Years

2. Tender Fees: Exempted

3. Earnest Money Deposit (EMD) (E)

Item	Amount In Lakh INR	Instrument
3.0 TESLA MRI MACHINE	30.00	Online

4. Annual Turnover Requirements: (E)

The Tenderers should have annual sales turnover (i.e. total turnover of the company) of minimum on an average of last three financial years (2014-15, 2015-16 and 2016-17) as per the Audited Accounts of the Organization as mentioned in the table below:

Item	ANNUAL TURNOVER in Crore Rs.
3.0 TESLA MRI MACHINE	25.0

5. (a) Time for Supplies & Commissioning of work from the date of issuance of Award of Contract

Item	Timeline in days
3.0 TESLA MRI MACHINE	90

5. (b) Payment Terms

3.0 TESLA MRI MACHINE
<p>I. General Terms</p> <p>(i) The payment to manufacturing company or its subsidiary in India will be made under Delivered Duty Paid contract.</p> <p>(ii) The Tenderers should only quote in INR.</p> <p>II. Payment terms for Manufacturer/Indian Distributor</p> <p>A. 80 % of the Basic Price of the Equipment along with applicable GST shall be paid after delivery and furnishing of Consignee Receipt Certificate (CRC) as per Format 3a.</p> <p>B. Remaining 20% of the Basic Price of the Equipment along with applicable GST shall be paid after successful installation, commissioning and initial end user training by the authorised service team of the supplier and furnishing of Satisfactory Installation Certificate (SIC) as per Format 3b. The SIC shall be signed by the facility / end user or issued after joint certification by authorised official(s) from WBMSCL and the facility/</p>

end user.

However, WBMSCL will furnish to the seller Bank Guarantee to be issued by United Bank of India, Sector-V, and Salt Lake Branch for payment of the price of the goods on its delivery to the WBMSCL at the place of installation.

Note:

Submission of required Performance Bank Guarantee and signing of Agreement are mandatory for the processing of any Payment.

Basic Price of Equipment(s) includes value of goods, accessories & ancillaries, freight charges, installation, commissioning, end user training as many number of times as required during the period of warranty and any other charges as applicable excluding GST. Applicable GST will be paid extra.

6. Performance Security (PS)

(In the form of unconditional and irrevocable Bank Guarantee)

3.0 TESLA MRI MACHINE
10% of the Bid Value (Validity should be till the completion of Warranty + 60 days).

7. Who can Bid (E)

3.0 TESLA MRI MACHINE
a. Manufacturing Company
b. Manufacturer's subsidiary in India
c. Manufacturer's sole Authorized Distributor / Business Partner, in case the manufacturing Company do not have registered office / commercial establishment in India.
Bidders would be only be eligible, if they agree to remain responsible for providing Comprehensive Maintenance Services (including all spares) for the entire useful life of the Equipment after expiry of the Warranty Period

8. Service Up time in Warranty & CMC

3.0 TESLA MRI MACHINE
Working condition for a minimum period of 354 days out of a period of 365 days.(i.e. 97% uptime)
The response time to any fault should be not more than 6 hrs. Time for rectification should not be more than 24 hours
Maximum Downtime allowed without penalty : 48 hours
Penalty beyond 12 hours of non physical attendance: Rs 25,000 per 24 hour per machine Penalty for Downtime after 48hrs: Rs 50,000 per 24 hour per machine

9. Liquidated damages for Delayed Delivery / Delayed setting up of Services

3.0 TESLA MRI MACHINE
The percentage of 0.5% of the Invoice price for each week or part thereof, of delay until actual delivery or performance, up to a maximum deduction of 5% of the Invoice price.

10. Experience and Technical Capacity (E)

3.0 TESLA MRI MACHINE
The manufacturer whose product is offered by the Tenderer must have commissioned at least 3 (three) units of the offered model which are functioning satisfactorily as on the date (s) of physical inspection & functional/working demonstration. At least 1(one) machine out of the 3(three) should be installed in India.
The offered equipment must be from manufacturing unit having the ISO quality certification. The certificate must be valid and issued by an internationally recognised inspection company.
Tenderers shall invariably furnish documentary evidence / Client's certificate of at least 3 (three) users of the quoted model in support of the satisfactory operation. At least 1(one) certificate out of the 3(three) should be from healthcare facility in India.
Confirmation from manufacturer of the offered equipment that all the facilities exist in his factory for inspection and testing and these will be made available to WBMSCL or his representative, if inspection is considered necessary.

11. Preference for S.S.I. units registered in West Bengal & PSUs in West Bengal:

Preference will be given to the S.S.I. units registered in West Bengal & PSUs in West Bengal and State Based Other Manufacturers as per West Bengal Financial Rule incorporated under notification No. 10500-F dated 19.11.04 as amended hereafter.

Note1: Warranty and CMC includes the equipment, all accessories and ancillaries as given in the specification of the equipment and all the accessories and ancillaries required for the site preparation and interiors (if included in the particular schedule of requirement).

Note2: The Tenderers, who have downloaded the bid documents, shall be solely responsible for checking these websites for any amendment, addendum issued subsequently to the bid document and takes into consideration the same while preparing and submitting the bids.

Bids will be opened in the presence of Tenderers' representative who chooses to attend on the specified date and time. However, opening of bids will not be stopped for absence of any bidder or his authorized representative at the notified time.

Note3: In the case of a Tenderer offering to supply Goods under the Contract that the Tenderer does not manufacture or otherwise produce, the Tenderer should be duly authorized by the manufacturer of the Goods who meets the criteria under

- (1) Above (all supporting documents / information as asked above for manufacturer shall be submitted with the bid) and
 - a) The manufacturer furnishes a legally enforceable authorization in the prescribed Form assuring full guarantee and warranty obligations as per GCC and SCC for the goods offered; and
 - b) The Tenderer, as authorized by the manufacturer, must have supplied and provided after sales service of similar model in the Schedule of Requirements in any one of the institutions within India in the last five (5) years, which must be in satisfactory operation at least for one year on the date of bid opening.

Note 4: Service Centres

- Preferably at Kolkata, Siliguri and Durgapur

Note 5: In case extension required, Bank Guarantees are to be renewed prior to 30 days of their expiry.

12. General Instructions

- a) Bidders are requested to study the tender document, terms & conditions carefully before submitting their bids. Submission of tender shall be deemed to have been done after careful study and examination of the tender document with full understanding of its implications.
- b) Tender documents should be downloaded from the E-tender portal of Govt. of West Bengal at wbtenders.gov.in and the website of WBMSCL at www.wbmsc.gov.in. The submission of bids should only be through online at wbtenders.gov.in.
- c) All pages of the bid submitted must be signed and sequentially numbered by the Bidder. All information in the offer must be in English. Information in any other language must be translated to English. Failure to comply with this may render the offer liable to be rejected. In the event of any discrepancy between the offer in a language other than English and its English translation, the English translation will prevail.

B. General

11. Scope of Bid

- 11.1 The type of goods and related services to be purchased is: **Supply and Commissioning of 3.0 TESLA MRI MACHINE in the hospitals of the Govt. of West Bengal** as per the Schedule of Requirements.

12. Source of Funds

- 12.1 Funds received from the **Department of H & FW**, for the procurement of Medical Equipments on behalf of the **Department of H & FW**.

13. Fraud and Corruption

- 13.1 It is WBMSCL policy to require that Tenderers, suppliers and contractors and their subcontractors under WBMSCL contracts, observe the highest standard of ethics during the procurement and execution of such contracts. In pursuance of this policy, WBMSCL:
- (a) Defines, for the purposes of this provision, the terms set forth below as follows:
 - (i) Bribery is the act of unduly offering, giving, receiving or soliciting anything of value to influence the process of procuring goods or services, or executing contracts;
 - (ii) Extortion or coercion is the act of attempting to influence the process of procuring goods or services, or executing contracts by means of threat of injury to person, property or reputation;
 - (iii) Fraud is the misrepresentation of information or facts for the purpose of influencing the process of procuring goods or services, or executing the contracts, to the detriment of WBMSCL or other participants;
 - (iv) Collusion is the agreement between Tenderers designed to result in bids at artificial prices that are not competitive.
 - (b) Will reject a proposal to award a contract if it determines that a vendor recommended for award has engaged in corrupt practices in competing for the contract in question;
 - (c) Will declare a vendor ineligible, either indefinitely or for a stated period of time, to become a WBMSCL registered Vendor if it at any time determines that the vendor has engaged in corrupt practices in competing for or in executing a WBMSCL contract;
 - (d) Will cancel or terminate a contract if it determines that a vendor has engaged in corrupt practices in competing for or in executing a WBMSCL contract;
 - (e) Will normally requires a WBMSCL vendor to allow WBMSCL, or any person that WBMSCL may designate, to inspect or carry out audits of the vendor's accounting records and financial statements in connection with the contract.
- 13.2 Any vendor participating in WBMSCL's procurement activities, shall facilitate to WBMSCL personnel upon first request, all documents, records and other elements needed by WBMSCL to investigate the

allegations of misconduct by either vendors or any other party to the procurement activities. The absence of such cooperation may be sufficient grounds for the debarment of the vendor from WBMSCL vendor roster and may lead to suspension following review by WBMSCL Vendor Review Committee.

- 13.3 It is required that Vendors, their subsidiaries, agents, intermediaries and principals cooperate with WBMSCL Internal Audit Group as well as with other investigations authorized by WBMSCL or by the Government of West Bengal or the Central Government as and when required. Such cooperation shall include, but not be limited to, the following: access to all employees, representatives, agents and assignees of the vendor; as well as production of all documents requested, including financial records. Failure to fully cooperate with investigations will be considered sufficient grounds to allow WBMSCL to repudiate and terminate the contract and to debar and remove the supplier from WBMSCL's list of registered vendors.

14. Eligible Tenderers

- 14.1 A Tenderer and all parties constituting the Tenderer may have the nationality of any country.
- 14.2 A Tenderer shall not have a conflict of interest. All Tenderers found to have conflict of interest shall be disqualified. Tenderers may be considered to have a conflict of interest with one or more parties in this bidding process, if they:
- ii) are or have been associated in the past, with a firm or any of its affiliates which have been engaged by WBMSCL to provide consulting services for the preparation of the design, specifications and other documents to be used for the procurement of the Goods to be purchased under these Bidding Documents ; or
- 14.3 A Tenderer that is under a declaration of ineligibility by WBMSCL in accordance with Instructions to Tenderers at the date of contract award, shall be disqualified. Tenderers shall not be eligible to submit a bid when at the time of bid submission:
- Suppliers are already suspended by WBMSCL; or,
 - Suppliers are suspended by the Government of West Bengal or Central Government or any other State Government or WBMSCL,
 - Suppliers have been declared ineligible by Government of West Bengal or Central Government or any other State Government or WBMSCL.

15. Eligible goods and related services

- 15.1 All the goods and related services to be supplied under the Contract may have their origin in any country.
- 15.2 For purposes of this Clause, the term "origin" means the country where the goods have been mined, grown, cultivated, produced, manufactured or processed; or, through manufacture, processing, or assembly, another commercially recognized article results that differs substantially in its basic characteristics from its components.

C. Contents of Bidding Documents

16. Sections of Bidding Documents

16.1 The Bidding Documents consist of:

- Section I. Instructions to Tenderers
- Section II. General Conditions for Goods (GCG)
- Section III. Special Conditions of Contract (SCC)
- Section IV. Schedule of Requirements
- Section V. Bidding Forms
- Section VI. Contract Forms

16.2 The Tenderer is expected to examine all instructions, forms, terms and specifications in the Bidding Documents. Failure to furnish all information or documentation required by the Bidding Documents may result in the rejection of the bid.

16.3 Tenderers are cautioned to read the specifications carefully (see Section IV - Schedule of Requirements), as there may be special requirements. The technical specifications presented herein are not to be construed as defining a particular manufacturer's product. Tenderers are encouraged to advise WBMSCL, if they disagree.

16.4 The specifications are the minimum requirements for the products. Products offered must meet or exceed requirements mentioned in technical specifications. The products shall conform in strength, quality and workmanship to the accepted standards of the relevant industry. Modifications of or additions to basic standard products of less size or capability to meet these requirements will not be acceptable.

17. Clarification of Bid Document

17.1 A prospective Tenderer requiring any clarification of the Bidding Documents shall contact WBMSCL in writing at procurement@wbmsc.gov.in

18. Amendment of Bid Document

18.1 At any time prior to the deadline for submission of bids, WBMSCL may amend the Bid Document by issuing amendment to be uploaded in the e-tender portal & website of WBMSCL.

18.2 To give prospective Tenderers reasonable time in which to take an amendment into account in preparing their bids, WBMSCL may, at its discretion, extend the deadline for the submission of bids.

D. Preparation of Bids

19. Tenderers are to prepare and submit the following:
- i) Non statutory documents to be submitted under My Space
 - ii) BID – A (Should be in multiple page single PDF file)
 - iii) BID – B (Should be in multiple page single PDF file)
 - iv) BID – C (BOQ, the Statement of Breakup of Taxes & Duties and Cost of Consumables & Spares)
Details are given in “Submission and Opening of Bids”
20. **Cost of Bidding**
The Tenderer shall bear all costs associated with the preparation and submission of its bid.
21. **Language of Bid**
The Bid, as well as all correspondence and documents relating to the bid exchanged by the Tenderer and WBMSCL, shall be written in the English language. Supporting documents and printed literature that are part of the Bid, if submitted in any other language provided they are accompanied by an accurate translation of the relevant passages into English and duly authenticated.
22. **Alternative Bids**
Alternative Bids will not be accepted.
23. **Bid Prices**
- 23.1 The prices in the BOQ shall conform to the requirements as specified in the tender.
 - 23.2 The Incoterms shall be governed by the rules prescribed in the Incoterms 2010, published by The International Chamber of Commerce.
 - 23.3 Prices quoted by the Tenderer shall be fixed during the Tenderer’s performance of the Contract and not subject to variation on any account.
24. **Currencies of Bid**
The Tenderer shall quote
- (i) The rate of the equipment in INR only.
 - (ii) Cost of site Preparation including interiors in INR only
 - (iii) Cost of year wise CMC charges in INR only.
25. **Documents Establishing the Conformity of the Goods and Related Services**
To establish the conformity of the goods and related services to the Bidding Documents, the Tenderer shall furnish as part of its Bid the documentary evidence that the Goods conform to the technical specifications and standards specified in Section IV, Schedule of Requirements.
26. **Documents Establishing the Qualifications of the Tenderer**
As per Form 1: Check-List of Section V of the bid document.
27. **Period of validity of Bids**
- 27.1 Bids shall remain valid for a period of **2 year** from last date of online submission of bid. A bid valid for a shorter period than specified in previous lines shall be rejected by WBMSCL as non responsive.
 - 27.2 In exceptional circumstances, prior to the expiration of the bid validity period, WBMSCL may request Tenderer to extend the period of validity of their bids and EMD. In the event of the request for such extension beyond bid validity period, the bidder may or may not

accept such request. In case, the bidder refuses to accept the request, the EMD of the bidder shall not be forfeited.

28. Period of validity of the bid price:

The bid price shall remain valid for a period of **2 (two)** years from the date of signing of the contract.

29. Earnest Money Deposit (EMD):

29.1 The EMD shall be through online from any scheduled commercial bank as per prescribed form in Section V, in favour of "**West Bengal Medical Services Corporation Limited**", GN-29, Swasthya Sathi, Sec-V, Salt Lake, Kolkata- 700091 in the amount as provided in the **Schedule of Requirements** and denominated in INR.

29.2 Any bid not accompanied by a substantially responsive EMD in accordance with Instructions to Tenderers shall be rejected by WBMSCL as non-responsive.

29.3 The EMD of unsuccessful Tenderers shall be returned as promptly as possible upon the successful Tenderer's furnishing of the Performance Security pursuant to Instructions to Tenderers

29.4 The EMD may be forfeited:

- (a) if a Tenderer withdraws its bid during the period of bid validity specified by the Tenderer on the Bid Submission Form, except as provided in Instructions to Tenderers; or
- (b) if the successful Tenderer fails to:
 - (i) sign the Contract in accordance with Instructions to Tenderers;
 - (ii) furnish a Performance Security in accordance with Instructions to Tenderers ;

30. Signing of Bid – The bid document should be digitally signed and uploaded on the E-tender portal.

31. Withdrawal, Substitution and Modification of Bids

31.1 The bid once submitted cannot be withdrawn but prior to the deadline prescribed for submission of bids, a Tenderer may substitute, or modify its Bid after it has been submitted.

31.2 The objective of this bid is to ensure supply of best quality equipment at the most competitive price. If at any stage of the bidding, including at the stage of financial evaluation, it appears that the tendered rate is artificially hiked or is much lower compared to the prevailing market price and available rates of similar or identical composition with the government, WBMSCL reserves the right to cancel the bids.

32. Confidentiality

Any effort by a Tenderer to influence WBMSCL in the examination, evaluation, comparison of the bids or contract award decisions may result in the rejection of its Bid.

Notwithstanding Instructions to Tenderers from the time of bid opening to the time of Contract Award, if any Tenderer wishes to contact WBMSCL on any matter related to the bidding process, it should do so in writing.

E. Submission and Opening of Bids

33. The following are to be submitted:

i) Non statutory documents to be submitted under My Document

(Each sub-category item should be in multiple page single PDF file)

Guidelines for uploading documents in My Document:

Sl. No.	Category Name	Sub - Category Name	Document Name
1	CERTIFICATES	CERTIFICATES	a) PAN Card b) 15 – digit Goods and Services Taxpayer Identification Number (GSTIN)
2	COMPANY DETAILS	COMPANY DETAILS 1	a) License from Government / Statutory Authority as applicable. OR b) Registration with the Registrar of Companies, if applicable.
		COMPANY DETAILS 2	a) Manufacturing Licence (National/International). (In case, manufacturing licence is not required / applicable for production of the quoted item, notarized declaration from the manufacturer is to be submitted) b) Current registration as SSI (if any)
3	CREDENTIAL	CREDENTIAL 1	Performance Statement Form (For the period of last three calendar years ending December 2016) - Form 6 of Section V
		CREDENTIAL 2	ISO certificate of the manufacturing factory
4	DECLARATION	DECLARATION1	Income Tax returns for assessment year (2014-15, 2015-16 & 2016-17) or (2015-16, 2016-17 & 2017-18)
		DECLARATION2	
		DECLARATION3	
		DECLARATION 4	Legal declaration affirmed before a First Class Magistrate / Notary on non judicial stamp paper of Rs. 100/- for, a) Acceptance of Terms and Conditions of NIT and its Amendments and Addendums thereto. (As per Form 1 of Section V). Note: Technical evaluation of the bid will be taken up only after scrutiny of Form 1 (NIT Acceptance Form) duly notarized.

		DECLARATION 5	Legal declaration affirmed before a First Class Magistrate / Notary on non judicial stamp paper of Rs. 100/- as per Form – 7.
5	EQUIPMENT	MACHINERIES 1	Manufacturer's Authorization (If applicable) as per Form 5 of Section V
		MACHINERIES 2	List of installations & commissioning of offered model. (self declaration)
		MACHINERIES 3	Satisfactory Performance Certificate from at least 3 (three) users of the quoted model in support of the satisfactory operation in India.
6	FINANCIAL INFO	P/L & BALANCE SHEET 2013-2014	P/L & Balance sheet 2014-2015
		P/L & BALANCE SHEET 2014-2015	P/L & Balance sheet 2015-2016
		P/L & BALANCE SHEET 2015-2016	P/L & Balance sheet 2016-2017

(ii) Statutory Documents

(a) **BID – A** (Should be in multiple page single PDF file)

1. EMD (Scanned copy of the instrument through which EMD have been submitted)
2. Declaration of the bidder on letter head that "We agree to submit a copy of the Tender Documents and its Amendments and Addendums thereto duly initialled by us in all pages with our seal/ rubber stamp affixed thereto, in token of acceptance thereof."

(b) **BID – B** (Should be in multiple page single PDF file)

1	Model of the equipment offered for (Self Declaration) with Technical Data Sheet
2	Compliance Sheet of the Technical Specifications (Form No. 3 of Section V) duly signed & stamped by Authorized signatory of the Manufacturing company
3	2 sets of Brochure of the offered product / model.
4	CE ("Conformité Européene") Certificate & US FDA approval Certificate of the offered model (as applicable) <small>CE ("Conformité Européene") Certificate should be from EU Notified Bodies authorized to conduct audits</small>
5	Self declaration by the Tenderer for agreeing that WBMSCL will do Quality Check by Third Party Agency, If required
7	Confirmation from manufacturer of the offered equipment that all the facilities exist in his factory for inspection and testing and these will be made available to WBMSCL or his representative, if inspection is considered necessary.
8	Pre-requisites of installation like [Power (KVA, Phase, Hz) and any other requirement, if any].
9	Average Annual Turnover of the Company in medical equipment division during the last 3 Financial Years (2014-15, 2015-16 & 2016-17) / (2015-16, 2016-17 & 2017-18) (in INR) - to be certified by practicing Chartered Accountant as per format given in FORM 11
10	Form 12: Declaration of Quality Certification of Equipment (as applicable)

Non-statutory document (document uploaded in My Space), Bid – A & Bid – B constitute the technical bid

iii) **BID – C [Bill of Quantity (BOQ), the Statement of Breakup of Duties and Taxes and Cost of Consumables & Spares]**

BOQ shall contain the financial quotes in respect of (a) **Basic Price of Equipment** (b) **Cost of site preparation including interiors etc.**

- (a) **Basic Price of Equipment (BP):** includes value of goods, accessories & ancillaries, duties, freight charges, insurance, installation, commissioning, end user training as many number of times as required during the period of warranty and any other charges as applicable excluding applicable GST.
- (b) **Cost for site preparation including interiors etc.** excluding applicable GST

Comparison of Financial Bids would be based on the sum of ‘a’ & ‘b’ taken together. The rates quoted shall be firm and no variation will be allowed during the period of contract.

iv) **Comprehensive Maintenance Contract (CMC):** The selected bidder shall enter into CMC after expiry of warranty and CMC should include the following:

- a) The equipment including all other accessories and ancillaries as given in the specifications of the equipment including UPS, UPS Battery, refilling of Helium etc.
- b) All the accessories and ancillaries including Air conditioning machine required for the site Preparation and interiors

The equipment wise list for preventive maintenance is given in the table below,

Equipments	Mandatory preventive Maintenance Service (PMS) visit per year		Remarks
	Warranty	CMC	
3.0 TESLA MRI MACHINE	4	4	1. Supplier/ authorised service provider must attend all breakdown calls during warranty / CMC period. 2. The supplier should provide Preventive Maintenance Services (PMS) yearly / bi - yearly / quarterly in equal interval to fulfill the minimum number of mandatory PMS as recommended in the previous column during warranty / CMC period. 3. The preventive maintenance includes testing & calibration as per technical / service / operational manual, spares, all software updates and labour.

The cost for Year wise CMC charges after completion of 2 years warranty in percentage of the sum of quoted price of the equipment and Turnkey work in the BOQ to be paid to the selected bidder(s) in the following manner:

ITEM	CMC charges in percentage(%) for Year1	CMC charges in percentage(%) for Year2	CMC charges in percentage(%) for Year3	CMC charges in percentage(%) for Year4	CMC charges in percentage(%) for Year5	CMC charges in percentage(%) for Year6	CMC charges in percentage(%) for Year7	CMC charges in percentage(%) for Year8
3.0 TESLA MRI MACHINE	3.0	3.2	3.4	3.6	3.8	4.0	4.2	4.4

The Tenderer should upload the following statements in PDF in addition to BOQ in .xls

- (i) Breakup for Duties and Taxes (as per Form 8 of Sec V) for equipment.
- (ii) Cost of Consumables & Spares as per Form 10(a) and 10(b).

Detailed list of documents annexed at Form 1 Check-List Form, Section V

N.B.

- a) Any wrong or misleading information provided by the Tenderer during submission of bids may lead to summary cancellation of bid, blacklisting in WBMSCL for at least 5 years and forfeiture of EMD.
- b) Each scanned documents should have an index page indicating the name of the documents enclosed with Page no.
- c) The Earnest Money will be refunded within 15 days after finalization of the tender and / or submission of Performance Bank Guarantee.

F. Evaluation and Comparison of Bids

34. Evaluation of Bids

(A) Technical Evaluation

During the tender evaluation process **Non-statutory document** (document uploaded in My Space), **Bid – A & Bid – B** constituting the technical bid will be opened first and evaluated. The determination of Technical qualification status of a bidder will be based on the following:

- i) **Scrutiny of Form 1 (NIT Acceptance Form) duly notarized**
- ii) **Scrutiny of documentary evidence as per Form 2: Check-list, Section V of Bid document submitted by the Tenderers**
- iii) **Evaluation of Spec indicator:**

Bidders should arrange for an onsite functional demonstration of the offered equipment on the notified date at facility where installed in India. The evaluation of the spec indicator will be made based on the reports of the onsite functional demonstration of the equipments. The onsite functional demonstration of the equipment is purely at the discretion of the Technical Bid Evaluation Committee and its input shall be treated as only corroborative in nature and will not be a substitute for technical evaluation of the document submitted along with the bid. The decision of the Technical Committee in this regard will be final.

Bidder has to comply with all parameters of the technical specifications except deviation(s) which will be considered minor and acceptable by the team of Experts to be engaged by WBMSCL to take working / functional demonstration of the offered equipments.

A bidder will be considered technically qualified if,

1. **Comply with i) & ii) and qualify in iii) above**

B. Financial Evaluation

Financial Bids (Bid - C) of the technically qualified Bidders would only be opened. **Comparison of Financial Bids would be based on the sum of 'a' & 'b' taken together as mentioned in "Submission and Opening of Bids" quoted by the tenderers.**

THE DECISION OF THE WBMSCL AUTHORITY WILL BE FINAL AND BINDING IN THIS MATTER.

35. Responsiveness of Bids

- 35.1 WBMSCL's determination of a bid's responsiveness is to be based on the contents of the bid itself.
- 35.2 A substantially responsive Bid is one that conforms to all the terms, conditions and specifications of the Bidding Documents without material deviation, reservation, or omission. A material deviation, reservation, or omission is one that:
 - (a) Affects in any substantial way the scope, quality, or performance of the goods and related services specified in the Contract; or
 - (b) Limits in any substantial way, inconsistent with the Bidding Documents, WBMSCL's rights or the Tenderer's obligations under the Contract; or
 - (c) If rectified would unfairly affect the competitive position of other Tenderers presenting substantially responsive bids.

WBMSCL considers material deviation to include but not to be limited to the following situations:

(d) During technical evaluation of bids (verification of formal criteria):

- Lack of proper bid securities in terms of change in the wording (not consistent with the prescribed format), amount or validity period.
- Absence of bid form, change in the wording (not consistent with the prescribed format) or lack of signature in the key portions of the bid form,
- The Tenderer does not accept important Contract conditions, i.e. related to Performance Security, Warranty, Force Majeure, Applicable Law, Delivery Schedule, Payment Terms, Limitation of Liability, etc.
- Specifications of the item quoted vary in one or more significant respect(s) from the minimum required Technical Specifications.

35.3 If a bid is not substantially responsive to the Bidding Documents, it shall be rejected by WBMSCL

36. Examination of Terms and Conditions and Technical Evaluation

36.1 WBMSCL shall examine the Bid to confirm that it does not contain material deviation or reservation related to the conditions and requirements specified in the GCC of Section II, SCC of Section III and in the Schedule of Requirements of Section IV.

36.2 If, after the examination of the terms and conditions and the technical evaluation, WBMSCL determines that the Bid is not substantially responsive in accordance with Instructions to Tenderers, it shall reject the Bid.

37. Domestic Preference

Companies based and manufacturing in India shall not be a factor in bid evaluation.

38. Financial evaluation

The financial bids of only those bidders who qualify in the technical bids will be opened.

39. WBMSCL's Right to Accept Any Bid and to Reject Any or All Bids

WBMSCL reserves the right to accept or reject any bid and to annul the bidding process and reject all bids at any time prior to contract award, without thereby incurring any liability to Tenderers.

G. Award of Contract

41. Award Criteria

- 41.1 In the event of a Contract award, WBMSCL shall award the Contract to the Tenderer whose offer has been determined to be the lowest evaluated bid substantially responsive to the Bidding Documents, provided further that the Tenderer is determined to be qualified to perform the Contract satisfactorily.
- 41.2 Before the award of Contract, WBMSCL may inspect the manufacturing facilities of the lowest evaluated responsive Tenderer to assess his capability to successfully perform the Contract as per the terms and conditions specified in the NIT.

42. WBMSCL's right to vary quantities

- 42.1 WBMSCL reserves the right to increase or decrease the quantity of goods and related services originally specified in Section IV, Schedule of Requirements, provided this does not exceed 40 % (Forty percent) above or 40% (Forty percent) below the original required quantity and without any change in the unit prices or other terms and conditions of the Bid Documents.

43. Publication of Award of Contract

- 43.1 WBMSCL shall publish the Award of Contract in e-tender portal and its website wbtenders.gov.in, www.wbmsc.co.in, www.wbmsc.gov.in, www.wbhealth.gov.in.

44. Signing of Contract

- 44.1 Prior to the expiry of the period of bid validity, WBMSCL shall issue Award of Contract (AOC). The draft agreement will be sent to the successful Tenderer along with the AOC and Special Conditions for Goods, if any.
- 44.2 Within 14 (fourteen) days of receipt of the AOC, the successful Tenderer shall sign and return the agreement to WBMSCL along with the required value of Performance Security in full or in parts in the event of a staggered supply as decided by WBMSCL.

45. Performance Security

- 45.1 Within 14 days of receipt of the AOC from WBMSCL, the successful Tenderer, if required, shall furnish the Performance Security in full or in parts in the event of a staggered supply as decided by WBMSCL in accordance with the GCC, using for that purpose the Performance Security Form included in Section VI, Contract forms, or another Form acceptable to WBMSCL. WBMSCL shall promptly discharge the Bid Securities of the unsuccessful Tenderers pursuant to Instructions to Tenderers.
46. Failure of the successful Tenderer to submit the above-mentioned Performance Security in full or in parts in the event of a staggered supply as decided by WBMSCL and sign the agreement within 14(fourteen) days of issue of AOC shall constitute sufficient grounds for the annulment of the award and forfeiture of the EMD.

Note: - Working demonstration of all the offered goods within India shall be required to be arranged by the Tenderer before the Bid Evaluation Committee for technical evaluation, as and when requested to do so by the Technical Bid Evaluation Committee prior to the opening of the financial bids. The cost incurred for the tour of the members of technical bid evaluation will be entirely borne by WBMSCL. Choosing of site for onsite physical demonstration from the list of installations submitted by a bidder / the institutes who have issued satisfactory certificate to the bidder shall be on the discretion of WBMSCL.

Section II. General Conditions of Contract

In the event of an order and any dispute arising out of the same, the FIRST PARTY General Conditions of Contract will apply as under and all references to the General Conditions of Contract include (subject to all relevant approvals) a reference to these terms and conditions as amended, supplemented, substituted, novated or assigned from time to time. Each schedule and annexure referred to in these terms and conditions shall form part of these terms and conditions. The documents forming the supply contract shall be construed and interpreted so that, in the event there is any conflict or ambiguity between them, these terms and conditions shall prevail.

1. APPLICATION AND LEGAL STATUS OF THE PARTIES:

The General Conditions of Contract incorporated in section –II shall be applicable for the purchase and / or supply of **3.0 TESLA MRI MACHINE** in the hospitals of the Government of West Bengal and to the extent the same are not superseded by the Special Conditions of Contract prescribed under section III, section IV or Schedule of Requirement of this document.

WBMSCL and VENDOR shall respectively be referred to as “FIRST PARTY” & “SECOND PARTY” hereunder and each party acknowledges and agrees that:

- 1.1 Nothing contained in or relating to the contract shall be construed as establishing or creating between the Parties the relationship of employer and employee or of principal and agent. The officials, representatives, employees, or subcontractors of each of the Parties shall not be considered in any respect as being the employees or agents of the other Party and each Party shall be solely responsible for all claims arising out of or relating to its engagement of such persons or entities.

2. DEFINITIONS:

GOODS: Goods are hereinafter deemed to include, without limitation, such medicines, raw materials, components, intermediate products and products which the Tenderer is required to supply pursuant to the Purchase Order or Special Conditions of this Contract to which these General Conditions are attached. Services are hereinafter deemed to include services ancillary to the supply of the Goods including, without limitation transportation and supply at the point of consignee and such other obligations as required under this Contract.

- 2.2 **TRADE TERMS:** Whenever an Inco term is used in this Contract it shall be interpreted in accordance with the Incoterms 2010 and as the same has been judicially interpreted in India.

3. CONTRACT PRICE;

Prices charged by the Tenderer for the Goods supplied and the related services performed under the Contract shall not vary from the prices quoted by the Tenderer in its bid, with the exception of any price adjustment authorized in writing by FIRST PARTY.

4. PACKAGING OF THE GOODS:

4.1 The SECOND PARTY shall package the Goods for delivery with the best materials that are adequate to safeguard the Goods while in transit and with all due care and according to the highest standards of export packaging for the type and quantities of the Goods. The Goods shall be packed and marked in a proper manner in accordance with the instructions stipulated in the Contract or, otherwise, as customarily done in the trade and in accordance with any requirements imposed by applicable law or by the transporters and manufacturers of the Goods. The packing, in particular, shall mark the Contract or Purchase Order number and any other identification information provided by FIRST PARTY as well as such other information as is customary for the Goods in question. During transit, the packing shall be sufficient to withstand, without limitation, rough handling and exposure to extreme temperatures, salt, precipitation and open storage. The SECOND PARTY shall have no right to any return of the packing materials.

5. TRANSPORTATION AND FREIGHT:

Unless otherwise specified in the Contract (including in any INCOTERM 2010) the SECOND PARTY shall be solely liable for making all transport arrangements and for payment of freight and insurance costs for the shipment and delivery of the Goods in accordance with the requirements of the Contract and as defined in table 5. **(b) Payment Terms of Section I: Instructions to Tenderers, under sub-section A. Important information at a glance.**

6. DELIVERY OF GOODS:

The SECOND PARTY shall hand over or make available the goods and the Consignees shall receive the goods (as per detail annexed in Section IV), at the place for the delivery of the Goods and within the time for delivery of the Goods specified as **per table 5 (a) Date of supplies & commissioning at Section I: Instructions to Tenderers under sub-section A. Important information at a glance**, in the Contract. All manuals, instructions, displays and any other information relevant to the Goods shall be in the English language unless otherwise specified in the Contract. Unless specifically stated in the Contract (including in any INCOTERM 2010) the entire risk of loss, theft, damage to, or destruction of the Goods shall be borne as defined in **table 5. (b) Payment Terms of Section I: Instructions to Tenderers, under sub-section A. Important information at a glance.**

7. INSPECTION OF THE GOODS:

7.1 All goods may be subjected to inspection and testing by FIRST PARTY or its designated representatives at all times and places including the period of manufacture and in any event prior to final acceptance by FIRST PARTY.

7.2 Neither the carrying out of any inspections of the Goods nor any failure to undertake any such inspections shall relieve the SECOND PARTY of any of its warranties or the performance of any obligations under the Contract.

7.3 **For goods supplied from within or outside India.**

a) For goods supplied from within or outside India, Purchaser retains the right to perform pre-shipment inspection at the manufacturer's premises and an independent quality control laboratory testing **at its own cost**.

b) The Purchaser will retain the right to perform further inspections and quality testing at any time till the satisfactory installation of Goods, as it deems fit, **at its own cost**.

7.4 Should any inspected or tested goods fail to conform to the specifications, the purchaser shall reject them and the supplier shall replace the rejected goods free of cost to the purchaser, within a period of 30 (thirty) days of intimating such rejection.

8. ACCEPTANCE OF GOODS:

Under no circumstances shall FIRST PARTY be required to accept any goods that do not conform to the specifications of or requirements of the Contract. FIRST PARTY may condition acceptance of the goods upon the successful completion of acceptance tests, as may be specified in the Contract or otherwise agreed in writing by the Parties. In no case shall FIRST PARTY be obligated to accept any goods unless and until FIRST PARTY has inspected the goods following commissioning of the goods in accordance with the requirements of the Contract. The goods shall be deemed to be accepted only after FIRST PARTY provides written acceptance.

9. REJECTION OF GOODS:

Notwithstanding any other rights of, or remedies available to, FIRST PARTY under the Contract, in case any of the goods is defective or otherwise does not conform to the specifications or other requirements of the Contract, FIRST PARTY may, at its sole option, reject or refuse to accept the goods and the SECOND PARTY agrees promptly to replace the goods with goods of equal or better quality.

10. TITLE:

Unless otherwise expressly provided in the Contract, title including the incidentals of the title and any legal or inchoate right and interest which may accrue in the said Goods shall pass from the SECOND PARTY to the FIRST PARTY upon delivery of the Goods and the acceptance of the same by the FIRST PARTY in accordance with the requirements of the Contract.

11. PERFORMANCE SECURITY:

11.1 Within 14 days of receipt of the AOC from WBMSCL, the successful Tenderer, if required, shall furnish the Performance Security in full or in parts in the event of a staggered supply as decided by WBMSCL as **per table 6, Performance Security (PS) at Section I: Instructions to Tenderers under sub-section A. Important information at a glance** for an amount of 10% of the Contract Price (bid value) in full or in parts, valid up to 60 days after the date of completion of all contractual obligations, till the last date of warranty obligations.

11.2 Banks issuing Performance Securities must be acceptable to the FIRST PARTY, i.e. they have to be scheduled commercial banks.

- 11.3 Discharge of the Performance Security shall take place upon expiry of the Performance Security or the completion of all contractual liabilities of the Supplier as per as per table 6, Performance Security (PS) at Section I: Instructions to Tenderers under sub-section A. Important information at a glance.
- 11.4 In the event of any amendment issued to the Contract, the Supplier shall, within 14 (fourteen) days of issue of the amendment, furnish the corresponding amendment to the Performance Security (as necessary) rendering the same valid in all respects in terms of the Contract, as amended.

12. WARRANTIES:

12.1 Goods Warranties:

Without limitation of any other warranties stated in or arising under the Contract, the SECOND PARTY warrants and represents that:

- 12.1.1 The Goods, including all packaging and packing thereof, conform to the specifications of the Contract, are fit for the purposes for which such Goods are ordinarily used and for the purposes expressly made known in writing by FIRST PARTY to the SECOND PARTY and shall be of even quality, free from faults and defects in design, material, manufacture and workmanship under normal use in the conditions prevailing in the country of final destination;
- 12.1.2 If the SECOND PARTY is not the original manufacturer of the Goods, the SECOND PARTY shall provide FIRST PARTY with the benefit of all manufacturers' warranties in addition to any other warranties required to be provided hereunder;
- 12.1.3 The Goods are of the quality, quantity and description required by the Contract;
- 12.1.4 The Goods are free from any right of claim by any third-party and unencumbered by any title or other rights, including any liens or security interests and claims of infringement of any intellectual property rights, including, but not limited to, patents, copyright and trade secrets.
- 12.1.5 Unless otherwise indicated in the Technical Specifications, this warranty shall remain valid for 1 (One) year after the Goods have been commissioned at the final destination indicated in the Contract subject to issue of certificate regarding date of commissioning issued by the consignee.
- 12.1.6 During the warranty, free comprehensive annual maintenance and repairs services including testing and calibration, labour and spares shall be provided by the supplier during the period of warranty.
- 12.1.7 If the Supplier, having been notified, fails to remedy the defect(s) within the stipulated period, the Purchaser may proceed to take such remedial action as may be necessary, at the Supplier's risk and expense and without prejudice to any other rights which the Purchaser may have against the Supplier under the Contract.
- 12.1.8 The supplier shall, at all times, indemnify and keep indemnified the purchaser , free of cost, against all claims which may arise in respect of goods and services to be provided by the supplier under the contract for infringement of any intellectual property rights or any other right protected by patent, registration of designs or trademarks. In the event of any such claim in respect of alleged breach of patent, registered designs, trademarks etc. being made against the purchaser, the purchaser shall notify the supplier of the

same and the supplier shall, at his own expenses take care of the same for settlement without any liability to the purchaser.

- 12.1.9 The supplier shall be responsible and shall indemnify on account of any emission or radiation that may cause harm to the user of the supplied product.
- 12.1.10 The supplier shall visit each installation site as recommended in the manufacturer's technical/ service operational manual and as mentioned in the bid document during the warranty / CMC period for preventive maintenance.
- 12.1.10 The Goods shall be new and unused. The SECOND PARTY shall remain responsive to the needs of FIRST PARTY for any services that may be required in connection with any of the SECOND PARTY's warranties under the Contract. All warranties will remain fully valid following any delivery of the Goods and for a period of not less than one (1) year following acceptance of the Goods by the Consignee in accordance with the Contract. During any period in which the SECOND PARTY's warranties are effective, upon notice by FIRST PARTY that the Goods do not conform to the requirements of the Contract, the SECOND PARTY shall replace the defective Goods with Goods of the same or better quality or fully reimburse FIRST PARTY for the purchase price paid for the defective Goods; and if having been notified by any means, the SECOND PARTY fails to replace the defective Goods within 30 days, FIRST PARTY may proceed to take such remedial action as may be necessary, at the SECOND PARTY's risk and expense and without prejudice to any other rights which FIRST PARTY may have against the SECOND PARTY under the Contract. The reasonableness of the period of time granted for remedy is subject to the discretion of FIRST PARTY according to the circumstances of the Contract.

13.

INDEMNIFICATION:

The SECOND PARTY shall indemnify, defend and hold the FIRST PARTY, the DoHFW and the Government agencies harmless against any or all proceedings, actions and third party claims arising out of a breach by the SECOND PARTY of any of its obligations under this agreement.

This indemnity shall be limited in respect of making harmless to the FIRST PARTY, the DoHFW and the Government agencies.

The bidder shall indemnify the FIRST PARTY against all actions, suits, claims and demands brought or made against it, in respect of anything done or committed to be done by the SECOND PARTY in execution of or in connection with the work of this contract and against any loss or damage to the FIRST PARTY in consequence to any action or suit, or a legal proceeding, being brought against the bidder for anything done or committed to be done in the execution of this contract. The bidder will abide by the job safety measures prevalent in India and will free the FIRST PARTY from all demands or responsibilities arising from accidents or loss of life, on account of the bidder's negligence and responsibility. The bidder will pay all indemnities arising from such incidents without any extra cost to FIRST PARTY and will not hold the FIRST PARTY responsible or obligated. The FIRST PARTY may at its discretion and entirely at the cost of the bidder defend such suit, either jointly with the bidder or severally in case the latter chooses not to defend the case and /or proceeding.

14.

LIQUIDATED DAMAGES:

Except under the circumstances of force majeure as described, if the SECOND PARTY fails to deliver any or all of the Goods by date(s) of delivery as per conditions of the

contract, FIRST PARTY may, without prejudice to any or all its other remedies under the contract, deduct from the contract price, as per **table - 9 Liquidated damages for delayed delivery / delayed setting up of Services at Section I: Instructions to Tenderers under sub-section A. Important information at a glance**

15.

BLACKLISTING:

Any manufacturer/ its subsidiary which has been black-listed by any Government Department/Agency in India during the last five years, would not be eligible to participate in the tender.

16.

PENALTY FOR DEFAULT:

In case of failure by the Tenderer to perform according to this Contact to keep Service Up time in Warranty & CMC of all of the Goods, the Company may exercise one or several of the penal provisions as per **table – 8, Service Up time in Warranty & CMC at Section I: Instructions to Tenderers under sub-section A. Important information at a glance.**

16.1

In addition to what has been stated above, the following penalties shall be imposed against offences mentioned against each:

Nature of offence	Penalty to be imposed
Any wrong or misleading information provided by the Tenderer during submission of bids	a. Forfeiture of EMD b. May lead to blacklisting in FIRST PARTY for at least 5 years
Non execution of agreement within 14 days of issue of AOC	a. Forfeiture of EMD b. Blacklisting for 5 years in FIRST PARTY c. Blacklisting to be circulated to all procurement agencies throughout the country
Supplying refurbished goods instead of new	a. Termination of Contract. b. Blacklisting for 5 years. c. Blacklisting to be circulated to all procurement agencies throughout the country. d. Forfeiture of the Performance Bank Guarantee. Lodging FIR.
Breach of Agreement	a. Termination of Contract. b. Blacklisting for 5 years. c. Blacklisting to be circulated to all procurement agencies throughout the country. d. Forfeiture of the Performance Bank Guarantee e. Lodging FIR

16. CHANGES IN QUANTITY:

FIRST PARTY may at any time by written instruction vary the general scope of this Contract by 40% (forty percent) above or 40% (twenty percent) below the original Contract quantity at the accepted terms & conditions. The price for the additional quantity will be as per the contracted price of this bid.

17. TERMINATION FOR CONVENIENCE:

17.1 FIRST PARTY may, upon notice to the Tenderer, terminate this Contract, in whole or in part, at any time for its convenience. The notice of termination shall state that termination is for FIRST PARTY's convenience, the extent to which performance of the SECOND PARTY under the Contract is terminated and the date upon which such termination becomes effective.

17.2 In the event of Termination for Convenience, no payment shall be due from FIRST PARTY to the Tenderer except for Goods satisfactorily delivered and for the cost of such necessary work as FIRST PARTY may request the Tenderer to complete.

18. TERMINATION FOR DEFAULT:

18.1 FIRST PARTY, without prejudice to any other remedy for breach of Contract, by written notice of default sent to the SECOND PARTY, may terminate the Contract, in whole or in part if:

18.1.1 The SECOND PARTY fails to deliver any or all of the Goods within the period specified in the Contract:

18.1.2 The SECOND PARTY fails to perform any other obligation under the Contract;

18.1.3 The SECOND PARTY, in the judgment of FIRST PARTY, has engaged in fraud and corruption, in competing for or in executing the present Contract;

18.1.4 The SECOND PARTY attempts to offer any direct or indirect benefit arising from or related to the performance of the Contract or the award there of to any representative, official, employee or other agent of FIRST PARTY or any organization of Health & Family Welfare Department, Government of West Bengal;

18.1.5 The SECOND PARTY is adjudged bankrupt, or is liquidated, or becomes insolvent, applies for moratorium or stay on any payment or repayment obligations, or applies to be declared insolvent;

18.1.6 FIRST PARTY reasonably determines that the SECOND PARTY has become subject to a materially adverse change in its financial condition that threatens to endanger or otherwise substantially affect the ability of the SECOND PARTY to perform any of its obligations under the Contract.

18.1.7 Non-compliance of all statutory norms and applicable laws relating to the said contract will entitle FIRST PARTY to terminate the contract.

18.2 Upon occurrence of one or more of the events specified above, FIRST PARTY shall follow the procedure of issuing notice or show cause specifying the time frame and on

being not satisfied with the explanation, be entitled to terminate the Contract immediately. The decision of FIRST PARTY shall be final and binding on the Tenderer.

19. CONSEQUENCES OF TERMINATION:

- 19.1 In the event of any termination of the Contract, upon receipt of notice of termination by FIRST PARTY, the SECOND PARTY shall, except as may be directed by FIRST PARTY in the notice of termination or otherwise in writing:
 - 19.1.1 Take immediate steps to bring to a close in a prompt and orderly manner the performance of any obligations under the Contract, including, but not limited to, fulfilling any outstanding orders for Goods under the Contract and in doing so, reduce expenses to a minimum;
 - 19.1.2 Place no further orders for Goods or other materials, except as FIRST PARTY and the SECOND PARTY agree in writing are necessary to fulfil any outstanding order or to complete any portion of the Contract that has not been terminated;
 - 19.1.3 Transfer title and deliver to FIRST PARTY any Goods remaining to be delivered as stipulated in the notice of termination; and
 - 19.1.4 Take any other action that may be necessary or that FIRST PARTY may direct in writing, for the protection and preservation of any property, whether tangible or intangible, related to the Contract that is in the possession of the SECOND PARTY and in which FIRST PARTY has or may be reasonably expected to acquire an interest.
- 19.2 In the event of any termination of the Contract, FIRST PARTY shall not be liable to pay the SECOND PARTY except for those Goods delivered to FIRST PARTY in accordance with the requirements of the Contract, but only if such Goods were ordered, requested or otherwise provided prior to the SECOND PARTY's receipt of notice of termination from FIRST PARTY.

20. CONFIDENTIALITY:

- 20.1 FIRST PARTY and the SECOND PARTY, its agents, employees, sub-contractors and servants shall keep confidential and shall not, without the written consent of the other party hereto, divulge to any third party any documents, data, or other information furnished directly or indirectly by the other party hereto, in connection with the Contract, whether such information has been furnished prior to, during or following competition or termination of the contract. Notwithstanding the above, the SECOND PARTY may furnish to its sub-contractor such documents, data and other information it received from FIRST PARTY to the extent required for the sub-contractor to perform its work under the contract, in which event the SECOND PARTY shall obtain from such sub-contractor an undertaking of confidentiality similar to that imposed on the SECOND PARTY.
- 20.2 FIRST PARTY shall not use such documents, data and other information received from the SECOND PARTY for any purpose unrelated to the contract. Similarly, the SECOND PARTY shall not use such documents, data and other information received from FIRST PARTY for any purpose other than the performance of the contract.

- 20.3 The obligation of a party under the two foregoing paragraphs shall not apply to information that:
- 20.3.1 Now or hereafter enters the public domains through no fault of that party;
- 20.3.2 Can be proven to have been possessed by that party at time of disclosure and which was not previously obtained, directly, from the other party, or
- 20.3.3 Otherwise lawfully becomes available to that party from a third party that has no obligation of confidentiality.

21. FORCE MAJEURE:

- 21.1. Force majeure as used herein means any unforeseeable and irresistible act of nature, any act of war (whether declared or not), invasion, revolution insurrection, flood earthquake or any other acts of a similar nature or force, provided that such acts arise from causes beyond the control and without the fault or negligence of the SECOND PARTY. The SECOND PARTY acknowledges and agrees that, with respect to any obligation under the contract that the SECOND PARTY must perform any delays or failure to perform such obligation arising from or relating to harsh conditions within such areas shall not, in and of itself, constitute Force majeure under the contract. Further the SECOND PARTY acknowledges and agrees that scarcity of raw materials, power cut, workers unrest (even if wide spread) will not constitute force majeure under the contract.
- 21.2. In the event of and as soon as possible after the occurrence of any cause constituting Force majeure, the SECOND PARTY shall give notice and full particulars in writing to FIRST PARTY, of such occurrence or cause if the SECOND PARTY is thereby rendered unable, wholly or in part to perform its obligations and meet its responsibilities under the contract. The SECOND PARTY shall also notify FIRST PARTY of any other changes in condition or the occurrence of any event which interferes or threatens to interfere with its performance of the contract. Not more than fifteen (15) days following the provision of such notice Force majeure or other changes in conditions or occurrence, the SECOND PARTY shall also submit a statement to FIRST PARTY of estimated expenditure that will likely be incurred for the duration of the change in condition or the event. On receipt of notice or notices required hereunder, FIRST PARTY shall take such action as it considers, in its sole desertion, to be appropriate or necessary in the circumstances, including the granting to the SECOND PARTY of a reasonable extension of time in which to perform any obligations under the contract.
- 21.3. If an event of force majeure exists and the SECOND PARTY fails, within seven (7) days of such event to give notice in writing to FIRST PARTY and if the SECOND PARTY is rendered permanently unable, wholly, or in part, by reason of force majeure to perform its obligations and meet its responsibilities under the Contract, FIRST PARTY shall have the right to suspend or terminate the contract on the same terms and conditions except that the period of notice shall be seven (7) days. In any case, FIRST PARTY shall be entitled to consider the SECOND PARTY permanently unable to perform its obligations under the contract in the case of the SECOND PARTY's suffering any period of suspension in excess of ninety (90) days.

- 21.4. In the event of force majeure the benefit accrued to the FIRST PARTY in terms of title and any accrued right thereof including all inchoate rights shall remain with the FIRST PARTY and the SECOND PARTY shall not have any claim with the same.

22. SOURCE OF INSTRUCTIONS:

The SECOND PARTY shall neither seek nor accept instructions from any authority external to FIRST PARTY in connection with the performance of its obligations under the contract. Should any authority external to FIRST PARTY seek to impose any instructions on the SECOND PARTYs regarding the SECOND PARTY's performance under the contract, the SECOND PARTYs shall promptly notify and shall provide all reasonable assistance required by FIRST PARTY. The SECOND PARTY shall not take any action in respect of its performance of the contract or otherwise related to its obligations under the contract that may adversely affect the interests of FIRST PARTY and the SECOND PARTY shall perform its obligations under the contract with the fullest regard to the interests of FIRST PARTY.

23. BENEFITS, CORRUPTION AND FRAUD:

- 23.1 The SECOND PARTY warrants that it has not and shall not offer any direct or indirect benefit arising from or related to the performance of the contract or the award thereof to any representative, officials, employee, or other agent of FIRST PARTY or any official of the Health & Family Welfare Department, Government of West Bengal or any organization engaged in the procurement process whether during the period the contract is in process or before or after the contract is over. The SECOND PARTY acknowledges and agrees that any breach of this provision is a breach of an essential term of the contract as specified.
- 23.2 Corruption means the offering, giving, receiving or soliciting of, directly or indirectly, anything of value to influence the action of any FIRST PARTY representative, official, employee or agent of FIRST PARTY or any official of the Health & Family Welfare department, Government of West Bengal or any organization engaged in the selection process or in the execution of the contract.
- 23.3 Fraud means a misrepresentation or omission of facts in order to influence the selection process or the execution of the contract.

24. USE OF NAME OR OFFICIAL SEAL OF FIRST PARTY:

The SECOND PARTY shall not advertise or otherwise make public for purpose of commercial advantage or goodwill that it has a contractual relationship with FIRST PARTY, nor shall the SECOND PARTY, in any manner whatsoever use the name or official seal of FIRST PARTY, or any abbreviation of the name of the FIRST PARTY or Health & Family Welfare department, Government of West Bengal in connection with its business or otherwise without the written permission of FIRST PARTY.

25. ASSIGNMENT:

- 25.1 The SECOND PARTY shall not, except after obtaining the prior written approval of FIRST PARTY, assign, transfer, pledge, or make any other disposition of this contract or any part hereof or of any of the SECOND PARTY's right or obligations hereunder,

except with the prior written authorization of FIRST PARTY. The SECOND PARTY may assign or otherwise transfer the contract to the surviving entity resulting from a reorganization of the Party's operations.

- 25.2 Prior to the written approval of FIRST PARTY, the SECOND PARTY shall promptly notify FIRST PARTY of such assignment at the earliest opportunity subject to the assignee or transferee agrees in writing to be bound by all of the terms and conditions of the contract and such writing is promptly provided to FIRST PARTY following the assignment or transfer and FIRST PARTY finds that the SECOND PARTYs has the financial and technical capacity as laid down in the tender document to carry out the assignment provided that:
- 25.2.1 Such reorganization is not the result of any bankruptcy, receivership or other similar proceedings; and
- 25.2.2 Such reorganization arises from sale, merger, or acquisition of all or substantially all of the SECOND PARTY's assets or ownership interest; and
- 25.2.3 Such reorganization is not taking place with any of the Tenderer who had participated in the Tender or who will be deemed to have conflict of interest as defined in the tender documents process for the same tender.
- 25.3 However, should the SECOND PARTY become insolvent or should control of the SECOND PARTY change by virtue of insolvency, FIRST PARTY may, without prejudice to any other right or remedy, terminate this contract.

26. AMICABLE SETTLEMENT:

When a dispute arises under this agreement, the parties shall make all reasonable efforts to resolve through good faith negotiation, failing which they will attempt at dispute resolution with the intervention of the Principal Secretary, the DoHFW, GoWB

27. ARBITRATION:

1. Except for a dispute in connection with termination in which respect the decision of FIRST PARTY shall be final, any dispute between the parties arising out of or relating to this agreement which cannot be resolved through good faith negotiation shall be settled in arbitration, in terms of the provisions of the Arbitration and conciliation Act-1996(no.26 of 1996). The arbitration hearing shall be held in Kolkata only. The award of the arbitrator (s) shall be binding on both the parties. The cost of arbitration shall be borne by the respective parties.
2. Pending the submission of and / or decision on a dispute, difference or claim, or until the arbitral award is published, the party shall continue to perform all of their obligations under this agreement without prejudice to a final adjustment in accordance with such award.

28. COURT OF LAW:

In case of any dispute in between the parties, the matter will be settled in appropriate Court of Law within Kolkata Jurisdiction.

29. QUALITY OF EQUIPMENT:

The equipment should have compliance with CE (European Conformity) standards & safety. In case the name of the offered model is not under the scope of the certification, the bidder will submit a notarized declaration in the format given in FORM 12: Declaration of Quality Certification of Equipment

Section III. Special Conditions of Contract

The following Special Conditions of Contract (hereinafter referred to as SCC) shall supplement the General Conditions of Goods (hereinafter referred to as GCC). Whenever there is a conflict, the provisions herein shall prevail over those in the GCC. The corresponding clause number of the GCC is indicated in the first column.

SCC -1	GCC - 6	<p>Delivery of Goods shall be made by the SECOND PARTY in accordance with the Schedule of Requirement</p> <p>However, the FIRST PARTY may swap facilities between phases and/or substitute any facility by a new one if deemed necessary.</p> <p>The details of shipping and/or other documents, as applicable under I or II, to be furnished by the SECOND PARTY are:</p> <p>I. For Goods supplied from abroad :</p> <p>(A) Upon shipment, within 24 hours the SECOND PARTY shall notify the Purchaser in writing the full details of the shipment including Contract number, description of the Goods, quantity, date and port of shipment, mode of shipment, estimated dates of arrival at the port of entry and the place of destination. In the event of Goods sent by airfreight, the SECOND PARTY shall notify the Purchaser a minimum of forty-eight (48) hours ahead of dispatch, the name of the carrier, the flight number, the expected date and time of arrival, the Master airway-bill and the House airway- bill numbers. The SECOND PARTY shall first fax the above details and then send to the Purchaser, by courier, two sets of documents comprising one original and one copy of the following:</p> <p>(i) Commercial invoice, indicating the West Bengal Medical Services Corporation Limited as the Purchaser on behalf of the Department of Health and Family Welfare, Government of West Bengal; the Contract number, Goods description, quantity, unit price and total amount. Invoices must be signed in original and stamped, or sealed with the company stamp/seal;</p> <p>(ii) Negotiable, clean, on-board through bill of lading marked "freight prepaid" and indicating the West Bengal Medical Services Corporation Limited as the Purchaser on behalf of the Department of Health and Family Welfare, Government of West Bengal and notify Consignees as stated in the Contract, with delivery through to final destination as per the Schedule of Requirements and non-negotiable bill of lading, or railway consignment note, road consignment note, truck or air waybill, or multimodal transport document, marked "freight prepaid" and showing delivery through to final destination as per the Schedule of Requirements;</p> <p>(iii) Packing list identifying contents of each package;</p> <p>(iv) Manufacturer's Warranty Certificate covering all items supplied;</p> <p>(v) Manufacturer's Certificate of Origin covering all items supplied;</p> <p>(vi) Original copy of the certificate of weight issued by the port authority/licensed authority and six copies;</p>
--------	---------	--

		<p>(vii) Manufacturer will submit a pre-shipment advisory note to purchaser & consignee at least 15 days prior to the scheduled delivery of the equipment at the door step of the consignee;</p> <p>(viii) Acknowledgement of receipt of Goods by the Consignees, i.e. Consignment Receipt Certificate (CRC).</p> <p>The above documents shall be received by the 'Purchaser' at least 15 days before arrival of Goods at the port or place of arrival and, if not received, the SECOND PARTY will be responsible for any consequent expenses.</p> <p>II. For Goods from within India</p> <p>(A) Upon the delivery of the Goods, the SECOND PARTY shall notify the Purchaser in writing and deliver to the Purchaser two sets of documents comprising one original and one copy of the following:</p> <p>(i) Commercial invoice, indicating the West Bengal Medical Services Corporation Limited as the Purchaser on behalf of the Department of Health and Family Welfare, Government of West Bengal, the Contract number, loan number; Goods' description, quantity, unit price and total amount. Invoices must be signed in original and stamped or sealed with the company stamp/seal;</p> <p>(ii) Railway consignment note, road consignment note, truck or airway bill, or multimodal transport document showing the Purchaser as the West Bengal Medical Services Corporation Limited on behalf of the West Bengal Medical Services Corporation Limited and delivery through to final destination as stated in the Contract;</p> <p>(iii) Acknowledgement of receipt of Goods by the Consignees, i.e. Consignment Receipt Certificate (CRC);</p> <p>(iv) Packing list identifying contents of each package;</p> <p>(v) Manufacturer's or SECOND PARTY's Warranty certificate covering all items supplied;</p> <p>(vi) Manufacturer will submit a pre-shipment advisory note to purchaser & consignee at least 15 days prior to the scheduled delivery of the equipment at the door step of the consignee.</p>
SCC -2	GCC - 30	<p>Add clause 30- Payment Terms -</p> <p>As per table 5 (b) Payment Terms at Section I: Instructions to Tenderers under sub-section A. Important information at a glance.</p>

SCC-3	GCC - 31	<p>Add clause GCC 31- The Comprehensive Maintenance Contract (Including Spare parts)</p> <p>(i) The Consignees/ Government of West Bengal / Operation & Maintenance (O&M) Partner, may, at his own and sole discretion enter into a Comprehensive Maintenance Contract (CMC) with the SECOND PARTY, three months prior to the completion of Warranty Period, at the contracted price, for a period as specified in the tender after the expiry of the warranty period provided that this service shall not relieve the SECOND PARTY of any warranty obligations under this Contract. Wherever the Technical Specifications lay down a different period of CMC, this latter period shall prevail. The CMC will commence from the date of expiry of warranty period. The CMC includes preventive maintenance including testing & calibration as per technical/service/operational manual, labour and spares.</p> <p>(ii) The CMC includes repairs of entire system, preventive maintenance, testing & calibration, labour and spares and all software updates.</p> <p>(iii) The Comprehensive Annual Maintenance and Repair charges (after Warranty period) shall be paid quarterly in four equal installments.</p> <p>(iv) Details of CMC requirements or otherwise, as spelt out in the Technical Specifications, will prevail over those given in this section.</p>
SCC - 4	GCC -31.1	<p>Add GCC clause 31.1</p> <p>For both Warranty and CMC, as per table 8. Service Up time in Warranty & CMC at Section I: Instructions to Tenderers under sub-section A. Important information at a glance.</p>
SCC - 5	GCC - 32	<p>Add GCC clause 32</p> <p>The successful bidder shall be required to undertake supplies of quantity as per schedule of requirement in phases spread over a period of 2 years from the date of signing of the contract.</p>
SCC - 6	GCC -33	<p>Add GCC clause 33</p> <p>The successful bidder shall undertake to commission the equipment including the interiors of the site as per technical specification mentioned in Section IV. Schedule of Requirements.</p>

Section IV. Schedule of Requirements

Contents

1. List of Consignee
2. Technical Specifications
3. Standard requirements
4. List of related services
5. Inspections and Tests

1. List of Consignee

Description of Goods	Quantity	Unit	Consignee List
3.0 TESLA MRI MACHINE	2	Number	IPGME & R and SSKM Hospital (1 unit); Consignee for remaining 1 (one) unit will be notified later

Please note:

- The Consignee Receipt Certificate (CRC) will be issued to the Vendor within 72 hours of the delivery at the Consignee address.

2. TECHNICAL SPECIFICATIONS

Equipment Name: 3T MRI Scanner

WHOLE BODY 32 CHANNEL 3 TESLA MAGNETIC RESONANCE IMAGING SYSTEM WITH STATE OF ART ADVANCED TECHNOLOGY

The essential features mentioned below should be verifiable from catalogue / technical brochures of equipment for the offered model of the equipment and vendor must specify page, paragraph and line number in technical literature that indicates compliance. The system should have True 32 channels RF system with 32 independent ADC. The system should be totally new and should not contain refurbished or having recycled items. Latest version of quoted equipment and accessories available in the market at time of issue of supply order is to be supplied by the firm.

I. MAGNET

- 3 Tesla (superconducting) Magnet with approximately 70 cm or more bore diameter.
- a) **Field Strength:** Helium only 3 T (superconducting) Magnet along with Magnet Power supply Facility for quick shutdown of the magnet in case of emergency.
 - b) **Field Stability over time:** Should have active shielding, external interference shielding with good field stability
 - c) **Homogeneity:**
 - (i) Best homogeneity possible should be given.
Specify homogeneity in VRMS at 10 cm, 20 cm, 30 cm, 40 cm and 45 cm DSV and at maximum FOV achievable with the quoted scanner.
 - (ii) Should be very good for Single voxel and CSI spectroscopy. Specify values
 - (iii) Please specify the homogeneity at 40 cm FOV. (Guaranteed homogeneity).
 - (iv) Automatic shimming in phantom.
 - d) **Magnet Bore:** 70 cm or more magnet bore diameter, after positioning of gradient, shim and RF coils.
 - e) **Active Shielding/ Fringe field:** Quote values for 5 Gauss and 1 Gauss line.
 - f) **Ext. Shielding:** Ext. interference shield (sufficient to house the Magnet, Anaesthesia and physiologic monitors) should be provided.
 - g) **Magnet Cooling System:**
 - (i) Specify the boil off rate

- (ii) Devices for helium level monitoring in the magnet should be supplied.
- (iii) Liquid helium should be supplied during warranty period and Comprehensive AMC.
- (iv) The vendor should include the Cold Head maintenance and replacement during warranty period and also during comprehensive AMC.

h) Shim System:

- (i) High performance and highly stable shim system with global and localized manual and auto-shimming for high homogeneity magnetic field required for imaging (MRI / fMRI), single voxel spectroscopy (MRS), and spectroscopic imaging (MRSI).

3D shimming for volume imaging and CSI.

- (ii) Auto shim (global and voxel shim) should take less than 3 seconds to shim the magnet with patient in position (specify the time).
- (iii) Specify number of shim coils including higher order.
- (iv) Please specify whether second order shimming is required for better results on spectroscopy studies for the quoted model. If yes the same should offered as standard
- v) Specify whether off centre FOV shimming is required, if yes, please specify the details.

i) Others:

- i) The scan range for imaging should be 200 cm or more
- ii) The FOV (Field of view) in X, Y & Z directions should be 45 cm or more
- iii) Please specify up to what FOV Gradient linearity is maintained

II. PATIENT TABLE MONITORING & COMMUNICATION

- a) The vendor should supply Dockable Table/ Table Top/ Table Top with dockable trolley for the main MRI patient table. One MR compatible patient shifting trolley and one MR compatible wheel chair should be offered as standard.
- b) The table should be fully motorized, computer controlled table movements in vertical and horizontal directions.
- c) Table should be able to handle patient weight of 200 kg or above.
- d) Cushions and other patient comfort accessories. All parts of the table should be protected from liquid spill

- e) The table should have facility for manual traction in case of emergency.
- f) Patient monitoring devices for ECG, respiratory, pulse rate, oxygen saturation, at the console etc. A comprehensive solution at patient side and at main console capable of gating the sequence protocols with respect to patient's heart (ECG) and respiratory rates. Remote display of gating signals on magnet and at console.
- g) The table should deliver the protocols for automatic bolus chasing in peripheral angio with automatic table movement.
- h) Table should be compatible for future advanced applications like MR guided Focused US/RT Planning.
- i) The table should have facility for manual traction in case of emergency.
- j) A CCTV system with LCD display to observe the patient should be provided.
- k) Two-way Patient communication with headphone, microphone and necessary accessories
- l) There should be a hand held alarm for patients.

III. GRADIENT SYSTEM

a) General:

- i) Actively shielded gradient system in X, Y, Z planes.
- ii) Minimum Gradient Strength should be 40 mT/m or more along each axis and a slew rate of 200 T/m/s in each axis.
- iii) In case of dual gradient systems, please mention the details in each axis separately.
- iv) Quote the minimum rise time at the maximum gradient strength offered. The rise time should not be more than 250 ms, to reach the maximum gradient strength
- v) Quote the Slew rate at the maximum gradient strength
- vi) Specify the linearity of the gradients at full FOV.
- vii) 100% duty cycle for full FOV.
- viii) Effective cooling system for gradient coil and power supply, for uninterrupted operation. The system should have efficient and adequate provision for eddy current compensation.

b) Resolution Parameters:

- i) Field of View should be at least 45 cm in all three axes. Specify the minimum and maximum FOV achievable for the quoted MR system.

- ii) Minimum TE & TR in 2D/3D should be specified in relation to the sequences.
- iii) Specify min. slice thickness in 2D and 3D modes at 128x128, 256x256, 512x512 and 1024x1024 matrices (quote higher matrix resolution, if available)
- iv) The system should be capable of performing single shot EPI (in 64x64, 128x128, and 256x256 matrixes) including conventional and fluoroscopic imaging in the three orthogonal and also oblique planes.
- v) Echo Train length in both Spin echo and Gradient Echo should be at least 255 or more.
- vi) The measurement matrix should be from 128x128 to 1024x1024 in both 2D and 3D imaging as well.

IV. RF SYSTEM

a) RF Transmitter:

- (i) A fully digital RF System capable of transmitting power of at least 25 KW or more (Single/dual) with a combination of RF power amplifiers (please quote the value). System should be capable of Multi Transmit / Multi Drive / True shape. (as per FDA guidelines), and the operating frequency should cover ^1H , and ^{31}P nuclei (for multinuclear spectroscopy of $^1\text{H}/^{31}\text{P}$).
- (ii) Specify max transmitter RF power available (at 50 Ω impedance)

b) RF Receiver:

- (i) Optical / Digital RF receiver system with / high efficient RF receiver system / or its equivalent located on the magnet inside the shielded scan room.
- (ii) It should also have minimum of 32 independent ADC hardware RF channels without multiplexing, with each having bandwidth of 1MHz or more. Please provide the list of coils / coil-combinations that use this configuration
- (iii) Specify the RF receiver bandwidth for each channel.
- (iv) The system should have necessary hardware to support quadrature phased array and flex coils.
- (v) It should support Parallel acquisition techniques like ASSET / SENSE / iPAT / Sence /Grappa / dS-Sense with a factor of a least 4. Higher sectors if available should be offered optionally.
- (vi) Should allow remote selection of coils and/or coil elements.

c) RF Transmit technology:

- (i) Latest RF transmit system (like Multi-transmit/ Multi Drive transmit system or its equivalent/Trueshape multi-transmit, etc) with at least two independent output channels should be offered to improve RF uniformity and signal homogeneity and to reduce patient induced in-homogeneities.
- (ii) If the vendor has two RF amplifiers for producing better image quality/ features, the same should be quoted.

d) SAR limits:

SAR limits should be as per FDA guidelines for all protocols, including neuro / abdominal imaging.

e) Coils:

- (i) The system body coil integrated to the magnet must be Quadrature / CP. In addition to this coil, following Coils (Preferably be with equal number of elements as the channels) be quoted. RF coils in addition to main body coil (Transmit / Receive or receive coils) auto tune, array or no tune coils. Coils for the following applications should be available with the system. Coil / RF design should support compatibility to coils manufactured by other manufacturers. If one coil can cover multi applications please mention it. Please confirm that the system can adapt to coils developed and manufactured by other manufactures. All coils offered should have SENSE / Ipat2 / GEM compatibility. All coils (other than coils for exclusive spectroscopy, like surface coils) should be compatible for parallel acquisition. Please specify the number of channels and elements available for each coil. Please mention the true acceleration factor for each of the array coil.
- (ii) Standard Head coil (15-channel or more)
- (iii) Head coil (32-channel or more) for EPI / DTI applications. Compatible with FMRI projection device quoted with the system.
- (iv) Neurovascular coil (20-channel or more) for neurovascular applications. If separate neck coil can work in combination with head coil, then the neck coil is to be quoted, and the vendor should make sure NV application is satisfied.
- (v) Spine Array / Matrix Coils for thoracic and lumbar spine imaging - at least 32 elements or more.
- (vi) Body Array / Matrix coil with at least 45 cm coverage for imaging of whole

abdomen (32 Channel or more).

If a single coil is not available with the vendor, then a combination of coils should be quoted (capable of single station Cardiac / abdominal imaging), so that the resolution over 40 FOV is not compromised.

- vii) Dedicated Bilateral Brest Coil with at least 16 channel or more (Coils should be capable of doing breast spectroscopy).
- viii) Specify the coil to perform cardiac imaging (32 channel and above).
 - ix) Coils to do high resolution head, neck, spine in one go up to 75 cm coverage with parallel imaging techniques.
 - x) Additional coil / coils for peripheral angiography application (32 Channel or more) with high SNR and high spatial resolution without venous contamination to be offered from pelvis to ankle. The coverage should be at least 60 cm and it should be possible to combine this coil with body coils to perform whole body angio exam.
 - xi) Shoulder array coil (16 Channel or more)
 - xii) Dedicated wrist coil (8 Channel or more)
 - xiii) Extremity coil (8 Channel or more)
 - xiv) Knee imaging (Transmit/Receive 16 Channel or more)
 - xv) Set of large and medium/small general purpose wrap around flexible coils (minimum 2) for extremity Imaging (4 channel or more)
 - xvi) Endorectal coil (quantity-10) with hardware for prostate imaging.

f) **Coil technology:**

Integrated coil technology, latest as available with the vendor to be quoted:
Equivalent of TIM / GEM / D Stream to be offered.

g) **Table technology:**

Bolus chasing with automatic / continuous moving table should be offered and should be available with fluoro triggered MR angiography for manual and fast switchover in less than 1 sec for CE-MRA.

V. COMPUTER CONTROL SYSTEM

The vendor should supply the latest computer system along with the MR system, to handle all the latest applications available on the MR platform. During the warranty period, any hardware updates that are launched globally should be supplied and installed.

a) **Host Computer and Array Processors:**

- (i) Latest state-of-art computer system with sufficient RAM (8 GB or more) and computational speed to match the single shot Echo Planar Imaging (EPI), interactive angiogram, multi-planar three dimensional (3D) reconstruction, surface rendering and dynamic imaging, vascular imaging / angiography, and adequate storage for images and other applications.
- (ii) Necessary image processor with large RAM for ultra-fast image reconstruction capable of performing real-time image reconstruction should be provided. It should be at least 12 or higher GB RAM. Please specify RAM and reconstruction speed in images per second for full FOV 256 matrix.
- (iii) Total hard disk memory to be sufficient to store at least 250,000 images of 256x256 matrix data size.
- (iv) Monitor: 19" or more TFT monitor with enhanced graphics accelerator.
- (v) One measurement (Main) console capable of data acquisition and all online calculations (as required for all sequences in the tender), and Post processing (as required for all applications in the tender).
- (vi) Licenses for acquisition (as required for all sequences in the tender), post-processing and for special packages should be given explicitly (as required for all applications in the tender), listing all the capabilities of the vendor's quoted product (basic standard package, premium packages, etc).
- (vii) The main console/workstation should have pulse sequence software license that may be required to modify and run pulse sequences. If this is not possible, the vendor should provide the necessary hard and softwares necessary for such application (like laptop with system interface solution). Appropriate procedures (like research agreement) should be finalized before the installation of the equipment, so that there is no delay in operation of any requirement.
- (viii) Dual DVD write/CD Read/Rewrite drive for writing of images, spectra and raw data along with the necessary software for reading the images and spectra on DVD/CD storing capabilities. Provision for archival of k-space data and raw (unprocessed) images.
- (ix) The main console should have facility and controls for music system for the patient in the magnet room.
- (x) There should be a provision of retrieval of the reconstruction data (raw files) in

a user friendly manner.

- (xi) Dicom interface to hook DICOM dry/laser camera capable of storing printing 1024x1024 matrix size images at least in 16 format without loss of digital resolution.
- (xii) The system should be capable to connect to PACS through RIS/HIS at no extra cost. Highest version of DICOM connectivity be provided.

b) Additional workstations: Price to be quoted separately.

- (i) Two workstations with color TFT display (19" or more) with evaluation capabilities as required for all applications in the tender. Server based workstations with 24,000 concurrent slice rendering capacity (like Advantage Windows/ syngo.via MR/Intellispace portal) It should be with preferably the same user interface as of main console.
- (ii) Hard disk of at least 120 GB for at least 250,00 image storage in 256 matrix < and 48 GB RAM capacity or more, with self-playing DVD/CD achieving facility.
- (iii) Separate licenses for all the post-processing (including special packages) should be provided for this workstation also as required for all applications in the tender.
- (iv) Specify clearly the algorithms that require extra license and list them whether these have been included or not.
- (v) Filming also should be possible from this workstation.
- (vi) The vendor should mention which are the post-processing capabilities that have been quoted with additional workstation.
- (vii) The workstation should display cardiac cine images in movie mode with rapid avi creation.

c) Networking:

- (i) The vendor should provide Level 3 network Switch (with 32 nodes) or latest, to integrate the network.
- (ii) Protocol - Ethernet TCP/IP standards - based image transfer with DICOM 3.0 over standard Ethernet IEEE 903 (DICOM send, receive and DICOM query modes).
- (iii) The vendor should provide the connectivity with PACS available currently in

the department.

- (iv) The network speed and cables should match the latest industry standards (eg.10BaseT/100BaseT/1 GB)
- (vi) System should be configured with different IP series, so as not to clash with different equipments already existing in the department.

d) Printing solutions:

- (i) DICOM interface to hook DICOM compatible, dockable, latest state-of-art Dry Laser Camera with integrated processor, of resolution more than 600 dpi, capable of storing/printing images of 1024 x 1024 (or higher, if available) matrix size in various matrix formats (including 16 format) without loss of digital resolution to be made available for filming from main console & workstation.
- (ii) It should be possible to connect other imaging modalities to the printer. 5000 compatible films to be provided.
- (iii) Colour Laser Network Printer (PCL6/PS) for printing of colour CSI/ Perfusion/BOLD maps and images on A4 size paper (minimum 24 ppm).

e) Data archival for long time storage:

Archival system (complete PACS server with hot swappable drives) for archival & storage of image and spectroscopy with a minimum capacity of 200TB

VI. IMAGING

a) Data Acquisition:

- (i) The system should have basic sequences package with Spin Echo, Inversion Recovery, Turbo Spin Echo with high turbo factor of 256 slice or more. Gradient Echo with ETL of 255 or more.
- (ii) Single Slice, multiple single slice, multiple slice, multiple stacks, radial stack and 3D acquisitions for all applications.
- (iii) Single and multi shot EPI imaging techniques with ETL factor of 255 or more
- (iv) The system should be capable of 2D and 3D acquisitions in conventional, fast & ultra-fast spin echo and gradient echo modes so that real-time online images can be observed if needed. All the sequences that are available with the vendor at the time of quote/ delivery should

- be provided as per their manual.
- (v) 2D multi-slice imaging should be possible in all planes (axial, sagittal, coronal, oblique and double oblique).
 - (vi) Up to 1024 × 1024 matrix acquisitions preferred for all applications. Wherever 2048 matrix available, please mention.
 - (vii) Half Fourier or other techniques to reduce scan acquisition time while maintaining adequate SNR
 - (viii) 3D volume, multiple contiguous slabs, multiple interleaved and multiple overlapping slabs
 - (ix) Fat suppression for high quality images both STIR and SPIR
 - (x) Slice thickness in 2D and partition in 3D to be freely selectable.
 - (xi) Dynamic acquisition (pre and post contrast scans and time intensity studies) with capability to initiate scan sequences either from the magnet panel or from the console. The number of repeat scans and the delay time (identical or variable) should be selectable
 - (xii) The system should acquire motion artifact free images in T2 studies of brain and whole body in restless patients also less sensitive to motions such as CSF and blood flow or flow artifacts, breathing, patient tremor or voluntary movements. Motion artifact free imaging with Diffusion weighted studies and musculoskeletal studies will be added advantage. This application to be compatible with PAT - like Propeller 3.0 Multivane XD, Presto, Snapshot, Blade etc. Body navigators may be offered as well.
 - (xiii) Auto-slice positioning from the localizer images.
 - (xiv) Maximum-off center positioning both anterior-posterior and lateral direction and should be selectable
 - (xv) Gating: physiological signals like ECG, pulse, respiratory, External signal triggering (interface for triggering input pulse from external source). The provision should be available at the console also (for fMRI, EEG, etc).
 - (xvi) Simultaneous acquisition, processing and display of image data in 2D multi-slice mode.
 - (xvii) Selection of voxels from oblique slices should be possible while doing spectroscopy
 - (xviii) Artefact reduction/imaging enhancement / image filtering /image

- subtraction / addition / multiplication/division techniques
- (xix) Flow: 1st and 2nd order flow artefact compensation
 - (xx) Presentation slabs: a number of relocatable saturation bands to be placed either inside or outside the region of interest
 - (xxi) Fat saturation techniques: frequency selective RF pulses to suppress fat signals in the measured image FOV. ROIselective (regional) fat suppression should also be given
 - (xxii) Magnetization transfer saturation: Off resonance RFpulses to suppress signals from stationary tissue in FOV
 - (xxiii) Phase contrast capability in 2D and 3D mode.
 - (xxiv) Image intensity correction Breath hold acquisition
 - (xxvi) EPI mode
 - (xxvii) DTI with MDDW or equivalent with a minimum of 12 and selectable up to 128 direction encoding
 - (xxviii) Data acquisition in all three standard planes (axial, sagittal, coronal) and oblique and double oblique planes or more oblique planes.
 - (xxix) Higher matrix acquisition capability in single shot EPI. Acquisition time, TR, TE and slice thickness should be clearly mentioned and supported by data sheet reference.
 - (xxx) Multi coil acquisition in order to optimize throughput increase and increased effective FOV. Individual acquisition elements of every coil should be mentioned.

b) Imaging Pulse sequences:

- (i) All standard and special pulse sequences available at the time of quote/delivery should be offered and quoted in the bid. If the vendor does not have any particular sequence/s but offers a work in progress (WIP) sequence/s, then it should be provided without any pre-condition. This also applies to any post-processing software that is offered which is WIP.
- (ii) The system should be capable of selecting TR and TEs as per requirement in majority of the pulse sequences.
- (iii) Spin echo (SE): multi-slice single echo, multi-slice multiecho (8 echo or more), SE with symmetrical and asymmetrical echo intervals and fast spin echo. MT-SE imaging sequence.

- (iv) Inversion recovery (IR): including short TI modified IRSE, FLAIR, DIR (Double Inversion Recovery).
 - (v) Gradient echo (GE): with transverse gradient/RF spoiling, and transverse gradient re-phasing, e.g., GRASE or equivalent etc. 3D gradient echo with shortest TR and TE, free choice of flip angle selection, while maintaining SNR.
- c) **Fast sequences:**
- (i) Fast spin echo and GE sequences in 2D and 3D mode with T1, T2 and PD contrast capable of acquiring maximum number of slices with a given TR a minimum TE, echo train should be at least 128 or more in fast spin echo mode
 - (ii) Half Fourier acquisition capabilities should be available with/without diffusion gradients and in combination with fast spin echo
 - (iii) Fast inversion recovery with spin echo
 - (iv) Fast gradient spin echo IR multi-slice multi-echo mode with maximum ETL. Sequences should incorporate RF focusing to acquire ultra-fast gradient spin echo.
Fast gradient echo sequence should incorporate RF spoiling and other technique to acquire images in ultra-fast 2D and 3D modes
 - (vi) Fat and water suppressed imaging sequences.
 - (vii) EPI optimized sequences (with and without fat suppression)
 - (viii) For T1, T2, PD imaging, perfusion, regular diffusion values (at least 5b values, 3 directions) EPI-FLAIR, EPI-IR, EPI-FLAIR diffusion tensor, EPI-MT-FLAIR, tensor diffusion (at least 16 bvalues, and 128 directions) and diffusion studies. Suitable artefact/ fat suppression techniques to be incorporated in the sequence to have optimum image quality.
 - (ix) There should be capability of calculating ADC map (isotropic and anisotropy from the regular diffusion and tensor data).
- d) **Optimized sequence Packages:**
(To be provided as part of standard equipment with no additional cost)

1) **Neuro:**

- (i) All T1 (2D, 3D), T2 (2D, 3D), IR (2D, 3D), Dual IR (2D, 3D) sequences
- (ii) Sequence for internal ear imaging for visualization of fine structures like cranial

nerves (appropriate sequences like CISS, etc or equivalent). Mention the sequences provided.

- (iii) 3D sequences for internal auditory canal imaging
- (iv) Dynamic imaging of pituitary using appropriate sequence
- (v) Whole spine T1, T2, IR sequences
- (vi) Whole neuro examination with automatic planning, scanning and post-processing, with single localiser positioning, without changing the coils/ repositioning

2) Angiography:

- (i) MR angiography: 2D/3D TOF, 2D/3D Phase contrast (with and without gating) and magnetization transfer saturation, black blood angiography for cerebral, pulmonary, abdominal and peripheral vessels.
- (ii) For peripheral moving table angiography should be offered covering hip to limbs to be examined in one go with high resolution and high SNR.
- (iii) Bolus tracking software package.
- (iv) Sequences for breath hold angiography with contrast enhancement.
- (v) Sequences for time resolved angiography with contrast kinetics.
- (vi) ECG triggered non-contrast angiography
- (vii) Contrast bolus tracking (including single shot whole body MRA, interactive and automatic tracking, etc.).
- (viii) Perfusion study in organ systems like kidney, brain, heart etc. with T1 perfusion with permeability maps, and quantitation of rCBF/ rCBV, MTT, etc, with color maps.

3) Cardiac package:

- (i) Full comprehensive cardiac sequences which includes, MR cardiology package for evaluation of heart in long and short axis with black blood cardiac imaging.
- (ii) Package for coronary artery imaging including sequences for motion compensation - prospective and retrospective gating, etc.
- (iii) EPI based sequence for stress perfusion MRI including ability to adjust the cardiac phases required with increasing HR
- (iv) Myocardial tagging sequence
- (v) 2D and 3D Sequences enabled with delayed enhancement

- (vi) 3D sequence of cine (bright blood & dark blood options)
- (vii) Rapid acquisition of heart using acceleration techniques
- (viii) STIR sequence for cardiac use
- (ix) 3D whole heart sequence (with & without contrast for coronary imaging)
- (x) Ability to acquire multiple arterial and venous phases on CE MRA.
- (xi) Quantitative flow analysis software
- (xii) 3D acquisition of whole heart in one breath-hold
- (xiii) 4D TRAK/ TRICKS-XV/ TWIST/ equivalent (with maximum FOV)
- (xiv) Provision for timing /Stop watch (MR compatible) for timing drug infusion

4) Diffusion / DTI:

- (i) Sequence package for diffusion including DTI (tractography) study in organs like brain, kidney, muscle, heart, spine, breast, prostate, etc.
- (ii) There should be capability of calculating ADC map (isotropic and anisotropic from the regular diffusion and tensor data).
- (iii) MR diffusion tensor imaging package with tractography
- (iv) MR neuro functional imaging sequence package (incl. Mosaic, etc)

5) Body imaging:

- (i) Flow quantification in vessels and CSF, hepatobiliary system.
- (ii) Fly-through facility with Flow analysis including display of various velocity values.
Optimized breath hold sequences for abdominal studies including angiogram.
- (iv) MR Cholangiography and Pancreatography: Specialized sequences and processing to perform MRCP.
- (v) Pulmonary 2D/3D MRA sequence, including single breath hold sequence.
- (vi) MR ventriculography, cisternography, myelography.
- (vii) Single sequence to acquire four different contrast (inphase, out-of-phase water only, fat only). The same technique should be used in other sequences, for dynamic angiography/T1 quantitative analyses.
- (viii) Parallel acquisition techniques including new sequences.
Specify the technique used and the factor by which the acquisition time is reduced for similar acquisition with and without parallel imaging technique.
Mention the sequences.
- (ix) Flow quantification packages for CSF with dynamic CSF flow imaging,

aqueduct and spinal canal.

- (x) Radial/Spiral pulse sequences for ultrafast imaging.
- (xi) Suitable artefact/fat suppression techniques to be incorporated in all the sequences to have optimum image quality.
- (xii) A sequence for differentiation of fluid and cartilage in ortho applications (sequence like DESS or equivalent)
- (xiii) Susceptibility artefact correction techniques to be incorporated in all the sequences to have optimum image quality.

6) **SWI**: Sequences for susceptibility imaging

7) **Prostate imaging**: Sequences for imaging of prostate

8) **Breast imaging**: Sequences for imaging of breast (including sagittal bilateral breast imaging in a single acquisition)

9) **Motion correction**:

- (i) Sequence for in-line motion correction for uncooperative patients/ children (with software and acquisition - sequences like BLADE, PROPELLAR, Multivane or equivalent)
- (ii) Sequence with ultra-short TE
- (iii) Sequence for nullifying CSF pulsation artifacts
- (iv) Sequence enabling prospective motion correction in quick time and in real time during f MRI.

10) **MR Spectroscopy**:

- (i) System should have capability to perform multiplanar proton and phosphorous spectroscopy (31P)
- (ii) Proton MRS Sequence for single-voxel acquisition, with selectable fat/lipid saturation bands, options of water saturation (eg. VAPOR, CHESS, etc) with all post-processing software
- (iii) Proton Multi-voxel CSI [2-D and 3-D] acquisition and metabolite mapping with all necessary RF sequences (and post processing algorithms) with all post-processing software
- (iv) If separate coils are needed for carrying out MRS, it should be provided.
- (v) Sequences for phosphorous single voxel and multi-voxel spectroscopy should be provided, with all post-processing software

- (vi) RF sequences for cardiac, prostate, breast, liver, musculoskeletal and brain (if there are any specialised/optimised sequence available, the same should be offered)- with all postprocessing software
- (vii) Water and lipid suppression in automated sequences.
- (viii) The pulse sequences for 31P MRS and 1H MRS for liver, cardiac spectroscopy, etc. should have external gating provision with triggering bases on ECG/ Respiratory, with all postprocessing software.
- (ix) Please give details of spectroscopic packages offered and not offered (but available with the vendor).

11) Others:

- (i) Sequence employing arterial spin labelling (ASL) technique
- (ii) Whole body imaging (using body coil and surface coils)
- (iii) Whole body diffusion weighted imaging (using body coil and surface coils)
- (iv) Automated fusion and composing for the above two (without any artefacts)
- (v) Volume acquisitions for neuro applications.
- (vi) Head first or feet first for all imaging
- (vii) Single sequence for Triglyceride Fat and iron quantification software for Liver should be offered - Like IDEAL IQ, Liver LAB, mDixon Quant
- (viii) MR Elastography - offer hardware and software
- (ix) Software for fusion imaging CT/PET/SPECT/USG
- (x) T2 cartilage coloured mapping sequences for early detection of osteoarthritis like Map IT, Cartigram, T2 Mapping

12. Post Processing and evaluation:

- (i) Licences of all the post processing and evaluation packages should be provided for the main and additional console/ workstation
- (ii) Specify clearly number wise the algorithms that need licenses and a statement whether these have been provided in both the main console and the additional workstation (satellite console/ extended workspace)

13. Special Application Packages:

The vendor must provide their specialized and optimized imaging sequences with post-processing packages of those listed below along with the standard or basic package at no additional cost.

Please quote prices separately if the packages are not part of standard equipment.

Please list other applications available with the Vendor, which are part of standard or basic package.

1) **MPR:**

- (i) Multi-planar reconstruction (MPR) in any arbitrary plane including curved planes with freely selectable slice thickness and slice increments.
- (ii) Surface Reconstruction and evaluation on reconstructed images with minimum time.
- (iii) MIP in displaying in cine mode 2D and 3D mode, targeted/ segmented MIP in any orthogonal axis with minimum processing time and capable of displaying in cine mode.

2) **Cardiac evaluation package:**

- (i) Full cardiac evaluation: Operator selective or automatic contour mapping and calculation of cardiac parameters like wall thickness, stroke volume, Ejection Fraction, filling rate, myocardial wall motion including display of data in table, graph and in cine mode, blood flow quantitation, velocity mapping, pressure gradient quantitation, shunt quantitation, regurgitation calculation, stenosis, blood flow, etc. These should be usable on main as well as on additional workstation/ satellite console.
- (ii) Semi- and fully quantitative analysis of myocardial perfusion
- (iii) Quantification of myocardial scar vs total myocardium volume / mass on DE imaging
- (iv) Analysis of both RC and LV volumes mass and ejection fraction to be available separately; should be possible to do manually and in fully automated manner
- (v) Non contrast coronary angiography with single vessel approach/whole heart approach
- (vi) Sequences and post-processing for T1, T2 mapping
- (vii) T2* imaging with quantification

3) **ADC, perfusion:**

- (i) Evaluation and display of diffusion images, ADC map, fMRI in reference of EPI optimized sequence.
- (ii) Perfusion image evaluation with time intensity graph and other statistical parameters

- (iii) Evaluation package for calculating rCBV, rCBF, MTT, perfusion map, Corrected CBV calculation; Fusion of perfusion map with Contrast enhanced 3D T1 images etc. Mention the package/software offered with brochure.
 - (iv) Flow quantification and evaluation for vascular (high & low) CSF, bladder outlet and cine display.
- 4) **BOLD analysis:** Evaluation of functional images of brain with appropriate statistical algorithms, colour display and overlay on base anatomical images.
 - 5) **VBM:** Voxel-based morphometry for segmentation and quantification.
 - 6) **Tractography:** Post-processing package for DTI and Tractography, estimation of ADC, FA (Lamda- parallel, perpendicular separately and combined), Fiber tracking, fiber statistics, and display of fiber tracks on anatomical image(s).
 - 7) **Co-registration:** Superimposition on Neurotractography geometry and tensor diffusion field on both functional BOLD mapping and neurometabolite (CSI) mapping.
 - 8) **Image statistics:**
 - (i) Measurement of distance, area, volume, angle, mean, SD, image addition, subtraction, multiplication, division, interpolation, segmentation, threshold, histogram.
 - (ii) Image filtering and Image fusion software.
 - (iii) Software for co-registering MRI/fMRI/MRS/Metabolite mapping images with images from CT, PET, and SPECT.
 - (iv) Evaluation features like zoom, rotation, scroll, roaming, image synthesis, multi point T1 and T2 calculation (more than 8) window stretching, text dialogues graphics, sorting, searching, archiving, recalling etc.
 - 9) **Spectroscopy:**
 - (i) Full post-processing for single-voxel MRS, CSI (multivoxel MRS), metabolite mapping with colour coding (metabolic images), 31P MRS post-processing, etc., for brain, breast, prostate and for other applications.
 - (ii) Post processing should include FFT, base line correction, curve optimization, automatic phase correction, metabolite imaging, spectral mapping, magnetic resonance spectroscopic imaging (molecular imaging) with naming and peak integral values for all in-vivo metabolites
 - 10) **Functional MRI accessories and post-processing:**
 - (i) Functional Imaging with package for BOLD imaging and processing package

(capable of real-time processing and display of colour overlay (in real time) using 32-channel Head coil being supplied with the system.

- (ii) Complete fMRI solution including audio-visual projection (3D capable) system, with headphones with very good noise suppression (> 30dB) (Preferable to have LCD/LED monitor for projection).
- (iii) Binocular eye tracker cameras, integrated with the visual system (preferable to have separate wearable eye-tracker cameras)
- (iv) The audio-video projection system should have the capability to project 3D images / movies to the subject, and should be compatible with the 32-channel head coil, and should all attachments that may be required for complete integration.
- (v) The system should be integrated with stimulus presentation/ paradigm generator software, along with permanent license (like Superlab, eprime, Presentation, etc), which is capable of presenting audio-visual picture, audio, video (multiple formats), etc.
- (vi) The paradigm presentation should be synchronised with the scanner (for starting along with measurements)
- (vii) Integration (and Provision near the console) for external trigger (of the sequence) for synchronising fMRI acquisition with paradigm
- (viii) Provision for serial ports and DB15 ports in the penetration panel for routing SVGA/EEG connections (one each for customer use).
- (ix) FMRI console should have all relevant functions to develop and integrate the paradigm, to deliver the paradigm and also, to monitor the task being presented. The volume control option should be available with the operator (at a convenient place at the console)
- (x) Post-processing workstation / server with post-processing software and hardware associated, with licenses for processing the BOLD data (with required licensed operating platform required like MATLAB, IDL, etc)
- (xi) The system should have integrated MR compatible binocular eye-tracker (binocular), along with eye-tracking software at the console (on a separate PC/laptop).
- (xii) The entire fMRI hardware package should be from a single vendor for complete integrated solution

- (xiii) Brain voyager post-processing software (along with permanent license)

VII. QUALITY ASSURANCE AND PHANTOMS

- (i) Phantoms for routine quality assurance for all coils (including body coil)
- (ii) Chemicals for spectroscopy phantoms for important short echo time neuro metabolites (NAA, Cho, Cr), breast (Cho, Cr) and prostate (Cho, Cr, Citrate, polyamines) with appropriate MR compatible phantom containers.
- (iii) Quality assurance as per AAMP standard for SNR for different coils and nuclei, spatial resolution, magnetic field inhomogeneity, eddy current compensation, RF power and inhomogeneity measurement. Specify the details of the QA package.

VIII. SAFETY FEATURES:

- (i) The magnet system should include an Emergency Ramp Down unit ERDU for fast reduction of the magnetic field with Ramp Down time below 3 minutes.
- (ii) The magnet should have quench bands that contain the fringe fields to a;
- (iii) specified value in the event of a magnet quench
- (iv) Real time SAR calculation should be performed by software to ensure that RF power levels comply with regulatory guidelines and are displayed on each image
- (iv) The system shall have manual override of the motor drive for quick removal of the patients from the magnet bore.
- (v) Temperature sensor (built in) for magnet refrigeration efficiency must be provided

IX. STANDARD MRI ACCESSORIES:

- (i) Hand held metal detectors (2 Nos.)
- (ii) MR compatible (minimum 2000 Gauss line) cardiac and physiological monitor (ECG, NIBP, SPO2,) for neonates/ infants and adults (with all accessories for five years) (MedRad/In-vivo/better models)
- (iii) MR compatible (minimum 1000 Gauss line) syringe/infusion pump. (MedRad/MRIdium/better models)
- (iv) MR Compatible Dual Pressure injector (minimum 2000 Gauss line) (MedRad / better models)
- (v) MR compatible anaesthesia machine (for Paediatric and adults use) with dual vaporisers (for isoflurane, halothane),and other accessories (minimum 1000

Gauss line). (Penlon/ Leon/better models)

- (vi) Non-ferromagnetic patient transfer trolley of international make should be provided (2 No's) & non-magnetic wheel chair - 1 no's
- (vii) Provision for external trigger (of the sequence) near the console
- (viii) Provision for serial ports and DB15 ports in the penetration panel for routing SVGA/EEG connections (one each for customer use).
- (ix) Two quantity: Non-magnetic IV stand
- (x) Two quantity: Digital Patient Weighing Scale (in the range between 0 to 200 kg)
- (xi) MR compatible storage carts and wall mounted cabinets.
- (xii) Water chillers for Cold Head/Gradients
- (xiii) Coil cabinets to be provided.
- (xiv) Network cable and other required materials for the complete installation to be provided by the supplier
- (xv) Ferromagnetic object detector and alarm (No door frame required), at the entry of MRI suite

X. OTHER ACCESSORIES:

- (i) Six Revolving chairs with ergonomic support
- (ii) Table for the MRI console, MRI additional console/ workstation, fMRI workstation Necessary Desk, chair and Rack for the PACS Server & Workstation to be provided by the supplier
- (iii) All the necessary interconnecting interfaces, cables, modules and other hardware and software to fully integrate the system for full operational status.
- (iv) Uninterrupted power supply (UPS) with sufficient capacity (appropriate rating as required with a minimum of 200 kVA or more UPS) for 30 minutes back up of the full load MR system and its accessories during patient MR imaging.
- (v) PACS system should be connected to the UPS (If a separate UPS is required for this purpose, this should be provided)
- (vi) Two (quantity) MR compatible oxygen cylinders (for the anaesthesia system)
- (vii) Good quality air curtain at MRI entrance (for patient entry), to filter the dust and prevent the leakage of a/c.

XI. SUITABLE RF ENCLOSURE

RF cabin: the system should be supplied with imported RF cabin with RF room

shielding. RF Door Screen, and interiors for the same should be carried out suitably

XII. INSTALLATION

A list of installations of the quoted model in Government / Private institutions in India to be provided. Satisfactory service back up facility should be available. Please mention the service centre available nearest to the site of installation and total No. of MRI service engineers employed in India.

XIII. APPLICATION SUPPORT

Applications specialist should be available on site for a total duration of 6 weeks to train the technical staff. Advanced training to be provided by the vendor at the site for Faculty, Residents, students and Radiographers, so as to benefit the latest applications available on the system. Please provide certificate from existing installed base for proof of providing application support time to time

XIV. STANDARD AND CERTIFICATION : Should be FDA or CE approved product.

XV. GUARANTEE: The vendor should guarantee the service and spare support for 10 years of the system including Helium and cold head and all accessories

XVI. DOCUMENTS: The offer should be accompanied by Original data sheet of the product.. Refurbished Units will not be accepted. Verification Certificate for authorized reputed third party agencies to be attached. Incomplete data sheets and offers which are speculative are not to be submitted.

XVII. TURNKEY OFFER:

- (v) Turnkey offer for Civil, Electrical and A.C works etc. are to be given after visiting the place of installation for which the bidder has to make a request. Please give drawings for civil, electric, and any other required for installation and reporting. Prior approval to be taken from the department

Specifications for Turnkey Work

Carpet Area in the Scope of work: 1000 Sq. feet

The room should be renovated according to the need of the specific machine. The indicative renovation work to be undertaken is noted below:

- i) **Painting of the rooms** – 2 coats synthetic enamel paints over 2 coats primer over wall putty (if required) wherever required
- ii) **Viewing Window: Viewing Window:** RF shielded Viewing window.

- iii) **Floor:** Floor (except of MRI room) should be of premier quality double charged joint less vitrified mirror polished tiles.
- iv) **Wall (except MRI room):** Walls should be of premier quality double charged joint less vitrified mirror polished tiles up to false ceiling other than the marble finished wall. Necessary replacement / repair of damaged marble before polish.
- v) **Ceiling:** Ceiling should be of Mineral fiber board with aluminum grid. 2/3 coats of distemper on true ceiling. / High quality Ceiling paneling as suitable with existing infrastructure.
- vi) **RF Cabin:** The RF cabin should be imported type. Required RF shielding should be as per the guidelines of the competent authority
- vii) All existing doors in the designated area are to be replaced with a new one.
- viii) **Height of the room (up to false ceiling):** 3.0 m and above
- ix) **Specification of Doors:**
Main Doors: Steel Door
Other Doors:
- x) Necessary electrical fittings are to be installed.
- xi) High quality room lighting (LED up to 400 LUX of illuminance)
- xii) Earthing: Two nos. Copper plate earthing as per PWD schedule

Wiring System:

- a) Light, Fan, 5 Amp Plug: 3 X 1.5 sq. mm copper conductor FRLS wire should be provided.
- b) Power Plug (15 Amp): 2 X 2.5 + 1 X 1.5 sq. mm copper conductor FRLS wire should be provided.
- c) Split AC wiring: 2 X 4 + 1 X 2.5 sq. mm copper conductor FRLS wire should be provided. Necessary electrical earthing facility is to be prepared.
- xiii) False ceiling / High quality Ceiling paneling should be provided in all rooms.
- xiv) Medical Gas Pipeline system [O₂, N₂O, Air (4 Bar) and Suction] with imported MRI compatible outlet points along with matching adapter etc. should be provided in MRI room and in patient waiting area. Inside pipeline in the Gantry room to be completed in all respect and the entry points of the pipelines should be terminated at a suitable place outside the Gantry room with medical grade isolation valves.

Air Conditioning

The entire carpet area of 1000 sq. feet should be air conditioned. The bidder should consider the tonnage calculation with 100% backup. The Air conditioning machines should be run on timer. High quality Mitsubishi / Hitachi / O-General make (Green Gas) Air Conditioning machine with 5 star rating should be installed.

Furniture item to be supplied:

- a) Executive revolving chair with arm rest: 6 Nos. (Godrej / Featherlite or equivalent)

- b) Storage Cupboard: 3 Nos. high quality customized Wall mounted with locking facility
- c) Shoe Rack: 2 No
- d) Corner Table: 2 Nos. (Godrej / Featherlite or equivalent)
- e) Change room with Mirror
- f) Kg Fire extinguisher cylinder: 4 Nos.
- g) Console Table from OEM
- h) Metal Detector-Hand Held: 01
- i) Music System(Sony/Philips/Bosch): 01
- j) LED view box for four films: 01
- k) MRI compatible Patient trolley with mattress, side rails, oxygen cylinder and fluid stand attachment: 2 Nos.
- l) MRI compatible Wheel Chair with mattress, side rails, oxygen cylinder and fluid stand attachment: 2 Nos.
- m) Dehumidifier (22 Ltrs): 4 Nos.
- n) Patient waiting chair for at least 30 nos.

1. STANDARD REQUIREMENTS

The following requirements with regard to inspection, quality, packing, warranty, maintenance and related services shall commonly apply to all the goods in all the Schedules:

i. QUALITY CERTIFICATION

Where ever appearing in the bid document, the "CE certificate" shall be read as: "CE mark for *conformité européenne*, (French for "European conformity").

ii. WARRANTY

Unless specified otherwise, warranty shall always be for a period of one (1) year from the date of acceptance after commissioning of the goods. During warranty, cost and responsibility of the transport/shifting of the equipment, in case so required for repair, etc, shall be entirely borne by the Supplier, without any liability on the consignee. In case of such shifting of equipment, alternative working equipment shall be first made available to the consignee to avoid any disruption in the clinical work

iii. MAINTENANCE

- a. CMC shall be as per the specification after the expiry of warranty, unless specified otherwise.
- b. During CMC, cost and responsibility of the transport/shifting of the equipment, in case so required for repair, etc, shall be entirely borne by the Supplier, without any liability on the


consignee. In case of such shifting of equipment, alternative working equipment shall be first made available to the consignee to avoid any disruption in the clinical work.

c. Subject to (b) above, CMC services shall be provided at the site of the equipment, within the prescribed response time.

iv. Labelling & Packing

The equipment should have a sticker on it with the following information:

- a. Procured by: WBMSCL Ltd.
- b. Serial Number:
- c. PMS Done:
- d. PMS Due:
- e. Toll Free No.:
- f. Mobile No.:
- g. Email:
- h. Facility Asset No.:
- i. Warranty upto:
- j. CMC Starts on:
- k. CMC Valid upto:
- l. Approved CMC Rate per annum:
- m. Service Engineer Contact detail:
- n. Standard format of sticker is attached here (N.B: Bidders are advised to approve the final format after discussion with WBMSCL officials)

		Procured by : West Bengal Medical Services Corporation Limited (WBMSCL)	
Reference No. : WBMSCL /NIT- 184 /2018, Dated-13.09.2018			
Machine Serial No. :		Facility Asset No. :	
Warranty (Xyrs) upto :		Approved CMC Rate:	per annum
CMC starts on :		CMC valid upto :	
Last PMS on :		Next PMS due on :	
Complaint logging at :	✉: abc@abc.com, XXXX	☎: (0XX) XXXX XXXX / XXXX	
Service Engineer :	Mr. XYZ, 09XXXXXXXXXX	✉: abc@abc.com,	☎:
Service Manager :	Mr. XYZ, 09XXXXXXXXXX	✉: abc@abc.com,	☎:

2. LIST OF RELATED SERVICES

i) Incidental Services

The supplier may be required to provide any or all of the following services, including additional services, if any, specified in SCC:

- (a) Performance or supervision of the assembly, installation and/or start-up of the supplied equipments
- (b) Furnishing of tools required for assembly and/or maintenance of the supplied Goods along with each equipment
- (c) Furnishing of detailed operations and maintenance manual for each appropriate unit of supplied equipments at the time of delivery
- (d) Successful tenderer shall be required to give an undertaking that suitable trained service engineers shall be posted at three places of West Bengal for providing prompt, effective and preventive maintenance during the period of warranty as defined as well as CMC period.
- (e) The Comprehensive maintenance Contract (Including Spare Parts)
 - (i) The Purchaser/ Consignees/ Government of WB, may, at his own and sole discretion enter into a Comprehensive Maintenance Contract (CMC) with the Supplier at the contracted price.
 - (ii) The supplier shall visit each consignee site as recommended in the manufacturer's technical/ service operational manual, but at least once in three months during the CMC period for preventive maintenance.
- (f) Training of the Purchaser's personnel, on-site, in assembly, start-up, operation, maintenance and/or repair of the supplied Good. This must be carried out at the time of commissioning of Equipment.

ii) Availability of Spare parts

Suppliers shall ensure the availability of spare parts for 10 (ten) years. Inventory of the Spare parts required for 8 years.

5. INSPECTIONS AND TESTS

- a) The Vendor shall get the Goods inspected in the manufacturer's works by a competent authority and submit a test certificate and also a guarantee/warranty certificate that the Goods conform to laid down specifications.
- b) WBMSCL or its representative may inspect and/or test any or all item of the Goods to confirm their conformity to the Contract, prior to dispatch from the manufacturer's premises or at the port of entry.

Section V. Bidding Forms

FORM 1

Tender Form

(To be furnished on non-judicial stamp-paper of Rs. 100/-, affirmed before a First Class Magistrate/ Notary/ Executive Magistrate)

Date:

Bid Reference No.: _____ Schedule - ____

Name of Contract: Supply and Commissioning of *[Name of the equipment]* for the 3.0 TESLA MRI MACHINE at Hospitals and Medical Colleges of the Govt. of West Bengal

To
Managing Director,
West Bengal Medical Services Corporation Ltd,
Swasthya Sathi,
GN- 29, Sector – V, Salt Lake,
Kolkata - 700 091

Sir,

I/We, the undersigned hereby accept all the terms and conditions of the Bid Reference No.: **WBMSCL/NIT-184 /2018, Schedule -, Dated -13.09.2018** and its Amendments and Addendum thereto are read and accepted without any modification or condition(s). We now offer to Supply and Commissioning of Medical Equipments for *[Name of the equipment]* for the Hospitals and Medical Colleges of the Govt. of West Bengal in conformity with your above referred document.

We also

1. certify that:

- a) We have examined and have no reservations to the Bidding Documents, including Addenda issued in accordance with Instructions to Bidders.
- b) The offered products are in accordance with the required specifications and technical requirements
- c) Our Bid consisting of the Technical Bid and the Price Bid shall be valid for a period of 120 days from the date fixed for the bid submission deadline in accordance with the Bidding Documents. However, the prices quoted by us and accepted by WBMSCL shall hold good and remain valid for a period of **2(two)** year from the date of signing of the contract and no additional claims will be made on account of any price variation or

fluctuation in market rates. The rate quoted shall remain binding upon us and may be accepted at any time before the expiration of **2(two)** year.

- d) If our bid is accepted, we will submit a performance security within 14 (fourteen) days of issuance of Award of Contract (AOC) in the prescribed format as given in the bid document.
- e) Our company has been incorporated in accordance with the laws of India and governed by them.
- f) Our Company have commissioned _____nos. of the offered model and providing necessary service support to the equipments.
- g) We have never been blacklisted by any Government Department/ Agency in India during the last 5 years.
- h) There is no adverse report against the equipment offered by us in any Govt. institution.
- i) We will permit WBMSCL or its representative to inspect our accounts and records and other documents relating to the bid submission and to have them audited by auditors appointed by the WBMSCL.
- j) All the statements made in the attached documents are true and correct. In case of any information submitted proved to be false or concealed, penalties shall be imposed in terms of the bid document.

2. understand that:

- (i) Partial or incomplete bid submission will lead to cancellation of our bid.
- (ii) The tender inviting and accepting authority can vary quantity up to 40% (forty percent) above or 40% (forty percent) below the required quantity under this tender.
- (iii) The tender inviting and accepting authority reserves the right to reject any application without assigning any reason.

Enclose:

- 1. Non Statutory Documents/ My Documents
- 2. Statutory Documents (Bid A & Bid B)
- 3. Forms & Annexure duly filled up, signed & notarized (where applicable)

Name.....

In the capacity of.....

Signed

Duly authorized to sign the Bid for and on behalf of (if applicable).....

Date.....

Form 2: CHECK-LIST

[Please fill in and include with your Bid]

Note 1: It is essential that all documents in hard copy are to be placed before the Committee and arranged in the same sequence as given in the Check List. All the documents should be appropriately flagged.

Note 2: If any document is written in any language other than English, an English translation of the document duly authenticated is to be submitted.

Note 3: All the documents mentioned below are essential for qualifying in the technical evaluation.

Note 4: After opening of the technical bids, if it is found that any of the documents required to be submitted with the bids is wanting, WBMSCL shall reserve the right to allow late submission of such document at its discretion within a specified time limit.

Non statutory documents to be submitted under <u>My Document</u>				
Sl. No.	Activity	Yes/No/NA	Page No in the Bid	Remark
1	PAN Card			
	15 – digit Goods and Services Taxpayer Identification Number (GSTIN)			
2	Service Tax registration Certificate			
3	VAT Registration Certificate			
4	CST registration Certificate			
5	Import License			
6	License from Government/ Statutory Authority as applicable OR Registration with the Registrar Of Companies, if applicable.			
7	a) Manufacturing Licence (National/International). (In case, manufacturing licence is not required / applicable for production of the quoted item, notarized declaration from the manufacturer is to be submitted) b) Current registration as SSI (if any)			
8	Performance Statement Form (For the period of last three calendar years ending December 2016) - Form 6 of Section V			
9	ISO certificate of the manufacturing factory			
10	Income Tax returns for assessment year 2014-15, 2015-16 & 2016-17			
11	Acknowledgement of VAT Returns for 2016-17/ VAT Clearance Certificate.			
12	Acknowledgement of CST Returns for 2016-17/ CST Clearance Certificate.			
13	Legal declaration affirmed before a First Class Magistrate / Notary on non judicial stamp paper of			

	Rs. 100/- for, Acceptance of Terms and Conditions of NIT and its Amendments and Addendums thereto. (As per Form 9 of Section V). <i>Note: Technical evaluation of the bid will be taken up only after scrutiny of Form 9 (NIT Acceptance Form) duly notarized.</i>			
14	Legal declaration affirmed before a First Class Magistrate / Notary on non judicial stamp paper of Rs. 100/- as per Form – 7.			
15	Manufacturer's Authorization (If applicable) as per Form no. 5 of Section V			
16	List of installations & commissioning in India of offered model. (self declaration)			
17	Satisfactory Performance Certificate from at least 3 (three) users of the quoted model in support of the satisfactory operation in India.			
18	P/L & Balance sheet 2014-2015			
19	P/L & Balance sheet 2015-2016			
	P/L & Balance sheet 2016-2017			
BID - A				
Sl. No.	Activity	Yes/No /NA	Page No in the Bid	Remark
20	Earnest Money Deposit (EMD)/ Bid Security in the form of Demand Draft (DD) or Bank Guarantee (BG)			
21	Declaration of the bidder on letter head that "We agree to submit a copy of the Tender Documents and its Amendments and Addendums thereto duly initialed by us in all pages with our seal/ rubber stamp affixed thereto, in token of acceptance thereof."			
BID - B				
Sl. No.	Activity	Yes/No/NA	Page No in the Bid	Remark
22	Model of the equipment offered for (Self Declaration) with Technical Data Sheet			
23	Compliance Sheet of the Technical Specifications (Form No. 3 of Section V) duly signed & stamped by Authorised signatory of the Manufacturing company			
24	2 sets of Brochure of the offered product / model.			
25	CE ("Conformité Européene") Certificate & US FDA approval Certificate of the offered model (as applicable) CE ("Conformité Européene") Certificate should be from EU Notified Bodies authorized to conduct audits			
26	Self declaration by the Tenderer for agreeing that WBMSCL will do Quality Check by Third Party Agency, If required			
27	AERB approval certificate of the offered model			
28	Confirmation from manufacturer of the offered equipment that all the facilities exist in his factory for inspection and testing and these will be made			

	available to WBMSCL or his representative, if inspection is considered necessary.			
29	Pre-requisites of installation like [Power (KVA, Phase, Hz) and any other requirement, if any].			
30	Average Annual Turnover of the Company in medical equipment division during the last 3 Financial Years 2014-15, 2015-16 & 2016-17 (in INR) - to be certified by practicing Chartered Accountant as per format given in FORM 11			
31	Form 12: Declaration of Quality Certification of Equipment (as applicable)			

Form 3a: Consignee Receipt Certificate (CRC)

(To be issued by consignee's authorized representative)



West Bengal Medical
Services Corp. Ltd.

[The consignee may issue an additional challan receipt if delivered by courier or transporter]

Date of supply by the Company Person or Courier:	
Name and Address of the Consignee:	
Name of the item supplied (with Make & Model & Model No.):	
Purchase Order / Contract No.:	
Name of the Supplier:	
No. of Units supplied:	
Place of destination (The dept. where the equipment will be actually installed):	
Invoice No. & Date:	
Details of Batch / Serial Numbers, if any of item supplied:	
<p>..... (Signature & Office Seal of authorized representative of Consignees with date) [Name and designation of the signatory to be written capital letter]</p>	
<p>..... (Signature & Office Seal of Head of the Institute / Hospital with date) [Name and designation of the signatory to be written capital letter]</p>	

Form 3b: Satisfactory Installation Certificate (SIC)

(To be issued by the consignee after successful commissioning of equipment)



West Bengal Medical
Services Corp. Ltd.

Bid Reference:

Award of Contract Reference:

Description of Equipment/Service:

Date of Commissioning:

This is to certify that the equipment(s) as detailed below has/have been received in good condition along with all the standard and special accessories, consumables, set of spares in accordance with the contract/technical specification of the equipment and site preparation including interiors as per bid document.

Details of equipment, accessories, consumables, spares, etc.

Sl. No.	Description	Quantity	Serial No. / Part No.
1			
2			
3			
4			
5			
6			
7			
8			
9			
10			

In case of space deficiency, another sheet with the same format can be annexed.

The supplier has also submitted the following,

1. Tools for maintenance
2. Detailed operation and maintenance manual both in hard and soft copy for each item of supply at each location

The proving test has been done to our entire satisfaction. The equipments, its accessories and ancillaries of the site preparation including interiors is functioning satisfactorily and faultlessly

Declaration by Unit Head (HOD/MO-IC/Others):

Sticker designed by WBMSCL is fitted with the equipment Yes No

Signature :

Name (in Block):

Designation :

P.T.O. →

The following operators/ end users have been trained to operate the equipment(s),

SI	Name	Designation	Contact No	E-mail ID (In CAPS)	Remarks
1					
2					
3					
4					
5					
6					
7					
8					
9					
10					
11					
12					
13					
14					
15					

In case of space deficiency, another sheet with the same format can be annexed.

Countersigned by the head of the institute/ hospital:

Signature _____ Name _____

Designation with Stamp _____

Date _____ Phone No: _____

Signature of Unit Head :
(HOD/MO-IC/Others)

Name (in Block) :

Designation with Stamp :

Form 4: TECHNICAL SPECIFICATION FORM
(Technical Compliance Sheet)

Tenderers must complete the right column of the below table and the compliance confirmation statement as included in Section IV, Schedule of Requirements; Technical Specifications.

Schedule No. ____ :

<i>WBMSCL's Specification</i>	<i>Compliance</i>

THE OFFERED PRODUCTS ARE IN ACCORDANCE WITH THE REQUIRED SPECIFICATIONS AND TECHNICAL REQUIREMENTS:

YES NO

ANY DEVIATIONS MUST BE LISTED BELOW:

Form 5: BID SECURITY (BANK GUARANTEE) FORM

[Insert: No Bid Security is requested or The Bank shall fill in this Bank Guarantee Form in accordance with the instructions indicated.]

[Bank's Name and Address of Issuing Branch or Office]

Beneficiary: _____ *[Name and Address of WBMSCL]*

Date: _____

BID GUARANTEE No.: _____

We have been informed that *[name of the Tenderer]* (hereinafter called "the Tenderer") has submitted to you its bid dated (hereinafter called "the Bid") for the execution of *[name of contract]* under Invitation to Bid No. *[NIT number]* ("the NIT").

Furthermore, we understand that, according to your conditions, bids must be supported by a bid guarantee.

At the request of the Tenderer, we *[name of Bank]* hereby irrevocably undertake to pay you any sum or sums not exceeding in total an amount of *[amount in figures]* (*[amount in words]*) upon receipt by us of your first demand in writing accompanied by a written statement stating that the Tenderer is in breach of its obligation(s) under the bid conditions, because the Tenderer:

- (a) has withdrawn its Bid during the period of bid validity specified by the Tenderer in the Form of Bid; or
- (b) having been notified of the acceptance of its Bid by WBMSCL during the period of bid validity, (i) fails or refuses to execute the Contract Form; or (ii) fails or refuses to furnish the performance security, if required, in accordance with the Instructions to Tenderers.

This guarantee will expire: (a) if the Tenderer is the successful Tenderer, upon our receipt of copies of the contract signed by the Tenderer and the performance security issued to you upon the instruction of the Tenderer; or (b) if the Tenderer is not the successful Tenderer, upon the earlier of (i) our receipt of a copy of your notification to the Tenderer of the name of the successful Tenderer; or (ii) twenty-eight days after the expiration of the Tenderer's Bid.

Consequently, any demand for payment under this guarantee must be received by us at the office on or before that date.

This guarantee is subject to the Uniform Rules for Demand Guarantees, ICC Publication No. 458.

[Signature]

Form 6: MANUFACTURER'S AUTHORIZATION FORM

[The Tenderer shall require the Manufacturer to fill in this Form in accordance with the instructions in the bid document. This letter of authorization should be on the letterhead of the Manufacturer and should be signed by a person with the proper authority to sign documents that are binding on the Manufacturer. Such certificate is not required where Manufacturer is the Tenderer.]

Date:
NIT No.:

To:
MD, WBMSCL

WHEREAS

We *[insert complete name of Manufacturer]*, who are official manufacturers of *[insert type of goods manufactured]*, having factories at *[insert full address of Manufacturer's factories]*, do hereby authorize *[insert complete name of Tenderer]* to submit a bid the purpose of which is to provide the following Goods, manufactured by us *[insert name and or brief description of the Goods]* and to subsequently negotiate and sign the Contract.

We hereby extend our full guarantee and warranty in accordance with the General Conditions for Goods, with respect to the Goods offered by the above firm.

Authorised Signatory of the Manufacturer _____
Name _____
Designation with stamp _____
Date _____

Form 7: PERFORMANCE STATEMENT FORM

(For the period of last three years, if applicable)

Bid no: _____

Date of Opening: _____

Name of the Firm _____

Order placed by (Full address of purchaser)	Order no & date	Description & quantity of ordered items	Value of Order	Date of completion of Delivery		Remarks indicating reasons of late delivery, if any	Was the supplies of goods satisfactory
				As per Contract	Actual		

Signature and seal of the Tenderer

Countersigned by and seal of Chartered Accountant _____

Form 8: STATEMENT OF BREAKUP OF DUTIES AND TAXES
(to be uploaded for each SCHEDULE)

Sl. No.	Particulars	Percentage	Price (In INR)
1	Basic Price of equipment including accessories, ancillaries & warranty (as mentioned)		
2	Applicable GST		
Gross Price			

Form 9:
Deleted

**Form 10(a): PRICES FOR CONSUMABLES
(to be uploaded for each SCHEDULE)**

Sl. No	Items	All inclusive cost of 1(one) number in INR
1		
2		
3		
4		

NOTE: The Tenderers should furnish the price of all the Consumables of the offered model in a separate sheet. If any Tenderer do not submit the price of any of the consumables, it will be presumed that those Consumables shall be supplied FREE OF COST by the Tenderer during Warranty & CMC period.

**Form 10(b): PRICES FOR SPARES
(to be uploaded for each SCHEDULE)**

Sl. No	Items	All inclusive cost of 1(one) number in INR
1		
2		
3		

Form 11: TURNOVER CERTIFICATE

I certify that Average Annual Turnover of *(insert the name of the company)* in India in medical equipment division during the last 3 Financial Years (2014-15, 2015-16 & 2016-17) / (2015-16, 2016-17 & 2017-18) (in INR) *(strike out which is not applicable)* is Rs. as per the Audited Accounts of the Organization.

Signature and seal of Chartered Accountant

Form 12: DECLARATION OF QUALITY CERTIFICATION OF EQUIPMENT

(To Be Notarised)

AFFIDAVIT

I am the authorised signatory of -----, (name of Company) and in the context of supply and/ or installation of the ----- (Name of the Equipment, Name of the offered model) which is an (name of class) solemnly affirm and declare as follows:

1. That the device is in conformity with the requirements of the medical devices directive 93/42/EEC which apply to it.
2. That the above mentioned model of (Name of the Equipment) is CE (European Conformity) approved.
3. That the vendor company and /or supplier on whose behalf I am the authorised signatory admit the responsibility on the subject relating to the standard and safety of the equipment.
4. That the original copy/ photo copy of the CE certification of the offered model is enclosed and / or shall be subject to inspection on demand.
5. That the statements made in the above paragraphs are true to my knowledge and belief.

DEPONENT"

Section VI. Contract Forms

Form 1: PERFORMANCE SECURITY

[Insert: No Performance Security shall be requested or the bank, as requested by the successful Tenderer, shall fill in this form in accordance with the instructions indicated]

Date: *[insert date (as day, month and year) of Bid Submission]*

ITB No. and title: *[insert no. and title of bidding process]*

Bank's Branch or Office: *[insert complete name of Guarantor]*

Beneficiary: *[insert legal name and address of WBMSCL]*

PERFORMANCE GUARANTEE No.: *[insert Performance Guarantee number]*

We have been informed that *[insert complete name of Supplier]* (hereinafter called "the Supplier") has entered into Contract No. *[insert number]* dated *[insert day and month]*, *[insert year]* with you, for the supply of *[description of Goods and related Services]* (hereinafter called "the Contract").

Furthermore, we understand that, according to the conditions of the Contract, a Performance Guarantee is required.

At the request of the Supplier, we hereby irrevocably undertake to pay you any sum(s) not exceeding *[insert amount(s)¹ in figures and words]* upon receipt by us of your first demand in writing declaring the Supplier to be in default under the Contract, without cavil or argument, or your needing to prove or to show grounds or reasons for your demand or the sum specified therein.

This Guarantee shall expire no later than the *[insert number]* day of *[insert month]* *[insert year]*,² and any demand for payment under it must be received by us at this office on or before that date.

This guarantee is subject to the Uniform Rules for Demand Guarantees, ICC Publication No. 458, except that subparagraph (ii) of Sub-article 20(a) is hereby excluded.

[signatures of authorized representatives of the bank and the Supplier]

¹ The Bank shall insert the amount(s) specified in the SCG and denominated, as specified in the SCG, either in the currency(ies) of the Contract or a freely convertible currency acceptable to WBMSCL.

² Dates established in accordance with Clause 12 of the General Conditions of Contract ("GCG"). WBMSCL should note that in the event of an extension of the time to perform the Contract, WBMSCL would need to request an extension of this Guarantee from the Bank. Such request must be in writing and must be made prior to the expiration date established in the Guarantee. In preparing this Guarantee, WBMSCL might consider adding the following text to the Form, at the end of the penultimate paragraph: "We agree to a one-time extension of this Guarantee for a period not to exceed [six months] [one year], in response to WBMSCL's written request for such extension, such request to be presented to us before the expiry of the Guarantee."