

## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Bhawan Complex  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340307/40340308

E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

SUPPLY OF BLOOD-BANK EQUIPMENTS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT- 25 /2015

Dated – 13.07.2015

The following amendment have been made in the tender document,

### Amendment - I

#### REVISED TECHNICAL SPECIFICATION

##### SCHEDULE – VI

#### INCUBATOR WITH THERMOSTATIC CONTROL (45 LTR.)

- a) Control: Equipped with Microprocessor controlled with a large digital display.
- b) Work Chamber Volume: minimum 45 Litres.
- c) Working temp.: Ambient temp. +5°C to 75°C whereas room temperature is around 20° – 24° C.
- d) Temp. deviation: ± 0.2°C.
- e) Display: Digital display located on front panel of the equipment.
- f) Easy to clean: An interchangeable silicon door gasket & removal of shelf supports make thorough cleaning and decontamination of the working chamber easy & safe
- g) Inner casing: Should be made of corrosion resistant 304 grade stainless steel.
- h) Shelves (adjustable): Perforated shelves made of corrosion resistant 304 grade stainless steel. Standard supply 2 nos. , inter shelves gap – 6" approx.
- i) Heat conduction: Gravity convection and air circulation.
- j) Heater on rear / side / bottom of the equipment
- k) Power requirement: 230 V, single phase, 50 Hz.
- l) After sales service: Prompt & efficient after sales service should be available from Kolkata.

- m) Product certification: CE / US FDA / BIS certified
- n) Quality Certification: ISO certificate of manufacturing unit
- o) Product should have alcohol based L type calibrated thermometer (temp. range 0°-100°C).