



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

Supply and Commissioning of Medical Equipment for Cath Lab of Midnapore Medical
College & Hospital, Paschim Midnapore

[Bid Reference No.: WBMSCL/NIT-235/2023,(Schedule-I to IV); Dated:10-05-2023](1st Call)]

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 322/2023	Dated-22.06.2023
---	------------------

Amendment-II

- i. **Comprehensive Maintenance Contract (CMC):** The selected bidder shall enter into CMC after expiry of warranty and CMC should include the following:
 - i. The equipment including all other accessories and ancillaries as given in the specifications of the equipment
 - Bidders must enter in to the CMC with WBMSCL directly / third party as authorized by H&FW Dept/ WBMSCL etc. Bidder must accept the same CMC rate, if the CMC is offered by the third party/WBMSCL then no escalation, no extra cost will be provided. If any bidder failed to execute the CMC with WBMSCL or third party (as authorized by WBMSCL or H&FW Department), WBMSCL may initiate legal action against the selected bidder. The selected bidder may be debarred to participate in the future tender
 - The execution of CMC of any equipment will be at sole discretion of the WBMSCL management.
 - As per tender terms and condition third party items should be included within the CMC rate.
 - **Breakdown Calls:** as many numbers as may be required to attend to resolve the complaint lodged by the end-users.
 - **Preventive Maintenance:** The selected bidder should attend periodic planned preventive maintenances in the following manner:

Equipment	Mandatory preventive Maintenance Service (PMS) visit per year		Remarks
	Warranty	CMC	
Defibrillator	2	2	1. Supplier/ authorised service provider must attend all breakdown calls during warranty / CMC period. 2. The supplier should provide Preventive Maintenance Services (PMS) yearly / bi - yearly / quarterly in equal interval to fulfill the minimum number of mandatory PMS as recommended in the previous column during warranty / CMC period. 3. The preventive maintenance includes testing & calibration as per technical / service / operational manual, spares, all software updates and labour.
BIPAP Machine			
ETO sterilizer Machine			
Echocardiogram Machine with Adult, Pediatric, vascular and TEE probe			

The cost for Year wise CMC charges after completion of 2 (two) years warranty in percentage of the sum of quoted price of the equipment in the BOQ to be paid to the selected bidder(s) in the following manner:

ITEM	CMC charge s in percent age(%) for Year1	CMC charge s in percent age(%) for Year2	CMC charge s in percent age(%) for Year3	CMC charg es in perce ntage (%) for Year4	CMC charg es in perce ntage (%) for Year5	CMC charge s in percent age(%) for Year6	CMC charge s in percent age(%) for Year7	CMC charge s in percent age(%) for Year8	CMC charge s in percent age(%) for Year9	CMC charges in percentag e(%) for Year10
Defibrillator	3.0	3.25	3.50	3.75	4.0	4.25	4.50	4.75	5.0	5.25
BIPAP Machine	3.0	3.25	3.50	3.75	4.0	4.25	4.50	4.75	5.0	5.25
ETO sterilizer Machine	3.0	3.25	3.50	3.75	4.0	4.25	4.50	4.75	5.0	5.25
Echocardiogr am Machine with Adult, Pediatric, vascular and TEE probe	4.0	4.25	4.50	4.75	5.0	5.25	5.50	5.75	6.0	6.25