

# Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340307/40340308

E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

SUPPLY OF MEDICAL EQUIPMENTS FOR MULTI / SUPER SPECIALITY HOSPITALS OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-15/2016

Dated – 14.04.2016

The following amendments have been made in the tender document,

## Amendment - I

### REVISED TECHNICAL SPECIFICATION

#### SCHEDULE - VI

#### Trial lens sets with frames

1. Should be supplied with set of 226 trial lens with durable metal frame

#### Technical Specification

2. Lens set
3. Set of 226 trial lenses with polished metal rims in an elegant carrying case.
4. **Minimum** 70 Spherical Concave Lenses: 0.12 to 20.00 Diopters (0.25D increments)
5. **Minimum** 70 Spherical Convex Lenses: 0.12 to 20.00 Diopters (0.25D increments)
6. **Minimum** 34 Cylindrical Concave Lenses: 0.12 to 6.00 Diopters (0.25D increments)
7. **Minimum** 34 Cylindrical Convex Lenses: 0.12 to 6.00 Diopters (0.25D increments)
8. Prismatical Lenses: 0.5(2), 1(2), 2(2), 3, 4, 5, 6, 7, 8
9. Accessory Lenses: Red, Green, Slit, Dark, **Aperture(1), Block (1)**

## Metal Frame

10. Should be durable and light weight (Adult and Paediatric)
11. Number of lenses which can be inserted in to left or right lensframe simulataneously:4 pieces
12. Degree of lens rotation around optical axis in the lens frame:360 degree
13. Horizontal and vertical bridge adjustments.
14. Free-rocking saddle bridge assures comfortable nose fit.  
Range of nose rest adjustment Length:0 ~ 14 mm,angle:0 ~ 30 degree
15. Separate PD adjustment for each eye to compensate for asymmetrical factors in facial structure.
  - a) PD of both eye: 48 ~ 80 mm
  - b) Left or right PD: 24 ~ 40 mm
16. Individual adjustments for temple length and angle.
17. Adjustment for rotating cylinders to correct axis.
18. Scales with large, easy-to-read numerals.
19. **Should be CE approved Product**