

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited Swasthya Sathi GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320 E mail: procurement@wbmsc.gov.in

Procurement of Medical Equipment and Urology Instruments set for Burdwan Medical College and Hospital, Purba Burdwan.

(Submission of Bid through online)

Bid Reference No.: WBMSCL/NIT-572/2023

Dated-20.09.2023

Amendment-I

REVISED TECHNICAL SPECIFICATION ITEM-1 High-frequency surgery units for resection

- Microprocessor based unit, modular construct, custom configurable, Facility with plug and play feature using digital instrument recognition and high resolution colour TFT display with supporting functions like focus view/ preview for better patient safety and instrument usage counter, Software and Hardware upgradable with latest model.
- 2. Output rating of 300 Watts under standard testing conditions.
- 3. The unit should be able to store minimum 100 Programmes with various settings. Programmable programs are to be available.
- 4. Monopolar cutting and coagulation.
- 5. Should have Bipolar Auto Start and Auto Stop facility.
- 6. The unit shall have Protective circuit and alarm system for the disconnection for patient plate.
- 7. It should utilize Tissue Density feedback circuitry & Microprocessor based controlled output for Automatic response to changes in tissues density.
- 8. To be supplied with two pedal Footswitch, separate for cut and coagulation.
- 9. Power output should be automatic depending on the tissue impedance.
- 10. Separate control to change the output voltage in steps.
- 11. Arc regulation should be available in the unit to provide the initial high power spike required for smooth cutting.
- 12. Suitable accessories, if required should be quoted.
- 13. Should be suitable for all surgeries with Bipolar cut and Coagulation.

14. Should be able to support both monopolar and bipolar TURP / TURBT

15. Equipment should be European CE/ US FDA / BIS / CDSCO Certified.

- 16. The unit must be UL Certified for patient safety.
- 17. The unit should confirm to IEC 60601-2-2 standard.
- 18. All Accessories should be from same Manufacturer
 - Following Accessories to be supplied along with the Unit
 - 1. Two Pedal Footswitch.
 - **2.** Single Pedal Foot Switch.
 - **3.** Mono polar hand switch with two buttons, Disposable (Qty-50).
 - 4. Reusable Bipolar Forceps, straight tip 1mm, Blunt, Length 190mm (Qty-1).
 - 5. Reusable Bipolar connecting cable for Bipolar Forceps, 4mts (Qty-1).
 - 6. Reusable monopolar connecting cable for TURP / TURBT (Qty-1).
 - 7. Disposable Patient plate with equipotential ring and cable (Qty-50).
 - 8. Optional upgrade for Vessel Sealing.

ITEM-2

1. Forceps for removal of foreign bodies of 5Fr each and 7Fr each with double action jaws, flexible, length 40 cm for Cystoscope – (Qty. – 2 each)

2. Sachse Urethrotome sheath, 20-22 Fr with channel. (Qty.-1 each)

3. Visual Obturator for urethrotome sheath. (Qty.-1 each)

4. Telescope Bridge for used with urethrotome sheath with 5 fr channel for instrument. (Qty.-1 each)

5. Supplementary Sheath, side open for introduction of balloon catheter, to slip on urethrotome sheath. (Qty.-1 each.)

6. Sachse cold knife, straight. (Qty-3)