



## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340307/20

E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

Supply and Commissioning of different types of Medical Equipment at Medical College and Hospital of the Govt. of West Bengal

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-215/2018

Dated-12.10.2018

### Amendment-IV

#### REVISED TECHNICAL SPECIFICATION

#### 800 mA X-RAY:

Parameter/Feature	Threshold Limit/Range
Fluoroscopic parameter	
Pulsed Fluoro in mA	(1 -150) or more
X- Ray Tube	
mAs in .6 Sec & at 90 KV	300 or more

#### ACCESSORIES

LEAD HANGER STAND(TO HOLD 6 HANGERS)-02 NOS	
--	--

