



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

SUPPLY OF EQUIPMENTS FOR MOTHER &CHILD HUB OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 109 /2017

Dated-01.12.2017

The following amendment has been made in the tender document,

Schedule-X

REVISED SPECIFICATION VACCUM EXTRACTOR

1. Foot operated electric sucker
2. Pressure 600 mm Hg
3. Power consumption 40 watt - Pump noise 60 db
4. Cup (Silicon Rubber):
Metal cup; chair & handle - 4, 5 cm
Silastic cup - 50 mm, 55 mm, 60 mm.
Smooth external surface i.e., smooth longitudinal ridge to observe rotation.
Suitable for autoclave sterilization.
5. Accurate functioning
Low maintenance
Tough construction
6. Digital display of pressure
7. All silicon tubes
8. Trolley
9. Product certification: CE / US FDA