

West Bengal Medical Services Corporation Limited Swasthya Sathi GN-29, Salt Lake, Sector-V Kolkata-700091

Phone No (033) 40340308/412 E mail: procurement@wbmsc.gov.in

Procurement of equipment for CCU/HDU and other for COVID 19 Hospitals of the State of West Bengal.

(Submission of Bid through online)

Bid Reference No.: WBMSCL /NIT-205/2020

Dated -13.10.2020

The following amendment have been made in the tender document,

<u> Amendment – I (Revision of Technical Specification)</u>

The revised technical specifications for the item is given below,

Cold storage for 6 dead bodies

- 1. Entire storage should be made of SS 304
- 2. Size of chamber (inside): 2100 2400 mm (length) \times 500 600 mm (breadth) \times 300 450 mm (height)
- 3. 2 storied (3 column & 2 rows; i.e. maximum 2 bodies vertically)
- 4. Temperature range: $(+1 \text{ to } +5)^{\circ}\text{C}$
- 5. Must be equipped with Digital temperature display along with high or low temperature alarm or hooter
- 6. Number of Refrigeration unit: 2 (1 working and 1 standby)
- 7. Must be plug-in type for easy operation
- 8. Refrigerant: Green gas (non CFC)
- 9. Insulation: High density PUF panel (40 kg/cu. m)
- 10. Insulation thickness: minimum 80 mm
- 11. CFC free Green PUF panel

- 12. Trolley must be mounted on wheel for easy movement
- 13. Door activated interior LED (2 sets)
- 14. Individual lockable chamber
- 15. Company or Product should be certified by ISO / appropriate authorities
- 16. Automatic defrost
- 17. Should be CE/BIS certified"