

Revised Technical Specifications

Microscopes - Pathological And Research As Per Is:4381, Is:5204-IS:4381, IS:5204

Specification	Specification Name	Bid Requirement (Allowed Values)
CONFORMITY / TYPE / EYE PIECE - OBJECTIVE DETAILS	Microscope Type	research binocular non hinged type with built in light and with light intensity regulator
	ISI MARKED	No
	Eye piece Type	Compensating
	Eye Piece with magnification	Set of Two For Binocular 10x
	Objective Type	achromatic
	Objective Magnification	100X
	Numerical Aperture of Objective	1.25
	Type of lamp for Illumination	led
ACCESSORIES	Extra eyepieces (inclusive in the scope of supply)	NA
WARRANTY	Warranty	3, 4, 5 Or higher
Additional Specification Parameters - Microscopes - Pathological And Research As Per Is:4381, Is:5204-IS:4381, IS:5204		
Specification Parameter Name		Bid Requirement (Allowed Values)
Objective Magnification		4X,10X,40X & 100X (oil immersion)
Certification		CE / US FDA / BIS
Binocular eyepieces confirming to the requirements of IS : 8275 / 1976(latest)		yes
Warranty		3 years (ignore 4 & 5 or higher years)