



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/319

E mail: procurement@wbmsc.gov.in / to@wbmsc.gov.in

Supply and Commissioning of Hysteroscopy and Laparoscopy Set for Bankura
Sammilani Medical College & Hospital

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-406/2023

Dated- 18.07.2023

Amendment-I

Schedule – I -Hysteroscopy Set

Technical Specification for diagnostic & operative Hysteroscopy set

| Sl. No. | Description | Qty |
|---------|---|---------|
| 1 | Hysteroscopy Diagnostic Sheath for 2.9 mm telescope | 01 Nos. |
| 2 | Full HD Telescope 2.9 mm 30 Degree with Rod Lenses, Optical system, Autoclavable with incorporated fiber optic light transmission length minimum 200 mm; | 01 Nos. |
| 3 | Operative sheath (Inner & Outer) for 2.9 mm forward oblique telescope 30 degree with 5 Fr instrument Channel Auto lock system with Discreet inflow & outflow channels | 01 Nos. |
| 4 | Semi Rigid Scissor Single Action 5Fr | 01 Nos. |
| 5 | Semi Rigid Alligator / Grasping Forcep 5 Fr | 01 Nos. |
| 6 | Semi Rigid Biopsy Forcep 5 Fr | 01 Nos. |
| 7 | Bipolar L Hook 5 Fr | 01 Nos. |
| 8 | Continuous Flow Inner outer Sheath with Obturator For 2.9mm Telescope for TCRC | 01 Nos. |
| 9 | Mono as well as Bipolar working Element | 01 Nos. |
| 10 | Cord For Working Element | 01 Nos. |
| 11 | Bipolar Cutting Loop | 01 set |
| 12 | Bipolar Ball Electrode | 01 Nos. |
| 13 | Collins knife | 01 Nos. |

| Sl. No. | Description | Qty |
|---------|--|---------|
| 14 | Tubal Recanalization Guide wire | 01 Nos. |
| 15 | Ellik Evacuator | 01 Nos. |
| 16 | Full HD Telescope 4 mm 30 Degree with Rod Lenses, Optical system, Autoclavable with incorporated fiber optic light transmission length minimum 200mm; | 01 Nos. |
| 17 | Diagnostic Sheaths | 01 Nos. |
| 18 | Operative sheath (Inner & Outer) for 4 mm forward oblique telescope 30 degree with 5Fr to 7 Fr instrument Channel Auto lock system with Discreet inflow & outflow channels Operative Sheath Continuous flow | 01 Nos. |
| 19 | Mono polar Bug Bee Electrode | 01 Nos. |
| 20 | Semi Rigid Scissor Single Action 5Fr to 7 Fr | 01 Nos. |
| 21 | Semi Rigid Alligator/Grasper 5Fr to 7 Fr | 01 Nos. |
| 22 | Semi Rigid Biopsy Forcep 5Fr to 7 Fr | 01 Nos. |
| 23 | Continuous Flow Inner Outer Sheath For 4 mm Telescope for TCRE | 01 Nos. |
| 24 | Mono as well as bipolar Working Element. For 4 mm Telescope | 01 Nos. |
| 25 | Loop Straight, 90 Degree | 01 Nos. |
| 26 | Collins knife | 01 Nos. |
| 27 | Endomat/ Hysteropump <ul style="list-style-type: none"> ➤ Flow & Pressure controlled Irrigation System ➤ Liquid Pressure Monitoring Technology ➤ Disposable Pressure transducer to avoid Patient Fluid Contamination ➤ Real Time Pressure & Flow Monitoring of Intra-Uterine Cavity I specific surgical site ➤ Built-in Low Noise irrigation pump ➤ Surgery Specific Custom Modes : Hysteroscopy & versatile ➤ Preferably indicator for amount of consumed fluid by the patient ➤ Preferably for Saline & Glycine Mode settings Alarm indication for consumed fluid after safety threshold ➤ Priming / Flush for removing air bubble from tubing at initialization ➤ Touch Screen LCD Interface <p> Pressure Range: 10 to 400 mmHg \pm 10% Flow Range: Irrigation : 100mL to 2000mL \pm 10% Pressure Setting: Irrigation : 10mmHg 1 step Flow Setting: Irrigation : 100mL 1 step Operating Modes: Only Irrigation Mode Filter: Disposable Cartridge Filter (0.2 micron) Tubing : Silicon Tubing Tube Specifications: Silicon tubing for Irrigation and Suction of 3Mtr Touch Control Panel: 5" LCD Screen with brightness control To be supplied with reusable Irrigation Tubing Set Certification – US FDA / European CE (4 digit notified body) </p> | 01 Nos. |

