



## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320

E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

Supply and Commissioning of Ophthalmic OT Table with Surgeon Stool for Satellite Eye OT Centre

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-081/2023

Dated-21.02.2023

### Amendment-I

#### Revised Technical Specification

#### Ophthalmic OT Table With Surgeon Stool

1. The versatile Operation Theatre table suitable for all varieties of eye surgeries.
2. The OT Table should have motorized up and down facility with foot switch control.
3. Motorized up and down movement to be soundless, smooth and without any jerk
4. OT table should have adjustable head rest and adjustable wrist support.
5. The OT table should have instrument tray for eye surgery fitted to table and flexible for attachment and detachable.
6. Overall length to be 1800mm or more and width 600mm or more.
7. Up and down movement range to be 690 mm to 900 mm or more.
8. Forward and reverse tilting with clutch break facility of at least 20 degree or more.
9. Weight bearing capacity of at least 150 kg or more.
10. The adjustable Cushioned Head Rest.
11. Equipment should be CE / USFDA / BIS certified.

#### Surgeon Stool

- Deluxe Model Pneumatic Operating Stool with backrest and Metal Piston for height Adjustment.
- Caster Wheel – 4 or more