



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

SUPPLY AND COMMISSIONING OF FACILITY BASED NEWBORN CARE (FBNC)
EQUIPMENTS IN THE HOSPITALS AND AND MEDICAL COLLEGES OF THE GOVT. OF
WEST BENGAL

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT- 38 /2017

Dated – 20.07.2017

The following amendment have been made in the tender document,

Amendment –II

Pulse Oxymeter

A. **Essential**

1. It should be portable, light weight and desktop pulse oxymeter.
2. It should have a large LED / LCD display (minimum size of 4.0 inches) so that numerical values are readable from a distance of at least 10 feet with naked eyes. Display should indicate numerical values of oxygen saturation and pulse rate along with graphical display of plethysmogram.
3. It should have a SpO₂ range of 1 to 100%.
4. It should have visual alarm for both upper and lower SpO₂ and Pulse rate and adjustable audio alarm volume.

5. It should have a facility to capture data for real time analysis at least for 24 hours.
6. SpO₂ cable and probe: 2 (two) reusable neonatal probes (flex with moulded or detachable type rubber or silicon made) & 1 (one) finger probe including cable with life span of at least one year. The cable length should not be less than 3 meters.
7. It should have a minimum 5 hours of internal battery back-up. The battery should be rechargeable type.
8. **Standards, Safety and Training**
 - i. The Product should have safety certificate from a competent authority US FDA and CE ("Conformite Europeene") from European Union notified body having 4 digit identification number (Copy of certificate to be produced).
 - ii. The medical devices should be ISO certified manufacture.
 - iii. Should have onsite service facility. The service provider should have the necessary equipment recommended by the manufacturer to carryout preventive maintenance test as per guidelines provide in the service / maintenance manual.
 - iv. Environmental Factors

Operating temperature.	10°C to 40°C
Storage temperature	-10°C to 50°C
Operating humidity.	15% to 95%
9. Warranty for 2 years and 5 years of CMC after warranty. The warranty and CMC includes entire equipment including battery, SpO₂ extension cable etc. except probes.
10. Bidder has to supply 02 (two) probes along with the machine. The warranty of each probe should be minimum of 01 (one) year. The supplied 02 (two) probes shall cover the entire equipment warranty of 02 (two) years. Bidders will have to replace the probe if it does not work within the warranty period of 02 (two) years.
11. Bidder should quote the price of different types of cables, battery and cost for upgradation to monitor methemoglobin (SpMet) etc. in Form 9 (a) & 9 (b). The same rate will be valid for 02 (two) years. After 02 (two) years, the rate may vary but should not exceed RBI inflation index.

B. Desirables (Qualifying marks in desirable is 65)

S. No.	Particulars/ Items	Characteristics	Variations	Points
1	Signal delay (Response time)	< 15 sec	NA	20
		15 sec to 25 sec	NA	10
		> 25 sec	NA	1
2	Averaging time	< 3 sec	NA	15
		3 sec to 8 sec	NA	10
		> 8 sec	NA	5
3	Accuracy of SpO ₂ (70% - 100%)	With motion	± 2 digits	5
			± 3 digits	3
			Not available	0
		With low perfusion	± 2 digits	5
			± 3 digits	3
			Not available	0
4	Accuracy of pulse rate	With motion	± 3 digits	5
			± 5 digits	3
			Not available	0
		With low perfusion	± 3 digits	5
			± 5 digits	3
			Not available	0
5	Frequency of capturing data for trending	Every 2 sec interval	NA	10
		Every 4 sec interval	NA	5
		Every 8 sec interval	NA	1
6	Frequency of capturing data for future review	Every 2 sec interval	NA	5
		Every 4 sec interval	NA	3
		Every 8 sec interval	NA	1
7	Duration of trending with data captured every 4 sec interval or equivalent duration if captured every 2 or 8 sec interval	> 120 hrs	NA	10
		48 hrs to 120 hrs	NA	8
		< 48 hrs	NA	5
8	Internal battery back up	> 9 hrs	NA	10
		5 – 9 hrs	NA	5
		< 5 hrs	NA	1
9	Screen Size	> 5 inches	NA	10
		4.0 – 5 inches	NA	5

Amendment - III

Form 9(a): PRICES FOR CONSUMABLES (to be uploaded for each schedule)

Sl. No	Items	Basic Price of 1(one) number in INR excluding GST
1		
2		
3		
4		

NOTE: The Tenderers should furnish the price of all the Consumables of the offered model in a separate sheet. If any Tenderer do not submit the price of any of the consumables, it will be presumed that those Consumables shall be supplied FREE OF COST by the Tenderer during Warranty & CMC period.

Form 9(b): PRICES FOR SPARES (to be uploaded for each Schedule)

Sl. No.	Items	Basic Price of 1(one) number in INR excluding GST
1		
2		
3		
4		
5		
6		