



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

Supply and Commissioning of different medical equipment for Critical Care Block(CCB)
project(Item category-Type B)

(Submission of Bid through *online*)

2nd call of Bid Reference No.: WBMSCL/NIT-176/2022; dated-18.04.2023

Bid Reference No.: WBMSCL/NIT-428/2023	Dated-26.07.2023
--	------------------

Schedule-II

AMENDMENT-I

REVISED TECHNICAL SPECIFICATION

BLOOD FLUID WARMER

1. Flow Rates should be maximum 1500 ml/ hour
2. Should have temperature prefixed at 37 Degree Temperature
3. Should be easily transportable
4. Should able to attach to I V pole and standard electrical sockets
5. Should use multichannel counter current heat exchanger/ equivalent technology
6. Should have audible and visual alarms for Temperature
7. Should have automatic cut off for set temperature
8. Should be easy to use and to clean
9. Calibration certificate should be issued during the installation

10. 5 disposable adult and 1 no. of pediatric warming sets should be supplied along with each machine
11. Warm up time at least 120 seconds
12. If available- Consumables with built in filter should be provide.
13. It should be US FDA or CE ("Conformite Europeene") from European Union notified body having 4 digit identification number or BIS approved (Certificate to be submitted).