



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340308/319

E mail: procurement@wbmsc.gov.in / mp@wbmsc.gov.in

Procurement of Medical Equipment for Medical College and Hospital

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 216/2020

Dated- 05.11.2020

AMENDMENT-I

TECHNICAL SPECIFICATIONS

Computerised Tonometer with Pachymeter

1. Should have fully automated operation.
2. It should have ocular pressure measurement facility with measuring range of 1-60 mm of Hg in different cases in non contact method.
3. It should have Corneal Thickness measurement facility in pachymeter device with measuring range of 0.4-0.75mm (0.001 mm step)
4. It should have standard LCD monitor with touch panel.
5. It should have data transport terminal with USB port, RS232 (Export), LAN (Export) as appropriately required for connectivity
6. It should have standard weight.
7. It should have table top motorised stand.
8. It should have standard 220 ± 5 V AC power supply
9. European CE ("Conformite Europeene") or US FDA Certificate or BIS