



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

SUPPLY AND COMMISSIONING OF MEDICAL EQUIPMENT AT MEDICAL COLLEGE AND HOSPITAL OF THE GOVERNMENT OF WEST BENGAL (Group 1).

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 216/2018

Dated-12.10.2018

The following amendments have been made in the tender document,

Amendment –II (Revision of Technical Specification)

The revised technical specifications for the item is given below,

BOD Incubator:

PARAMETER / FEATURE	THRESHOLD LIMIT / RANGE
RANGE OF AMBIENT TEMPERATURE IN ° C	10 TO 35
OPERATING /PRODUCIBLE/ ACHIEVBABLE TEMPERATURE RANGE IN BOD INCUBATOR IN ° C	0-70
RANGE OF TEMPERATURE SENSITIVITY IN ° C	± 0.1
RANGE OF TEMPERATURE UNIFORMITY IN ° C	± 0.5
MAXIMUM HEATING RATE IN ° C/ MIN	5 OR LESS
MINIMUM HEATING RATE IN ° C/ HOUR	2 OR MORE

MAXIMUM COOLING RATE IN ° C/ MIN	5 OR LESS
MINIMUM COOLING RATE IN ° C/ HOUR	2 OR MORE
AIR CIRCULATION FORCED TYPE WITH FANS	2 OR MORE
INNER CHAMBER VOLUME IN CUBIC FT	6 OR MORE
INNER CHABER SHELVES	4 OR MORE
SUPPLY VOLTAGE/ FREQUENCY	220 V/50 HZ
MUST HAVE SPECIFICATION	
PARAMETER / FEATURE	YES/NO
DIGITAL TEMPERATURE CONTROLLER DEVICE	
PROFILE WITH DIFFERENT RAMP UP RAMP DOWN AND SOAK FEATURES	
PROFILE CONTAINING MORE THAN 8 FEED POINTS	
INDEPENDENT OVERHEAT PROTECTION SYSTEM INSTALLED WITH SEPERATE TEMPERATURE SENSOR AND BUZZER ALARM	
HUMIDITY CONTROL FOR CHAMBER, INSTALLED DEHUMIDIFIER FOR LESS THAN 50% REL HUMIDITY	
INNER CHAMBER WALL WITH INSULATED STAINLESS STEEL (SS 304)	
BOD OUTER WALL RUST AND CORROSION PROOF	
LOCK AND BUZZER ALARM FOR DOOR	
CALIBRATION AND OPERATING MANUAL	
CFC FREE REFRIGERATION SYSTEM	
QUALITY STANDARD AND SAFETY CERTIFICATION	
EUROPEAN CE OR US FDA	