



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

SUPPLY AND COMMISSIONING OF MEDICAL EQUIPMENTS (SET-II) FOR FORTRAUMA CARE FACILITIES AND BRGF HOSPITALS FOR THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVT. OF WEST BENGAL

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 63/2017

Dated -22.09.2017

The following amendments have been made in the tender document. The changes are highlighted in yellow in the document below,

1. TECHNICAL SPECIFICATION

SCHEDULE – II

Horizontal Autoclave Machine (Flash Sterilizer)

1. Stand-alone floor top autoclave 90 Litres, automatic. Single door, self sealing with high-quality silicone gasket.
2. Dimension should match to suit the volume of autoclave
3. The autoclave should have double wall sheet metal with suitable insulation in between walls.
4. Pre-set automatic cycles for unwrapped instruments, wrapped instruments/packs
5. Two automatic programmes, approx: 2.2 bar at 134 degree C and 1.1 bar at 121degree C with 3 sub-programs in each cycle.
6. Should have Bowic and leak test

7. The equipment should have automatic pressure control switch /automatic water control device to ensure that the equipment does not run dry.
8. System should have digital display for Temperature and Time. System should have analogue display for Pressure
9. System should have audio visual alarm for low water level, and other system errors.
10. Fit with appropriate water reservoir, automatic fill, autonomy for at least 7 cycles
11. Should have Double safety door locking device
12. Should have safety feature protection against over-pressure and over-temperature
13. Audible and visual alert upon cycle interruption or completion
14. Control panel with cycle indicators
15. System Configuration Accessories, spares and consumables
16. US FDA or European CE or BIS approved product.
17. The system should have recording facility.
18. Should run on 3 phase power supply.