



## Notice Inviting e-Tender

**West Bengal Medical Services Corporation Limited**

**Swasthya Sathi**

**GN-29, Salt Lake, Sector-V**

**Kolkata-700091**

Phone No (033) 40340307/320

E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

Supply and Commissioning of different medical equipment for new modular OT and ICU setup at  
IPGME&R and SSKM Hospital

(Submission of Bid through *online*)

2<sup>nd</sup> call of Bid Reference No.: WBMSCL/NIT-476/2023; dated-14.08.2023

Bid Reference No.: WBMSCL/NIT-592/2023	Dated-03.10.2023
--	------------------

### **Amendment-I**

## **REVISED TECHNICAL SPECIFICATION**

### **Schedule-V**

#### **Orthopaedic Electric Drill**

- Driving Unit Accessories (All Items should be made of Stainless Steel or better material)
- Driving Unit for Electric Drill includes Stand, Base, Foot Control, Tool Kit
- Drilling Hand Piece
- Reaming Hand Piece
- Oscillating Hand Piece
- Dual trigger mode for forward and reverse
- Reaming Speed –Maximum 250 RPM or more,
- Drilling Speed – Maximum 1000 RPM or more
- IPX6 / IPX9 rating for long term maintenance against water and moisture
- Saw hand piece should have two speed / progressive controls with standard and fast mode free speed have 10,000 to 12,000 cycles/ Minutes
- Battery charger should have 4 charging slots (one extra battery charger and one sets of battery should be given)
- Drilling, reaming and cutting (blades) - 10 pcs of blades of each different sizes.
- Equipment should be European CE (4 digit notified body)/USFDA/BIS/CDSCO approved.