

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/40340308

E mail: procurement@wbmsc.gov.in

SUPPLY OF MEDICAL EQUIPMENTS FOR MULTI / SUPER SPECIALITY HOSPITALS OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-15/2016

Dated – 14.04.2016

The following amendment has been made in the tender document,

Amendment - I

REVISED TECHNICAL SPECIFICATION

SCHEDULE - XII

Class – II A2 Biological Safety Cabinet (Size: 4 ft.)

1. The cabinet should be advanced microprocessor control, which supervises operation of all cabinet functions. Temperature-compensated air velocity sensor monitors both exhaust and down flow. 24-hour clock, UV timer, UV run hour meter, and blower run hour meter are standard.
2. There should be programmable PIN, which restricts unauthorized cabinet access.
3. The cabinet should have energy efficient ECM blower motor.
4. The cabinet should have long life DUAL ULPA / HEPA Filter for supply and exhaust (per IEST-RP-CC001.3) with 99.999% efficiency for particle size 0.1 to 0.3 microns.
5. Should be raised armrest for elevates the operators arms to prevent inflow grille blockage for safety work.
6. Work tray should be made of single piece stainless steel type 304 , with 4B finish

7. Programmable automatic UV light timer should simplify operation and extending UV light life and saving energy.
8. The Cabinet outer surface should have germicide coating for minimizing contamination.
9. There should be sliding front sash which can be fully opened to insert and remove large instruments.
10. The cabinet should come with following accessories :
UV lamp, electrical outlet sockets and fully SS movable stand with wheels & brakes [antimicrobial coated] for easy movement.
11. Should have service station at Kolkata with trained engineer.
12. Should be provided original literature.
13. Should be CE & US FDA approved product.