

Amendment-I

Revised Technical Specifications

Specification	Specification Name	Bid Requirement (Allowed Values)
Performance Parameters	Sample loading	Direct tube
	Type of sample	Whole blood, Serum , Plasma, Diluted urine
	PH	Yes Or higher
	Throughput (Number of Sample) per hour	60.0 Or higher
	Reagent pack Type	Compact reagent pack
	Probe wiper system/ attachment	Yes Or higher
	OEM liquid controls	Tri-level
	Calibration	Both Automatic and On demand/manual
Data management and Display	Patient sample data entry	Both barcode reader attachment or manual
	Type of patient data entry	alpha-numerical
	Type of display	TFT
	Printer	In built
Power requirements	Online UPS provided	Yes Or higher
	UPS Back up Time	30 mins
	Type of battery	Lithium ion
Miscellaneous Parameters	Number of installations in Central /State/PSU Govt Hospitals (Hint: Seller should supply a performance certificate of the device to the buyer incase demanded after placement of order)	More than 3
	Availability of toll free facility for technical support maintained by OEM or authorised agencies	Yes Or higher
	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty	2

Specification	Specification Name	Bid Requirement (Allowed Values)
	in the catalogue)	
Additional Specification Parameters - Electrolyte Analyzers		
Specification Parameter Name	Bid Requirement (Allowed Values)	
Reagents for reportable tests	750 nos. to be supplied with the machine and the rest 750 nos. will be supplied after 3 months	
Necessary calibrator and quality control serum	for 3 months to be supplied with the machine and the rest necessary for 3 months to be supplied after that	
Certification	CE IVD / USFDA / BIS	
Warranty	3 years (ignore 2 years)	