

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Bhawan Complex
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/40340308

E mail: procurement@wbmsc.gov.in

SUPPLY OF PICU (PAEDIATRIC INTENSIVE CARE UNIT) EQUIPMENTS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/ NIT-19/2015

Dated-03.06.2015

The following amendment has been made in the amendment II of the technical specification of Computerized Radiographic (CR) System (Without Printer) published on 15.06.2015,

Amendment – IV

SCHEDULE – II

Computerized Radiographic (CR) System (Without Printer)

[This machine would be attached to Mobile X-Ray Machine at Schedule I. Both the systems should be compatible with each other]

1. The system should comprise of:
 - a. Image recording system (Image plate/cassette)
 - b. Image reader
 - c. Image processing work station
 - d. Accessories and consumables
2. Should be of Single cassette (size – 14" x 17") speed – 50 cassette/hr or more with standard resolution of 10 pixel / mm

3. Acquisition of image at monitor after insertion of the cassette: 40 sec or less
4. Cassettes: Should provide the following cassettes;-
 - i) 10 X 12 cassette – 2 no.
 - ii) 8 X10 cassette – 2 no.
5. Standard resolution of 10 pixel/mm
6. Console with monitor of 17" or more – 1 no.
7. UPS from reputed manufacturer having at least 10 minute back up with appropriate KVA for the entire system.
8. System should have capability to pan, zoom, annotate, crop, noise suppression/control multiple image layout, customized formatting, CD/DVD writing etc.
9. System should be scalable for connectivity to PACS.
10. The equipment should be new and unused. The manufacturing date should not be more than 180 days when it would reach the consignee address.
11. **Standard & Safety**
Should be of US FDA or CE ("Conformité Européene")
12. **Warranty & CMC**
The machine, UPS and all other accessories should be under warranty for 2 years and followed by 8 years CMC.